



Chir 3

The State of Wisconsin

Filed Aug 4 - 1960
10:35 am

STATE BOARD OF EXAMINERS IN CHIROPRACTIC

M. W. GROSOPP
CHAIRMAN
CLEAR LAKE, WISCONSIN

S. C. SYVERUD, SECRETARY
320 WEST MAIN STREET
MT. HOREB, WISCONSIN

R. P. HINTZ
SUPT. OF EXAMINATIONS
526 WASHINGTON ST.
BURLINGTON, WISCONSIN

STATE OF WISCONSIN

DEPT. OF STATE BOARD OF EXAMINERS IN CHIROPRACTIC

TO ALL TO WHOM THESE PRESENTS SHALL COME, GREETINGS:

I, S. C. Syverud, Secretary of the State Board of Examiners in Chiropractic, and custodian of the official records of said board, do hereby certify that the annexed rules and amendments to rules and regulations, relating to the State Board of Examiners in Chiropractic, were duly approved and adopted by this board on July 26, 1960.

I further certify that said copy has been compared by me with the original on file in this board and that the same is a true copy thereof, and of the whole of such original.

IN TESTIMONY WHEREOF, I have here-
unto set my hand and affixed the
official seal of the board at my
office, in the city of Mt. Horeb,
this 3rd day of August A.D., 1960.

S. C. Syverud D.C.
(Office)



The State of Wisconsin

STATE BOARD OF EXAMINERS IN CHIROPRACTIC

M. W. GROSKOPP
CHAIRMAN
CLEAR LAKE, WISCONSIN

S. C. SYVERUD, SECRETARY
320 WEST MAIN STREET
MT. HOLEB, WISCONSIN

R. P. HINTZ
SUPT. OF EXAMINATIONS
526 WASHINGTON ST.
BURLINGTON, WISCONSIN

ORDER OF THE STATE BOARD OF EXAMINERS IN CHIROPRACTIC ADOPTING AND AMENDING RULES

Pursuant to authority vested in the State Board of Examiners in Chiropractic by chapters 147 and 227 Wis. Stats., the agency hereby amends, and adopts rules as follows:

Chir. 3.07 Radiation (1) All x-ray equipment must be of shock-proof construction

(2) Lead cones or diaphragms should be used to confine the primary ray within the film area to reduce the secondary radiation to a minimum.

(3) A 2.5 m.m. external filter should be used on all x-ray tubes.

(4) Lead screen, or, preferably, lead lined booth with sufficient thickness, 1/16 inch lead.

(5) X-ray equipment should be located as to direct all primary rays to the outside or away from any occupied adjacent rooms.

(6) Lead, leaded material, steel, or concrete of sufficient thickness must be used where needed to protect persons in living quarters or adjacent offices in accordance with table on page 15 of the X-Ray Protection Handbook, No. 60, which is available for 20¢ in coin from the National Bureau of Standards, Department of Commerce, Government Printing Office, Washington, D.C.



The State of Wisconsin

STATE BOARD OF EXAMINERS IN CHIROPRACTIC

M. W. GROSKOPP
CHAIRMAN
CLEAR LAKE, WISCONSIN


S. C. SYVERUD, SECRETARY
320 WEST MAIN STREET
MT. HOREB, WISCONSIN

R. P. HINTZ
SUPT. OF EXAMINATIONS
526 WASHINGTON ST.
BURLINGTON, WISCONSIN

The rules and amendments contained herein shall take effect
on September 5, 1959 as provided in section 227.026 (1).

Dated: August 5, 1960

STATE BOARD OF EXAMINERS IN CHIROPRACTIC


S. C. Syverud, D.C., Secretary