

The scope statement for this rule, SS 078-12 was published in Register No. 682 on October 31, 2016 and approved by the Natural Resources Board on May 22, 2013. The rule was adopted by the Natural Resources Board on December 12, 2018 and approved by the Governor on June 14, 2019.

**ORDER OF THE STATE OF WISCONSIN NATURAL RESOURCES BOARD
REPEALING AND RECREATING AND CREATING RULES**

The Wisconsin Natural Resources Board proposes an order to **repeal** NR 247.03 (7); **amend** NR 106.145 (10)(a), (10)(d), (11), NR140.05 (13) note, NR 140.16 (1)(d) note, NR 157.20 (1), (3), NR 200.02 (6) note, NR 200.22 (1) (e) 6. b., NR 219.04 (1), NR 258.03 (15), NR 500.03 (146), NR 507.17 (4), (5), NR 507.26 (3) (b) 1., and NR 700.13 (1m) note; **repeal and recreate** Chapter NR 149, and NR 263.03 (4) relating to laboratory accreditation and affecting small business.

SS-22-12

Analysis Prepared by the Department of Natural Resources

1. Statute Interpreted:

s. 299.11, Stats.

2. Statutory Authority:

ss. 299.11 (3), 299.11 (4), 299.11 (5), 299.11 (7), 299.11(8), 299.11 (9), Stats.

3. Explanation of Agency Authority:

Section 299.11 (3) Stats. authorizes the department to seek recommendations of the certification standards review council for the general administration of the laboratory certification and registration program.

Section 299.11 (4), Stats. defines the applicability of the certification and registration rules to laboratories submitting data for covered programs.

Section 299.11 (5) Stats. allows the department to recognize certifications from other agencies, governments, and private organizations.

Section 299.11 (7) Stats. authorizes the department to promulgate rules for the certification of laboratories submitting data for covered program.

Section 299.11 (8) Stats. authorizes the department to promulgate rules for the registration of laboratories submitting data for covered programs.

Section 299.11 (9) Stats. authorizes the department to establish a regulated schedule of fees to cover the costs of administering a laboratory certification and registration program.

4. Related Statutes or Rules:

Sections 15.107 (12) and 93.12, Stats.

Chapters NR 101, 102, 106, 110, 123, 131, 132, 140, 150, 153, 155, 157, 182, 200, 204, 205, 206, 210, 211, 212, 214, 216, 217, 218, 219, 230, 233, 243, 254, 260, 261, 263, 273, 274, 290, 347, 500, 507, 518, 528, 538, 662, 664, 665, 700, 712, 716, 720, 738, 747, 809, 810, 811, 812 and 845.

5. Plain Language Analysis:

Chapter NR 149 sets requirements for the certification and registration of laboratories that submit data to the department for covered programs. Since the last major revision of the chapter, laboratory operations have undergone significant advances. Other state and national certification programs have promulgated and revised rules that reflect these advances. This version of ch. NR 149 incorporates many of those changes and, where appropriate, moderates them by incorporating suggestions expressed by our regulated community.

The proposed rule introduces efficiencies for administering the laboratory accreditation program, improves the structure used for certification and registration of laboratories, identifies clear steps and procedures for the certification and registration process, establishes a more equitable fee structure, clarifies requirements for proficiency testing of laboratories, stipulates procedures for on-site evaluations of laboratories, and adds specificity and flexibility to quality systems requirements for laboratories.

6. Summary of, and Comparison with, Existing or Proposed Federal Statutes and Regulations:

The US EPA has established a program for the certification of laboratories that analyze drinking water for compliance with the safe drinking water act. The US EPA delegates the authority to certify laboratories to states that have established equivalent programs. The proposed rule incorporates the latest changes in the regulations and the manual used by EPA to certify drinking water laboratories. Thus, this revision makes the Wisconsin laboratory accreditation program current with the US EPA's.

The US EPA sponsors a National Environmental Laboratory Accreditation Program (NELAP) for states that voluntarily seek such recognition. The procedures for accrediting laboratories under NELAP are contained in standards promulgated by the National Environmental Laboratory Accreditation Conference (NELAC). The proposed rule contains elements of the NELAC Standards recommended for incorporation by our regulated community. In most cases, the incorporated elements address standard practices commonly performed by laboratories.

7. Comparison with Similar Rules in Adjacent States:

All the adjacent states, Minnesota, Illinois, Michigan, and Iowa, have primacy from the US EPA to certify laboratories analyzing drinking water. Their rules must mirror federal requirements to maintain the states' authority. Our proposed revision makes the drinking water portion of our chapter current with those of the adjacent states.

Wisconsin, Minnesota, Iowa, and Illinois have similar certification, registration, or accreditation programs for laboratories analyzing wastewater, hazardous waste, and solid waste. Michigan requires certification only for those laboratories analyzing drinking water. Illinois and Minnesota are recognized NELAP accrediting authorities and NELAP rules agree or are stricter than those the department proposes for ch. NR 149. In addition, the Minnesota Pollution Control Agency (MPCA) accredits wastewater laboratories and its rules are like those proposed for ch. NR 149. Iowa has a certification program that is more limited in scope than ours because the state has few laboratories providing environmental analytical services other than the University of Iowa State Hygienic Laboratory.

To compare fees between the states, we used an average annual proposed fee for the 30 largest commercial laboratories as one category, and used a wastewater laboratory certified for BOD, TSS, Ammonia, and Phosphorous as an indicator of the typical wastewater laboratory fee. Using these two categories, Wisconsin proposed fees are \$5,311/\$1,114 for commercial/wastewater laboratories. Illinois assesses \$8,400/\$3,400 annually for these same types of laboratories. Minnesota's fees are \$10,900/\$1,800. Iowa's fees are lower than WI's for wastewater laboratories, but higher for commercial laboratories (\$10,800/\$800). Michigan charges \$6,729 for certification of drinking water laboratories; no certification for wastewater laboratories is required.

8. Summary of Factual Data and Analytical Methodologies Used and How Any Related Findings Support the Regulatory Approach Chosen:

To create this proposed rule, the Department engaged in a structured process to seek input from all stakeholders. The core of this effort consisted in convening a rule revision advisory committee (RAC) composed of all the members of the Certification Standards Review Council, a body authorized by s. 15.107 (1), Stats. The following constituencies were represented in the NR 149 RAC:

- Small and Large Municipal Wastewater Treatment Plants
- Commercial Laboratories
- Industrial Laboratories
- Municipal Environmental Group (MEG)
- Wisconsin State Laboratory of Hygiene
- Laboratory Accreditation Program

The NR 149 RAC envisioned a code that had greater specificity without sacrificing flexibility and alternatives for compliance. Over the course of approximately 30 meetings held from January 2014 to December 2017, the NR 149 RAC offered advice and guidance on every aspect of the Laboratory Accreditation Program. Meetings were facilitated by program staff. The agreements reached were captured in standardized documents reviewed and endorsed by the NR 149 RAC. These documents were used in drafting specific language included in the proposed rule.

The NR 149 RAC reviewed a complete final draft of the proposed chapter in February 2018. The comments received and the input received by the Laboratory Accreditation Program and other Department programs are reflected in this proposed rule.

The following table illustrates the methodologies and data considered in producing this proposed rule:

Methodology	Data Considered
Advisory Committee	Input from all stakeholders on all aspects of the Laboratory Accreditation Program.
Decision Making Rule	NR 149 RAC made decisions by reaching substantial agreement and when necessary, registering consensus.
Topic Prompts	Captured decisions made by NR 149 RAC on program administration, program structure, certification and registration process, proficiency testing, on-site laboratory evaluations and quality control.
Model Documents	Alternatives for certification and registration structure, fee structure, applications, and quality systems.
Comparative Analysis	Scope of certification and registration of current laboratories in the program to arrive at equitable fee structure. Analytical technologies were assessed and assigned a fee based on

	relative workload to evaluate them. Fee structure and assessments of certification programs in other states.
Feasibility	Certification and covered program staff reviewed changes endorsed by RAC to determine feasibility of implementation.

9. Analysis and Supporting Documents Used to Determine the Effect on Small Business or in Preparation of an Economic Impact Report:

In order to be equitable and efficient to all laboratories, the new fee schedule attempts to match the time spent by WDNR staff during laboratory inspections to fees paid (RVU’s) by removing the restrictive cap on the fees that larger laboratories are required to pay.

In terms of cost of certification, only 25 of commercial laboratories are expected to see a fee increase. The average fee increase for the commercial laboratories that are projected to see a fee increase, is estimated to be \$919 per laboratory per year. Two laboratories which certify for the maximum analytical methods will see their annual fees increase by \$1,222. The Department estimates that approximately 316 laboratories will see a decrease in fees by an average of \$72.

10. Effect on Small Business (initial regulatory flexibility analysis):

Small business laboratories are not likely to change their scope of certification under the proposed certification structure, as long as the costs for maintaining those certifications do not increase dramatically. In general, the proposed rule maintains these costs as in check.

When given options for quality control analyses that could reduce operation costs, laboratories are selective and respond that cost is not the sole determinant in selecting an option. Some small businesses are likely to continue to choose existing, more costly practices and may need to be educated in selecting valid and more economical alternatives.

Most operating costs in laboratories are associated with maintaining staff to perform analyses. The proposed rule does not require increases in staff to ensure compliance with it.

The Department concludes that the proposed rule provides flexibility in meeting many of its requirements. Small businesses may be able to realize some savings in implementing the proposed rule by judiciously selecting among the options contained in it. The specificity and flexibility contained in the proposed rule bring equity and uniformity to all laboratory operations and are likely to increase the competitiveness of small laboratories providing analytical services in and out of state.

11. Agency Contact Person:

Tom Trainor
 Certification Services/EAS
Tom.trainor@wisconsin.gov
 920-412-5970

WDNR
2984 Shawano Avenue
Green Bay, WI 54313

The consent of the Attorney General was requested on February 13, 2018 for the incorporation by reference of the following standards:

Standards

1. 40 CFR, Part, 136, Appendix B.
2. 40 CFR, Part, 141.
3. EPA publication "Manual for the Certification of Laboratories Analyzing Drinking Water," EPA815-R-05-004, fifth edition, EPA, Office of Ground Water and Drinking Water, January 2005.
4. EPA publication "Supplement 1 to the Fifth Edition of the Manual for the Certification of Laboratories Analyzing Drinking Water," EPA 815-F-08-006 EPA, Office of Ground Water and Drinking Water, June 2008.
5. EPA publication, Test Methods for Evaluating Solid Waste: Physical/Chemical Methods, also known as SW-846.
6. The State of Wisconsin Aquatic Life Toxicity Testing Methods Manual (Methods Manual).

Consent was received on May 9, 2018. Consent to Incorporation is attached.

SECTION 1. NR 106.145 (10) (a) is amended to read

In this subsection, "method blank", "~~matrix spike~~" and "limit of detection" have the meanings specified in s. NR 149.03.

Note: "Matrix spike" has the meaning specified in EPA Method 1631, Revision E: Mercury in Water by Oxidation, Purge and Trap, and Cold Vapor Atomic Fluorescence Spectrometry, August 2002, Office of Water, EPA-821-R-02-019

~~Note: "Method blank" is not defined as a subset of the definition of "Blank" in s. NR 149.03 (15).~~

SECTION 2. NR 106.145 (10) (d) is amended to read

The laboratory performing the analyses shall be certified for mercury under the cold vapor atomic fluorescence spectrophotometry technology of ch. NR 149 s. NR 149.42 for low level mercury analyses. ~~Until low level mercury certification is available, the lab shall be certified under ch. NR 149 for mercury and recognized by the department as having demonstrated its low level mercury capabilities under the emerging technology provision contained in s. NR 149.42.~~

~~Note: With the changes to ch. NR 149, effective 9-1-08, certification for low level mercury is now available. Certification for low level mercury under the emerging technology provision is no longer necessary or available.~~

SECTION 3. NR 106.145 (11) is amended to read

DATA REJECTION. The department may reject any sample results if data quality requirements specified in subs. (9) and (10) are not met or if results are produced by a laboratory that is not in compliance with ~~certification~~ requirements specified in ch. NR 149.

SECTION 4. NR 140.05 (13) Note is amended to read

Note: The limit of quantitation is ~~10/3 or 3.333 times the limit of detection~~ established as defined under s NR 149.48 (3).

SECTION 5. NR 140.16 (1) (d) Note is amended to read

Note: Refer to s. NR ~~149.46~~ 149.442 for sample preservation procedures and holding times.

SECTION 6. Chapter NR 149 is repealed and recreated to read:

Chapter NR 149

LABORATORY ACCREDITATION

TABLE OF CONTENTS. CHAPTER NR 149 IS REPEALED AND RECREATED TO READ:

Subchapter I — General Provisions

- NR 149.01 Purpose.
- NR 149.02 Applicability.
- NR 149.03 Definitions.

Subchapter II — Program Administration

- NR 149.05 Required accreditation.
- NR 149.06 Certificates.
- NR 149.08 Acceptance of other accreditations, licenses, or approvals.
- NR 149.09 Certification standards review council.
- NR 149.10 Enforcement.
- NR 149.11 Discretionary acceptance.
- NR 149.12 Variances.

Subchapter III — Program Structure

- NR 149.13 Fields of accreditation: certification and registration.

Subchapter IV — Accreditation Process

- NR 149.14 Application for accreditation.
- NR 149.15 Period, renewal, and expiration of accreditation.
- NR 149.155 Required notifications.
- NR 149.18 Subcontracting.
- NR 149.19 Requirements for certification in the drinking water matrix.
- NR 149.20 Requirements for whole effluent toxicity testing.
- NR 149.21 Fees.

Subchapter V — Proficiency Testing

- NR 149.22 Required proficiency testing samples and frequency of analysis.
- NR 149.23 Approval of proficiency testing sample providers.
- NR 149.24 Proficiency testing samples for applications and accreditation renewal.
- NR 149.25 Treatment of proficiency testing samples.
- NR 149.26 Reporting proficiency testing sample results.
- NR 149.27 Proficiency testing sample acceptance limits and grading.
- NR 149.28 Procedure for correcting unacceptable proficiency testing sample results.

Subchapter VI — On-Site Laboratory Evaluations

- NR 149.29 Purpose, type, and frequency.
- NR 149.30 Evaluation procedures and appraisal.
- NR 149.31 Evaluation reports.
- NR 149.32 Evaluation corrective action.
- NR 149.33 Conflicts of interest.
- NR 149.34 Evaluator qualifications.

Subchapter VII — Quality Systems

- NR 149.35 General requirements.
- NR 149.36 Laboratory personnel.
- NR 149.365 Laboratory ethics.
- NR 149.37 Quality manual.
- NR 149.38 Corrective action.
- NR 149.39 Records and documents.
- NR 149.40 Standard operating procedures.
- NR 149.41 Method selection.
- NR 149.42 Alternative methods.
- NR 149.43 Laboratory facilities.

- NR 149.44 Laboratory equipment.
- NR 149.442 Handling of samples.
- NR 149.444 Initial instrument calibration requirements.
- NR 149.446 Continuing instrument calibration requirements.
- NR 149.45 Measurement traceability.
- NR 149.47 Reporting results.
- NR 149.48 Quality control requirements for chemical testing.
- NR 149.50 Technology requirements.

Note: Chapter NR 149 as it existed on July 2014, was repealed and a new chapter NR 149 was created, Register [XX] No. [XX], effective [DATE].

SECTION 1. Chapter NR 149 is repealed and recreated to read:

SUBCHAPTER I - GENERAL PROVISIONS

NR 149.01 Purpose. The purpose of this chapter is to establish a program for the accreditation of laboratories performing testing under s. 299.11, Stats.

NR 149.02 Applicability. (1) This chapter specifies requirements for the administration of the laboratory accreditation program by the department.

(2) Unless otherwise exempted in this section, this chapter applies to all the following:

(a) Laboratories applying for accreditation.

(b) Laboratories holding an accreditation.

(c) Laboratories submitting data to the department for a covered program.

(d) Laboratories generating data that is necessary for the department to determine compliance with a covered program.

Note: Administrative codes and covered programs requiring analyses to be performed by an accredited laboratory are chs. NR 101 – Reports And Fees For Wastewater Discharges, 102 – Water Quality Standards For Wisconsin Surface Waters, 106 – Procedures For Calculating Water Quality Based Effluent Limitations For Point Source Discharges To Surface Waters, 110 – Sewerage Systems, 123 – Well Compensation Program, 131 – Nonferrous Metallic Mineral Prospecting, 132 – Nonferrous Metallic Mineral Mining, 140 – Groundwater Quality, 150 – Environmental Analysis And Review Procedures, 153 – Targeted Runoff Management And Notice Of Discharge Grant Programs, 155 – Urban Nonpoint Source Water Pollution Abatement And Storm Water Management Grant Program, 157 – Management Of Pcb's And Products Containing Pcb's, 182 – Nonferrous Metallic Mineral Mining Wastes, 200 – Application For Discharge Permits And Water Quality Standards Variances, 204 – Domestic Sewage Sludge Management, 205 – General Provisions, 206 – Land Disposal Of Municipal And Domestic Wastewaters, 210 – Sewage Treatment Works, 211 – General Pretreatment Requirements, 212 – Waste Load Allocated Water Quality Related Effluent Limitations, 214 – Land Treatment Of Industrial Liquid Wastes, By-Product Solids And Sludges, 216 – Storm Water Discharge Permits, 217 – Effluent Standards And Limitations For Phosphorus, 218 – Method And Manner Of Sampling, 219 – Analytical Test Methods And Procedures, 230 – Inorganic Chemicals Manufacturing, 233 – Pesticide Chemicals, 243 – Animal Feeding Operations, 254 – Iron And Steel Manufacturing, 256 – Metal Molding And Casting, 260 – Electroplating, 261 – Metal Finishing, 263 – Coil Coating, 273 – Nonferrous Metals Forming And Metal Powders, 274 – Nonferrous Metals Manufacturing, 290 – Steam Electric Power Generating, 347 – Sediment Sampling And Analysis, Monitoring Protocol And Disposal Criteria For Dredging Projects, 500 – General Solid Waste Management Requirements, 507 – Environmental Monitoring For Landfills, 518 – Landspreading Of Solid Waste, 528 – Management Of Accumulated Sediment From Storm Water Management Structures, 538 – Beneficial Use Of Industrial Byproducts, 662 – Hazardous Waste Generator Standards, 664 – Hazardous Waste Treatment, Storage And Disposal Facility Standards, 665 – Interim License Hazardous Waste Treatment, Storage And Disposal Facility Standards, 700 – General Requirements, 712 – Personnel Qualifications For Conducting Environmental Response Actions, 716 – Site Investigations, 720 – Soil Cleanup Standards, 738 – Temporary Emergency Water Supplies, 747 – Petroleum Environmental Cleanup Fund, 809 – Safe Drinking Water, 810 – Requirements For The Operation And Maintenance Of Public Water Systems, 811 – Requirements For The Operation And Design Of Community Water Systems, 812 – Well Construction And Pump Installation and 845 – County Administration Of NR 812 Private Wells Code.

Note: Links to the codes specified above can be found on the Wisconsin department of natural resources laboratory accreditation program website.

(3) Laboratories performing analyses for the safe drinking water program under ch. NR 809 or for the well construction and pump installation testing program under ch. NR 812 shall be certified; registration is not available

for these analyses. Additional requirements for laboratories performing compliance analysis under ch. NR 809 are specified in s. NR 149.19.

(4) Laboratories performing analysis for whole effluent toxicity testing shall meet the requirements specified in s. NR 149.20.

(5) This chapter applies to laboratories analyzing industrial pre-treatment samples when the department is the control authority of a pre-treatment ordinance or when another control authority requires it.

(6) Laboratories performing asbestos or radiological testing for a covered program shall be certified or approved by the EPA or the department.

Note: Laboratories performing bacteriological testing for a covered program are certified or approved under ch. ATCP 77 by the department of agriculture, trade, and consumer protection.

(7) This chapter establishes compliance requirements that shall be incorporated into the quality systems of all laboratories accredited by the department.

(a) Laboratories shall meet any requirements pertaining to analyses and analytical operations contained in the methods, regulations, or covered programs when those requirements are more stringent than the ones specified in this chapter, unless this chapter grants explicit, alternative allowances.

Note: Sources, including the following as updated, likely contain methods that are acceptable for testing under this chapter: The EPA, the department, Standard Methods for the Examination of Water and Wastewater, Test Methods for Evaluating Solid Waste, Physical/Chemical Methods – SW-846, American Society for Testing and Materials, and the U.S. Geological Survey Agency.

(b) When it is not apparent to the laboratory whether the minimum requirements of this chapter or those specified in the methods, regulations, or covered programs are more stringent, the department shall communicate which requirements are more stringent to the laboratories and the laboratories shall meet those requirements.

(c) When a laboratory incorporates a procedure that is neither expressly permitted nor prohibited by the method, the department will assess the scientific validity of the procedure to determine if the procedure is within the scope of the method. The underlying chemistry of the method shall remain unchanged. The department may seek the advice of the council in making determinations under this paragraph.

Note: For example, if a digestion time of 30 minutes is required, the laboratory is not to use less time for digestion. A second example, when a minimum volume of solvent is required to extract a specific amount of sample, then less solvent is not to be used unless the ratio of extraction solvent to sample amount is maintained.

NR 149.03 Definitions. In this chapter:

(1) “Acceptance limits” means limits established by the department that are used to determine if a laboratory has analyzed a proficiency testing sample successfully.

Note: Acceptance limits are specified in s. NR 149.27.

(2) “Accreditation” and “accredited” mean that the department has determined that an organization is competent to perform specific types of tests. “Accreditation” and “accredited” include “certification” and “registration.”

(3) “Accreditation matrix” means a matrix type that is part of the first tier of a field of accreditation under s. NR 149.13 (2). Accreditation matrices are drinking water, aqueous, and non-aqueous matrices.

(4) “Accuracy” means the closeness of a measured value to an accepted reference value or standard.

- (5) “Algorithm” means a process or set of rules to be followed in calculations for solving a problem.
- (6) “Analysis day” means the day in which a specific type of analysis is performed.
- (7) “Analyte” means the chemical substance, physical property, or organism analyzed in a sample.
- (8) “Analyte group” means a set of analytes that can be determined using the same method or technology and that constitute a unit, acknowledged by the department, of the third tier of accreditation under s. NR 149.13 (4).
- (9) “Analytical balance” means a balance that is capable of measuring masses to within 0.0001 g.
- (10) “Analytical class” means a set of analytes or analyte groups of similar behavior or composition, or a set of analytes or analyte groups regulated under the same provisions of the federal safe drinking water act, that is used to organize the third tier of accreditation under s. NR 149.13 (4).
- (11) “Analytical instrument” means any test instrument used to provide analytical results that is not support equipment.
- (12) “Analytical staff” means staff that includes laboratory directors, supervisory personnel, quality assurance personnel, technicians, chemists, biologists, preparation analysts, and instrument analysts.
- (13) “Aqueous” means an accreditation matrix that is water, is not drinking water, and can be reported in units of mass per volume.
- Note:** Leachates are not accredited under the aqueous matrix.
- (14) “Batch” means a set of environmental samples prepared or analyzed together using the same process, personnel, and lots of reagents.
- (a) A “preparation batch” means a set of one to 20 environmental samples of the same accreditation matrix, meeting batch criteria, and with a maximum time of 24 hours between the start of processing of the first and last sample in the batch.
- (b) An “analytical batch” means a set of environmental samples which are analyzed together as a group in an uninterrupted sequence.
- (15) “Bias” means the consistent deviation of measured values from a true value caused by systematic errors in a procedure or a measurement process.
- (16) “Calibration” means the process used to establish an observed relationship between the response of an analytical instrument and a known amount of analyte, or the process used to determine, by measuring or comparison with a reference standard, the correct value of each scale reading in an instrument, meter, or measuring device.
- (17) “Calibration blank” means an aliquot that consists of the same matrix as that used for the calibration standards, but without the analytes.
- (18) “Calibration function” means the specific mathematical relationship established to relate calibration standards to instrument response.

(19) “Calibration model” means an algorithm that is used to determine an average calibration factor, average response factor, linear regression, or non-linear regression.

(20) “Certificate” means a document owned by the department and issued to a laboratory that indicates the fields of accreditation granted to a laboratory.

(21) “Certification” or “certified” means certification, under s. 299.11 (7), Stats., of laboratories that perform compliance analyses for hire or to laboratories that perform compliance drinking water analyses in accordance with the standards and requirements of this chapter.

(22) “Coefficient of determination” means a quantity that measures the degree of agreement between the points in a calibration and the function derived to connect the points.

(23) “Confirm” means to verify the identity of a compound by an alternative procedure, column, detector, wavelength, or by a technology that bases detection on a different scientific principle from the one originally used for identifying the compound.

(24) “Continuing calibration blank” or “CCB” means an aliquot that consists of the same matrix as that used for the calibration standards, but without the analytes, analyzed during an analysis sequence to verify the continued absence of instrumental interferences.

(25) “Continuing calibration verification standard” or “CCV standard” means a standard of known concentration of analyte used to assure continued calibration accuracy during an analysis sequence.

(26) “Correlation coefficient” means a quantity that measures the degree of agreement between the points in a calibration curve and the linear function derived to connect the points.

(27) “Corrective action” means any measure taken to eliminate or prevent the recurrence of the causes of an existing nonconformity, defect, or undesirable condition.

(28) “Council” means the certification standards review council created under s. 15.107 (12), Stats.

(29) “Covered program” means a program listed or enumerated in s. 299.11 (1) (d) 1. to 9., Stats., and includes any department program, project, permit, contract, or site investigation that requires analytical work to be performed by an accredited laboratory.

Note: The note in s. NR 149.02 (2) (d) provides a list of department administrative rules of covered programs requiring accreditation under this chapter.

(30) “Deficiency” means a documented or verifiable deviation from the requirements of this chapter that is noted during an on-site evaluation or while reviewing analytical data produced by a laboratory.

(31) “Department” means the department of natural resources.

(32) “EPA” means the United States environmental protection agency.

(33) “Field of accreditation” means a 3-tiered unit by which the department uses to grant laboratories accreditation as specified under s. NR 149.13.

(34) “For hire” means offering analyses for payment or non-monetary compensation.

(35) “Initial calibration blank” or “ICB” means an aliquot that consists of the same matrix as that used for the calibration standards, but without the analytes, analyzed following the initial calibration and prior to quantitating any samples to verify the absence of instrumental interferences.

(36) “Initial calibration verification standard” or “ICV standard” means a standard of known concentration, prepared using second source standards, analyzed following the initial calibration and prior to quantitating any samples to assure initial calibration accuracy.

(37) “Internal standard” means a known concentration of standard added to a sample or quality control sample as a reference for evaluating and controlling the precision and bias of the analytical method.

(38) “Laboratory” means a facility that performs tests in connection with a covered program that requires data from an accredited laboratory. A facility consisting of a principal laboratory and annexes within 5 miles of the principal laboratory may be considered a single laboratory. When the terms laboratory or laboratories are used unmodified in this chapter, the terms include laboratories accredited under this chapter and those seeking accreditation under this chapter.

(39) “Laboratory control sample” or “LCS” means a sample of a matrix without the analytes of interest or a matrix with a consistent concentration of the analytes of interest, fortified with a verified known amount of the analytes of interest. The purpose of an LCS is to determine whether the methodology is in control and whether the laboratory can make accurate and precise measurements.

Note: In many EPA methods, the term “lab-fortified blank” is equivalent to an LCS.

(40) “Laboratory equipment” means any support equipment or analytical instrument necessary to or involved in generating the results of an analysis.

(41) “Limit of detection” or “LOD” means the lowest concentration or amount of analyte that can be identified, measured, and reported with confidence that the concentration is not a false positive value. The department considers the LOD to be equivalent to the method detection limit and is determined under the method cited in sub. (46).

(42) “Limit of quantitation” or “LOQ” means the lowest concentration or amount of an analyte for which quantitative results can be obtained.

(43) “Maximum contaminant level” or “MCL” means the maximum permissible level of a contaminant in water that is delivered to any user of a public water system.

(44) “Method” means a procedure used for measuring the presence and concentration of physical and chemical pollutants.

(45) “Method blank” means a clean matrix that is treated and processed exactly as a sample including exposure to all glassware, equipment, solvents, reagents, internal standards, and surrogates to measure artifacts in the measurement process.

Note: In many EPA methods, the term “laboratory reagent blank” is equivalent to a method blank.

(46) “Method detection limit” or “MDL” means the minimum measured concentration of a substance that can be reported with 99% confidence that the measured concentration is distinguishable from method blank results. The MDL is generated according to the procedure specified in the latest revision of 40 CFR Part 136, Appendix B.

Note: Links to the 40 CFR Part 136, Appendix B can be found on the Wisconsin department of natural resources laboratory accreditation program website.

- (47) “NIST” means the National Institute for Standards and Technology.
- (48) “Non-aqueous” includes all matrices that are not drinking water or aqueous. It includes soils, sediments, sludges, organic liquids, oils, solid waste and multi-phasic wastes. Leachates are accredited under the non-aqueous matrix.
- (49) “Nonconformance” means a documented or verifiable deviation from the requirements of this chapter or a deviation from the requirements of a quality system.
- (50) “On-site evaluation” means an assessment conducted by the department at a laboratory seeking or maintaining accreditations to determine actual or potential compliance with the requirements of this chapter.
- (51) “Ownership” means owning or controlling, directly or indirectly, a laboratory facility through an equity interest, or its equivalent, of 10% or more.
- (52) “Pesticide” means a chemical substance defined in s. 94.67 (25) and (25m), Stats., an isomer of a pesticide or a degradation product or metabolic product of a pesticide.
- (53) “Precision” means the degree to which a set of observations or measurements of the same property, obtained under similar conditions, conform to themselves. Precision is usually expressed as the standard deviation, variance, or range, in either absolute or relative terms.
- (54) “Proficiency testing sample” or “PT sample” means a sample obtained from an approved proficiency testing sample provider to evaluate the ability of a laboratory to produce an analytical test result meeting the definition of acceptable performance outlined in s. NR 149.27. The concentration of the analyte in the sample is unknown to the laboratory at the time of analysis.
- (55) “Qualify” means a written statement accompanying or referencing test results to identify anomalies or deviations from the requirements in this chapter that were encountered in generating the results.
- (56) “Quality assurance” means an integrated system of activities involving planning, control, assessment, reporting, and improvement to ensure that a product or service meets defined standards of quality.
- (57) “Quality control” means the overall system of technical activities designed to measure and control the quality of a product or service that meets the stated needs of users.
- (58) “Quality control limit” means the acceptance criteria used to evaluate quality control samples. Quality control limits may be those published by the department, referenced in a method or calculated by a laboratory. In this chapter, quality control limits calculated by a laboratory will be referred to as generated in-house control limits.
- (59) “Quality system” means a structured and documented management arrangement describing the policies, objectives, principles, organizational authority, responsibilities, accountability, and implementation plan of an organization for ensuring quality in its work processes, products, and services.
- (60) “Raw data” means any original information from a measurement activity or study recorded in media that allows the reconstruction and evaluation of the activity or study. Raw data include absorbance, emission counts, area counts, peak heights, abundance, and millivolts. Raw data may be stored in hard copy or electronically.
- (61) “Reagent water” means water which has been treated to remove any impurities that may affect the quality of an analysis.

(62) “Reference material” means a material that has one or more sufficiently well-established properties that can be used for calibrating or verifying the calibration of support equipment or analytical instruments.

(63) “Registration” or “registered” means registration under s. 299.11 (8), Stats., of laboratories that perform tests solely on its own behalf or that of a subsidiary under common ownership or control in accordance with the standards and requirements of this chapter. Registered laboratories do not perform drinking water testing.

(64) “Relative standard error” or “RSE” means the standard error divided by the mean for a set of calibration data and expressed as a percentage. The RSE is calculated according to the following formula:

$$RSE = 100 \times \sqrt{\sum_{i=1}^n \left[\frac{x'_i - x_i}{x_i} \right]^2 / (n - p)}$$

x_i = true amount of analyte in calibration level i , in mass or concentration units.

x'_i = measured amount of analyte in calibration level i , in mass or concentration units.

p = number of terms in the fitting equation.

Note: average = 1, linear = 2, quadratic = 3, cubic = 4

n = number of calibration points.

(65) “RV” means relative value.

(66) “Relocation” means a move by a laboratory resulting in a change in the laboratory’s physical address.

(67) “Replicate” means two or more substantially equal aliquots analyzed independently for the same analyte.

(68) “Residual” means the difference, expressed as a percent, between the theoretical concentration of a calibration standard and the value derived from the calibration function from the measured response of the calibration standard.

(69) “Result” means the quantitative or qualitative output of an analysis, including measurements, determinations, and information obtained or derived from tests.

(70) “Revocation” means cancellation of a laboratory’s accreditation.

(71) “Second source standard” means a standard procured from a supplier or manufacturer different from the supplier or manufacturer of a laboratory’s calibration standards, or a standard obtained from the same supplier or manufacturer of a laboratory’s calibration standards from a lot verifiably different from the lot of the calibration standards.

(72) “Sensitivity” means the capability of a method or instrument to discriminate between measurement responses representing different levels of analyte, or the capability of a method or instrument to detect an analyte at or greater than a stated quantity.

(73) “Shall” means a mandatory requirement.

(74) “Subcontract” means the act of procuring analytical services from a certified laboratory.

Note: Registered laboratories only do testing for their own facility. Another facility, under the same ownership, can procure analytical services from a registered laboratory.

(75) “Support equipment” means devices that may not be analytical instruments, but that are necessary to support laboratory tests and operations. “Support equipment” includes autoclaves, balances, ovens, refrigerators, freezers, incubators, water baths, temperature measuring devices, sample preparation devices, and volumetric dispensing devices when quantitative results depend on the accuracy of the support equipment.

(76) “Surrogate” means a substance unlikely to be found in environmental samples, with properties similar to those of analytes of interest, which is used to evaluate the bias of an analysis in the fortified sample.

(77) “Suspension” means a temporary cancellation of a laboratory’s certification.

(78) “Temperature blank” means a sample container, of at least 40 mL capacity, filled with water and transported with each shipment of collected samples to determine the temperature of other samples in the shipment on arrival at a laboratory.

(79) “Test” means any chemical, biological, physical, radiological, or microscopic assay, examination, or analysis conducted by a laboratory on water, wastewater, groundwater, a biosolid, a waste material, a hazardous substance, or any other matrix analyzed to determine compliance with a covered program.

(80) “X-intercept” means the point at which the plot of the calibration function crosses the x-axis.

SUBCHAPTER II - PROGRAM ADMINISTRATION

NR 149.05 Required accreditation. (1) All laboratories submitting data to the department for a covered program or generating data to determine compliance with a covered program, shall be accredited under this chapter for the fields of accreditation corresponding to the submitted or generated data, unless this chapter or a covered program exempts a test from requiring accreditation.

(2) The department may not accept data from a laboratory that is not properly accredited under this chapter, except as provided in s. NR 149.11.

(3) The department may initiate enforcement action against a laboratory which maintains accreditations under this chapter but performs testing for analytes for which it does not hold the appropriate accreditation and for which the department offers accreditation when that data will be reported to the department.

(4) A laboratory may not transfer its accreditation to any other entity unless the department expressly approves the transfer by the process specified in s. NR 149.14 (1) (a).

NR 149.06 Certificates. (1) The department shall issue certificates to accredited laboratories indicating or referring to the specific fields of accreditation for which laboratories have been granted accreditation. The department shall issue certificates annually, whenever the fields for which a laboratory is accredited change, and when a laboratory relocates or changes its name.

(2) The department shall issue certificates to the owner or legally responsible party of a laboratory.

(a) The department may not issue certificates to anyone who is not the owner or legally responsible party of a laboratory.

(b) The department may indicate in a certificate that a laboratory is managed by an outside contractor.

(3) Certificates are the property of the department and shall be returned to the department upon request.

(4) A laboratory may not alter or modify certificates issued by the department. A laboratory that alters or modifies a certificate, or that misrepresents the fields of accreditation contained or referenced in a certificate, may be subject to revocation of all its accreditations.

NR 149.08 Acceptance of other accreditations, licenses, or approvals. (1) AGRICULTURE, TRADE, AND CONSUMER PROTECTION AGREEMENT. The department shall accept the accreditation, licensure, or approval by the department of agriculture, trade, and consumer protection for microbiological testing performed by a laboratory submitting or generating data for a covered program.

(2) LABORATORIES ACCREDITED, LICENSED, OR APPROVED BY OTHER GOVERNMENTS. (a) The department may negotiate with and attempt to enter into agreements with federal agencies and agencies of other states to reciprocally accept accreditations of laboratories under this chapter.

(b) The department may accept the accreditation, licensure, or approval of a laboratory by another state or an agency of the federal government if the standards used for the qualification of a laboratory are substantially equivalent to those established in this chapter.

(c) The department may not accept the accreditation, licensure, or approval of a laboratory by another state or an agency of the federal government, unless that state or federal agency accepts laboratories accredited under this chapter.

(3) PRIVATE ORGANIZATION AGREEMENTS. (a) The department may enter into agreements with private not-for-profit organizations to accept accreditation of laboratories under this chapter.

(b) The department may accept the accreditation, licensure, or approval of a laboratory by a private not-for-profit organization if the organization's standards used for the qualification of a laboratory are substantially equivalent to those established in this chapter.

NR 149.09 Certification standards review council. (1) The council shall advise the department on the standards used to certify, register, suspend, and revoke laboratories.

(2) The council shall advise the department on training and outreach activities that the department may offer or sponsor to facilitate compliance of laboratories with this chapter.

(3) The council shall advise the department on the frequency and scope of evaluations necessary to determine compliance of laboratories with this chapter.

(4) The department shall seek the advice of the council before requiring the analysis of additional PT samples and approving PT sample providers.

(5) The department shall seek the advice of the council before implementing changes in the fees assessed to laboratories.

(6) The department shall seek the advice of the council in granting variances.

(7) The department shall prepare annually the following for review by the council:

(a) A summary of laboratory evaluations performed.

(b) A list of required PT samples and available PT sample providers.

(c) A summary of fees scheduled to be assessed to laboratories.

(d) A summary of variances issued.

NR 149.10 Enforcement. (1) ADMINISTRATIVE PROCEDURES. A laboratory's accreditation is valid until it expires, is suspended, or is revoked. If, after opportunity for a contested case hearing, the department finds that an accredited laboratory materially and consistently failed to comply with the provisions of this chapter, the department may suspend or revoke a laboratory's accreditation in whole or in part by matrix, analytical technology, method, analyte, or analyte group. Contested case hearings for out-of-state laboratories regulated under this chapter shall be held in Madison, Wisconsin.

(2) SUSPENSION OR REVOCATION OF CERTIFIED LABORATORIES. (a) Causes for suspension of certification include any of the following:

1. Material and consistent failure to comply with the requirements of this chapter.
2. Reporting data to the department after a laboratory is deemed temporarily incapable of performing analysis in any matrix, analytical technology, method, analyte, or analyte group.
3. Suspension of certification, accreditation, license, or approval by another state or agency of the federal government for which the laboratory holds certification if the grounds for suspension are substantially equivalent to any of those listed in this paragraph.

(b) Causes for revocation of certification include any of the following:

1. Material and consistent failure to maintain records as required in this chapter.
2. Failure to allow the department to perform on-site evaluations as specified in subch. VI.
3. Material and consistent failure to comply with the requirements of this chapter.
4. Material and consistent failure to submit requested records to the department.
5. Material and consistent failure to follow specified procedural or quality control requirements prescribed in methods.
6. Falsification of analytical results, testing dates, or any other information submitted to the department by the laboratory. Falsification includes alteration or modification of a certificate.
7. Failure of two consecutive PT samples for any method and analyte or analyte group combination for laboratories holding certification in the drinking water matrix.
8. Revocation of certification, registration, accreditation, license, or approval by another state or agency of the federal government for which the laboratory holds certification if the grounds for revocation are substantially equivalent to any of those listed in this paragraph.

(3) REVOCATION OF REGISTERED LABORATORIES. Causes for revocation of registration include any of the following:

- (a) Material and consistent failure to maintain records as required in this chapter.
- (b) Failure to allow the department to perform on-site evaluations as specified in subch. VI.
- (c) Material and consistent failure to comply with the requirements of this chapter.
- (d) Material and consistent failure to submit requested records to the department.
- (e) Material and consistent failure to follow specified procedural or quality control requirements prescribed in approved methods.
- (f) Falsification of analytical results, testing dates, or any other information submitted to the department by the laboratory. Falsification includes alteration or modification of a certificate.

(4) PROCEDURE FOR SUSPENSION OR REVOCATION OF ACCREDITATION. (a) An order suspending or revoking accreditation shall be mailed to the laboratory and shall state the reasons for suspension or revocation. The order shall include the conditions under which reapplication will be accepted. For orders suspending accreditation, the order may include a timetable for correcting the deficiencies that led to the suspension. For orders revoking accreditation, the department may set a time for the revocation.

(b) An order suspending or revoking an accreditation shall take effect on the 30th day after the order is mailed unless the owner of an accredited laboratory submits a petition for a hearing. Petitions for a hearing shall be submitted to the department within 30 days of receiving the order. The petition for hearing shall specify the findings or conclusions, or both, that the laboratory disputes and conform to the requirements of s. NR 2.05 (5).

(c) If a request for a hearing is submitted and meets the requirements of s. 227.42, Stats., the suspension or revocation shall be stayed, and the department shall conduct a contested case hearing on the matter. At least ten days prior to the date of the hearing, the department shall send a written notice to the laboratory indicating the date, time, and location of the hearing. The final determination of the department, including the basis for the decision, shall be provided by written order to the laboratory after the hearing.

Note: Refer to ch. NR 2 for additional information on the contested hearing process.

(d) The final determination of the department is subject to review under ch. 227, Stats.

(5) REAPPLICATION FOLLOWING SUSPENSION OR REVOCATION. (a) A laboratory that has had its accreditation suspended may reapply for accreditation if all the following are met:

1. The deficiencies that led to the suspension have been corrected in accordance with the timetable contained in the order.

2. Any conditions for reapplication specified in the order have been met.

(b) A laboratory that has had its accreditation revoked may reapply for accreditation if all of the following have been met:

1. The deficiencies that led to the revocation have been corrected.
2. Conditions contained in the order have been satisfied.
3. The time for which the revocation is in effect has expired.

(c) Laboratories reapplying for accreditation following suspension or revocation shall submit an initial application as identified in s. NR 149.14 (1) and (2).

(6) REFERRAL. (a) Any violation of this chapter may be referred to the attorney general's office for enforcement under ss. 299.95 and 299.97, Stats.

(b) Any laboratory operating without proper accreditation for which analysis results are submitted to the department for compliance monitoring or for analyses that require certification or registration under any covered program may be referred by the department to the attorney general's office for enforcement.

NR 149.11 Discretionary acceptance. (1) Except for results of tests required under ch. NR 809 the department may accept, on a case-by-case basis, the results of tests originating in a laboratory not accredited as required by a covered program if the results meet all other requirements of this chapter.

(2) The requirements of this chapter may be waived by the department when there is a multi-agency response to a hazardous substance discharged in boundary areas of the state.

(3) The requirements of this chapter may be waived by the department when the environmental protection agency national enforcement investigations center laboratory is utilized for EPA or department led enforcement cases.

(4) The department may not accept the results of tests originating in a laboratory not accredited, unless the results are generated in accordance with requirements substantially equivalent to those outlined in this chapter.

Note: Refer to s. NR 149.42 for additional information on the use of alternative methods.

(5) The department may charge a fee under s. 299.11 (5) (d), Stats., if it is necessary to verify the results of tests for which a laboratory requests discretionary acceptance.

NR 149.12 Variances. (1) **GENERAL.** The department may approve variances from non-statutory requirements of this chapter when the department determines that the variances have no effect on the department's objectives. Before granting variances, the department shall consider factors such as good cause, circumstances beyond the control of the laboratory, and financial hardship.

(2) **REQUEST FOR VARIANCE.** Requests for variances shall be submitted to the department. Each variance request shall contain all the following:

- (a) The name of the applicant or laboratory.
- (b) The section of this chapter from which a variance is sought.

(c) A description of the circumstances under which the variance will be exercised, including any pertinent background information relevant to making a justification.

(3) APPROVAL OF VARIANCE. The department shall approve or deny the requested variance to the applicant within 60 days of receiving all the information referenced in sub. (2). If the request is denied, the department shall state the reasons for the denial.

(4) REPEAL OF VARIANCES. The department will annually review approved variances and may repeal those where the initial justification for the variance no longer applies. Once the department notifies the laboratory of the repeal, the laboratory will have six months before the repeal is effective.

SUBCHAPTER III - PROGRAM STRUCTURE

NR 149.13 Fields of accreditation: certification and registration. (1) GENERAL. The department shall certify and register laboratories by specific fields of accreditation. Accreditation shall be by certification under s. 299.11 (7), Stats., or registration under s. 299.11 (8), Stats. Fields of accreditation consist of 3 tiers describing the analytical capability of laboratories.

(2) TIER 1 – MATRIX. The first tier of accreditation is comprised of aqueous, non-aqueous, and drinking water matrices.

Note: Biosolids and sludges are a non-aqueous matrix for accreditation purposes.

(3) TIER 2 – TECHNOLOGY OR METHOD. The second tier of accreditation is comprised of analytical technologies for the aqueous and non-aqueous matrices or methods for the drinking water matrix.

(a) The department may certify or register laboratories that analyze aqueous and non-aqueous matrices for the analytical technologies contained in this section, Table 1.

(b) The department shall include any associated sample preparation techniques, such as digestions, distillations, extractions, cleanups, concentrations, and dilutions as part of the certification or registration for a given field of accreditation.

(c) Laboratories may employ multiple methods of analysis for a given analytical technology under the same field of accreditation.

Table 1 - Analytical Technologies for Aqueous and Non-Aqueous Matrices

Number	Analytical Technology
General Chemistry	
1.	Oxygen Demand Assays (BOD or cBOD) ¹
2.	Colorimetric or Turbidimetric
3.	Electrometric Assays (i.e. ion-selective electrode)
4.	Gravimetric Assays – Residue (solids)
5.	Extraction/Gravimetric Assays – Oil & Grease as Hexane Extractable Materials (HEM) ¹
6.	Titrimetric or Potentiometric Titration Assays
7.	Flow Injection-Gas Diffusion–Amperometry ¹
8.	Nondispersive Infrared (NDIR) or Microcoulometry
9.	Ion Chromatography (IC)
Metals	

10.	Flame Atomic Absorption Spectrophotometry (FLAA)
11.	Flame Photometry Spectrophotometry (FP)
12.	Gaseous Hydride Atomic Absorption Spectrophotometry (GHAA)
13.	Graphite Furnace Atomic Absorption Spectrophotometry (GFAA)
14.	Cold Vapor Atomic Absorption Spectrophotometry (CVAA)
15.	Cold Vapor Atomic Fluorescence Spectrophotometry (CVAFS)
16.	Thermal Decomposition Atomic Absorption Spectrophotometry (TDAA)
17.	Inductively Coupled Plasma Emission Spectrophotometry (ICP)
18.	Inductively Coupled Plasma-Mass Spectrometry (ICP/MS)
Organics	
19.	Gas Chromatography (GC)
20.	Gas Chromatography-Mass Spectrometry (GC/MS)
21.	Liquid Chromatography (LC)
22.	Liquid Chromatography-Mass Spectrometry (LC/MS)
23.	High Resolution Gas Chromatography-Mass Spectrometry (HRGC/MS)
Other	
24.	Hazardous Waste Characteristics ²
25.	Solid Waste Leaching Procedures ²
26.	Whole Effluent Toxicity Assays ¹
27.	Other ³

1. Accreditation for this technology is only available for the aqueous matrix – Tier 1.

2. Accreditation for this technology is only available for the non-aqueous matrix – Tier 1.

3. The department may offer accreditation in other analytical technologies if the technology is approved by the EPA or is approved by the department as an emerging technology.

(d) The department may certify laboratories analyzing drinking water samples using methods promulgated or approved by the EPA under 40 CFR, Part 141.

Note: Links to the 40 CFR Part 141 can be found on the Wisconsin department of natural resources laboratory accreditation program website.

(4) TIER 3 – ANALYTES OR ANALYTE GROUP. The third tier of the accreditation is comprised of analytes or analyte groups. The department may certify or register laboratories by analyte groups if it improves the efficiency of administering accreditations.

(a) The analytes and analyte groups available for accreditation under this subsection are contained in Appendix I.

(b) The department, upon consultation with the council, may offer accreditation for additional analytes or analyte groups that are not contained in Appendix I upon request by the manager of a covered program or when the EPA requires the additional analytes or analyte group analysis.

(c) For aqueous and non-aqueous matrices, the analytes and analyte groups are organized into classes. Laboratories analyzing aqueous and non-aqueous matrices may be accredited for analytes and analyte groups belonging to the analytical classes contained in this section, Table 2.

Table 2 - Analytical Classes for Aqueous and Non-Aqueous Matrices

Number	Analytical Classes
1.	General Chemistry
2.	Metals
3.	Volatile Organic Compounds
4.	Base, Neutral, and Acid Extractable Semivolatile Compounds, including: a. Aldehydes and Ketones b. Benzidines c. Chlorinated Hydrocarbons d. Explosive Residues e. Haloethers f. Nitroaromatics g. Nitrosamines h. Non-halogenated Organics i. Phenols j. Phthalates k. Polynuclear Aromatic Hydrocarbons
5.	Polynuclear Aromatic Hydrocarbons
6.	Pesticides and Metabolites, including: a. Acid b. Nitrogen c. Carbamate d. Organochlorine e. Organophosphorus f. Triazine g. Other
7.	Persistent Organic Pollutants
8.	Hazardous Waste Characteristics ¹
9.	Leaching Procedures ^{1,2}
10.	Solvent Scans
11.	Toxicity, Acute
12.	Toxicity, Chronic

1. Hazardous Waste Characteristics and Leaching Procedures are only offered for non-aqueous matrices – Tier 1.

2. Leaching Procedures require that laboratories also maintain accreditation for any analyte to be determined in the resulting leachate in the non-aqueous matrix.

(d) For the drinking water matrix, the analytes and analyte groups are organized into classes. Laboratories analyzing drinking water may be certified for analytes or analyte groups belonging to the analytical classes contained in this section, Table 3.

Table 3 - Analytical Classes for the Drinking Water Matrix

Number	Analytical Classes
1.	Disinfection By-products
2.	Primary Inorganic Contaminants (Non-Metals)
3.	Primary Inorganic Contaminants (Metals)
4.	Secondary Contaminants (Non-Metals)
5.	Secondary Contaminants (Metals)
6.	Synthetic Organic Contaminants (SOC) – Dioxin
7.	Synthetic Organic Contaminants (SOC) – Organochlorine Pesticides
8.	Synthetic Organic Contaminants (SOC) – Nitrogen-Phosphorus Pesticides
9.	Synthetic Organic Contaminants (SOC) – Herbicides
10.	Synthetic Organic Contaminants (SOC) – Miscellaneous
11.	Volatile Organic Compounds (VOC)

SUBCHAPTER IV – Accreditation Process

NR 149.14 Application for accreditation. (1) GENERAL REQUIREMENTS. (a) Laboratories are required to do all the following:

1. Submit applications to seek, revise, or transfer accreditations.
2. Declare the fields of accreditation being sought, revised, or transferred in corresponding applications.
3. For drinking water, declare the methods of analysis for analytes and analyte groups in the fields of accreditation being sought, revised, or transferred.
4. Submit a current analytical instrument list.
5. Submit acceptable results for PT samples when the department requires the PT sample analysis.
6. For laboratories that are not physically located in Wisconsin, submit a statement of intent to perform analyses for regulatory samples originating in Wisconsin. Intent to perform analyses for regulatory samples originating in Wisconsin can be manifested by any of the following:
 - a. Referencing the affiliation of the applicant laboratory with a plant, office, laboratory, or engineering firm physically located in the state of Wisconsin.
 - b. Submitting a letter from a potential client requesting the applicant to perform analyses to determine compliance with a covered program.
7. Submit any information identified in an application or upon request of the department such as standard operating procedures or analytical data.
8. When the department determines that an evaluation is necessary to determine potential or actual compliance with this chapter, allow the department to perform an on-site evaluation.
9. Remit any necessary fees required under this chapter.

Note: Fee information is contained in s. NR 149.21 Tables 1, 2, and 3.

10. Agree to comply with this chapter by signing the application.

(b) The department may not accept applications from a laboratory to which any of the following apply:

1. The laboratory has been issued a notice of violation for nonconformance with this chapter if the nonconformance has not been corrected and the notice of violation has not been closed.

2. The laboratory has been issued an administrative order of suspension or revocation for a violation of this chapter when the violation has not been closed and the suspension or revocation period specified in an order has not expired.

3. The laboratory was not in compliance with this chapter at the time the laboratory voluntarily relinquished its accreditations, nonconformances existing prior to relinquishing the accreditations have not been corrected, and at least 6 months have not elapsed since the voluntary action was undertaken.

(c) The department shall expire any application from a laboratory that has not submitted all the information and materials required as part of the application, or subsequent audit process, within a year of the receipt of the application form.

(d) The department may require the submittal of additional information necessary, such as standard operating procedures or analytical data, to determine a laboratory's actual or potential compliance with the provisions of this chapter.

(2) INITIAL APPLICATIONS. (a) A laboratory seeking direct accreditation by the department that has never been accredited under this chapter, that has let its entire accreditation lapse for more than one year, or that has voluntarily relinquished all its accreditations, shall submit an initial application to become accredited.

(b) A laboratory seeking reinstatement of its accreditations, following a suspension or revocation, shall submit an initial application for the desired accreditations.

(3) REVISED APPLICATIONS. (a) A laboratory holding valid accreditations shall submit a revised application to seek additional accreditations in any of the following:

1. Matrices.

2. Technologies for an accredited matrix.

3. Analytes or analyte groups within an accredited analytical technology.

4. Analyte-method or analyte group-method combinations for the drinking water matrix.

(b) A laboratory seeking reinstatement of accreditations within a year after failing to renew those accreditations shall submit a revised application for the desired accreditations.

(c) A laboratory seeking to convert a valid certification into a registration or a registration into a certification shall submit a revised application.

(4) APPLICATIONS FOR ACCREDITATIONS THROUGH RECIPROCAL AGREEMENT ACCEPTANCE. (a) A laboratory holding valid accreditations, licenses, or approvals from government bodies or private organizations, with which the department has established a reciprocal agreement, may have its accreditations, licenses, or approvals considered for acceptance by the department by submitting a reciprocity application.

(b) A laboratory applying for acceptance under an existing reciprocal agreement shall do all the following:

1. Submit certificates or official documents of the laboratory's accreditations, licenses, or approvals with its application.

2. Agree to notify the department of any changes, within 30 days of a change in its accreditation, licensure, or approval status with the entity with which the department has the agreement.

3. Submit a copy of the report of the most recent on-site evaluation performed by the entity with which the department has the agreement.

(5) ISSUANCE OF ACCREDITATIONS. (a) The department shall issue accreditations to laboratories through certificates that meet the criteria specified in s. NR 149.06.

(b) The department shall issue a certificate to a laboratory submitting an initial, revised, or reciprocity application for accreditation within 30 days of the date by which the laboratory successfully completes an on-site evaluation or the date by which the department waives an on-site evaluation, subject to all the following:

1. The department may not schedule or waive an on-site evaluation of an applicant laboratory until all the requirements of sub. (1) have been completed.

2. A laboratory completes an on-site evaluation successfully when it addresses, to the department's satisfaction, any deficiencies encountered during the on-site evaluation.

(c) Following an on-site evaluation, the department may issue accreditations, on a case-by-case basis, that are unaffected by any deficiencies encountered during the on-site evaluation.

(d) The department shall issue a revised certificate of accreditation to an accredited laboratory within 30 days of the occurrence of any of the following:

1. Receiving notification from that laboratory that it is changing its name without changing ownership.

2. Approval of relocation to a new facility that does not compromise the laboratory's ability to meet the requirements of this chapter.

NR 149.15 Period, renewal, and expiration of accreditation. **(1) ACCREDITATION PERIOD.** (a) The accreditation period shall commence on September 1 and end on August 31 of the following year for all laboratories accredited by the department.

(b) The department shall renew the accreditations of laboratories that meet the requirements of this section prior to September 1 of each year.

(2) RENEWAL PROCESS. Annually, each laboratory holding valid accreditations under this chapter and wishing to renew its accreditations shall do all the following:

(a) Pay the required annual renewal fee and any assessed administrative fees prior to July 1. After July 1, a late renewal fee may be assessed to laboratories that have not paid all requisite fees. A laboratory is not eligible for renewal of accreditation if full payment is not received prior to September 1.

(b) Submit acceptable PT sample results as required in subch. V, no later than August 31.

(c) If accredited via reciprocal agreement, submit documentation of accreditations and a copy of the most recent on-site evaluation report from the entity with which the department has the agreement.

(3) EXPIRATION OF ACCREDITATIONS. On September 1 of each year, the department shall expire the affected accreditations of laboratories failing to provide the information and fees specified in sub. (2).

(4) VOLUNTARY WITHDRAWAL OF ACCREDITATIONS. Laboratories may voluntarily withdraw accreditations at any time by notifying the department in writing.

Note: Conditions associated with applying for analytes for which accreditation was voluntarily withdrawn are provided in s. NR 149.14 (1) (b) (3).

NR 149.155 Required Notifications. (1) LABORATORY NAME CHANGE. A laboratory that changes its name without changing ownership shall notify the department, in writing, within 30 days of the effective date of the name change. The department may not charge a fee for any processing resulting solely from a name change.

(2) LABORATORY OWNERSHIP CHANGE. A laboratory that changes its ownership shall notify the department, in writing, within 30 days of the effective date of the ownership change. Notification shall be in the form of a completed application for transfer of ownership.

(3) LABORATORY RELOCATION. A laboratory relocating shall notify the department, in writing, at least 30 days prior to the relocation. Notification shall include the new address and any changes in contact information.

(4) KEY PERSONNEL CHANGES. A laboratory making changes to key personnel, including lab director, lab manager, quality assurance manager, or whole effluent toxicity technical expert, shall notify the department within 30 days of these changes.

NR 149.18 Subcontracting. (1) Subcontracting samples shall be to a laboratory that holds valid certifications corresponding to the matrix, technology or method, and analyte requested.

(2) A laboratory accepting samples under a subcontract from another laboratory shall maintain any analytical records needed to determine compliance with this chapter. The records shall be made available to the laboratory providing the samples.

NR 149.19 Requirements for certification in the drinking water matrix. (1) This section contains additional requirements that apply to laboratories analyzing drinking water for compliance under ch. NR 809.

(2) GENERAL REQUIREMENTS. (a) The minimum criteria and procedures for certification in the drinking water matrix are specified in the following documents:

1. As updated, the "Manual for the Certification of Laboratories Analyzing Drinking Water," EPA 815-R-05-004, fifth edition, EPA, Office of Ground Water and Drinking Water, January 2005.

2. As updated, the "Supplement 1 to the Fifth Edition of the Manual for the Certification of Laboratories Analyzing Drinking Water," EPA 815-F-08-006 EPA, Office of Ground Water and Drinking Water, June 2008.

Note: The documents above can be found on the Wisconsin department of natural resources laboratory accreditation program website.

(b) The department may not grant either interim or provisional certifications.

(c) A laboratory shall follow any additional criteria and procedures identified in this chapter applying to drinking water analyses.

(3) REQUIREMENTS FOR INORGANIC CONTAMINANTS. (a) To receive and maintain certification to conduct analyses of inorganic contaminants, the laboratory shall achieve MDLs no greater than the MDLs specified in 40 CFR 141.23 (a) (4) (i) and 40 CFR 141.89 (a) (1) (iii) for each accredited method.

Note: Links to the 40 CFR Part 141 can be found on the Wisconsin department of natural resources laboratory accreditation program website.

(b) Each laboratory shall successfully analyze at least one PT sample annually according to criteria specified in 40 CFR 141.23 (k) (3) (ii) or 40 CFR 141.89 (a) (1) (ii) (A) and (B) for each accredited method.

Note: Links to the 40 CFR Part 141 can be found on the Wisconsin department of natural resources laboratory accreditation program website.

(4) REQUIREMENTS FOR VINYL CHLORIDE. (a) To receive and maintain certification to conduct analyses of vinyl chloride, the laboratory shall achieve a MDL no greater than 0.0002 mg/L for each accredited method.

(b) Each laboratory shall successfully analyze at least one PT sample annually for each accredited method according to criteria specified in 40 CFR 141.24 (f) (17) (ii) (B). Vinyl chloride is evaluated separately from the other regulated volatile organic compounds and certification for the regulated volatile organic compounds requires successful analysis of vinyl chloride in addition to requirements for the other regulated volatile organic compounds.

Note: Links to the 40 CFR Part 141 can be found on the Wisconsin department of natural resources laboratory accreditation program website.

(5) REQUIREMENTS FOR OTHER VOLATILE ORGANIC COMPOUNDS. (a) To receive and maintain certification to conduct analyses of volatile organic compounds, excluding vinyl chloride, the laboratory shall achieve MDLs no greater than 0.0005 mg/L for all regulated volatile organic compounds for each accredited method.

(b) Each laboratory shall successfully analyze at least one PT sample annually for each accredited method according to criteria specified in 40 CFR 141.24 (f) (17) (i) (B). Excluding vinyl chloride, a laboratory may be certified for all volatile organic compounds if the laboratory successfully analyzes at least 80% of the regulated volatile organic compounds.

Note: Some PT sample providers include the trihalomethanes in the sample for regulated volatile organic compounds. Trihalomethanes are not considered part of the "80%" rule. To be accredited for the regulated volatile organic compounds, vinyl chloride and 16 of the remaining 20 regulated volatile organic compounds are to pass in each PT sample.

Note: Links to the 40 CFR Part 141 can be found on the Wisconsin department of natural resources laboratory accreditation program website.

(6) REQUIREMENTS FOR SYNTHETIC ORGANIC CONTAMINANTS. (a) To receive and maintain certification to conduct analyses of synthetic organic contaminants, the laboratory shall achieve MDLs no greater than the MDLs specified in 40 CFR 141.24 (h) (18) for each accredited method.

Note: Links to the 40 CFR Part 141 can be found on the Wisconsin department of natural resources laboratory accreditation program website.

(b) Each laboratory shall successfully analyze at least one PT sample annually according to criteria specified in 40 CFR 141.24 (h) (19) (i) (A) and (B).

Note: Links to the 40 CFR Part 141 can be found on the Wisconsin department of natural resources laboratory accreditation program website.

(7) REQUIREMENTS FOR DISINFECTION BY-PRODUCTS. (a) To receive and maintain certification to conduct analyses of disinfection by-products, the laboratory shall meet the requirements specified in 40 CFR 141.131 (b) (2) (iv) for each accredited method. To receive certification to conduct analyses of trihalomethanes, the laboratory shall achieve MDLs no greater than 0.0005 mg/L for each regulated analyte for each accredited method.

Note: Links to the 40 CFR Part 141 can be found on the Wisconsin department of natural resources laboratory accreditation program website.

(b) For the haloacetic acid and trihalomethane PT samples, laboratories shall pass 80%, or 4 of the analytes present in each PT sample.

(8) The department may not renew the accreditation for analytes for which the laboratory fails consecutive PT samples.

(9) CERTIFICATION EXEMPTIONS. Certification is not required to perform any of the following analyses:

(a) Fluoride analysis required under s. NR 809.74.

(b) Analysis for free chlorine residual and total chlorine residual required under s. NR 809.74.

(c) Analysis for pH required under s. NR 809.548.

(d) Analysis for turbidity required under s. NR 809.113.

NR 149.20 Requirements for whole effluent toxicity testing. All the following apply to laboratories accredited to perform whole effluent toxicity testing:

(1) ACUTE AND CHRONIC WHOLE EFFLUENT TOXICITY TESTING BY SPECIES. Laboratories analyzing whole effluents for acute and chronic toxicity for a given species shall follow the quality control requirements referenced in the "State of Wisconsin Aquatic Life Toxicity Testing Methods Manual," as updated.

Note: The methods of analysis for determining the toxicity of effluents are referenced in the "State of Wisconsin Aquatic Life Toxicity Testing Methods Manual," which can be found on the Wisconsin department of natural resources laboratory accreditation program website.

(2) ACCREDITATION REQUIREMENTS FOR CHEMICAL TESTING IN SUPPORT OF WHOLE EFFLUENT TOXICITY TESTING. Water chemistry testing performed in support of whole effluent toxicity testing for ammonia, alkalinity, hardness, pH, dissolved oxygen, conductivity, and total residual chlorine do not require separate accreditation under this chapter.

NR 149.21 Fees. The department shall establish fees for laboratories participating in the laboratory accreditation program. These costs include those associated with laboratory evaluations, discretionary acceptance of data, reciprocity, training, and collection of fees. Fees may not be prorated and, except for overpayment, are not refundable.

(1) TOTAL FEE INCOME. (a) The laboratory accreditation program's total fee income shall be designed to generate revenues equal to the costs of administering this chapter. Any amendments to the formulas in this subsection shall be reviewed by the council prior to being proposed as rule amendments.

(b) The department may adjust the fee schedule according to the formulas in this subsection and the relative value unit items specified in Tables 1, 2, and 3. Annual fee adjustments shall be reviewed by the council and approved annually by the natural resources board.

(c) The following formulas shall be used to generate and adjust the laboratory accreditation program's fee schedule:

1. Fee Revenue Required = Projected Laboratory Accreditation Program Expenses - (Application Fees + Travel Reimbursement)

a. Fee Revenue Required is the total amount of revenue which shall be collected via fees to cover all laboratory accreditation program costs.

b. "Laboratory Accreditation Program Expenses" is the sum of all anticipated laboratory accreditation program expenses including salary, fringes, evaluation travel costs, supplies, and services. This includes travel costs for evaluation of out-of-state labs which are required to reimburse the laboratory accreditation program for laboratory evaluation travel costs. Application Fee revenues are excluded from the Fee Revenue required because the application fee revenues are variable and collected independently throughout an accreditation period.

Note: "Laboratory accreditation program expenses" may not exceed the legislature's approved spending authority for the laboratory accreditation program in fiscal year. The department of administration approved spending authority is given in s. 20.370(3)(f), Stats., and may be revised by the department of administration to recover laboratory accreditation program cost.

c. Application Fees is a three-year moving average of application fees received for the three most recent fiscal years.

d. Travel Reimbursement is a three-year moving average of out-of-state travel reimbursements for the three most recent fiscal years. Laboratory accreditation program costs related to travel for out-of-state audits are negated because the department recovers these costs directly from each lab.

Note: For example, given the following,

Projected Laboratory Accreditation Program Expenses = \$612,121

Application Fees (three-year average) = \$31,681

Travel Reimbursement (three-year average) = \$17,079

Fee Revenue Required would be \$612,121 - (\$31,681 + \$17,079) = \$563,361.

2. Total # RV Units = $\sum [(\# \text{Laboratories in Item}) \times (\text{RV of Item})]$.

a. Total # RV Units is the total number of RV units available for the fiscal year. It is the mechanism by which fees are distributed to individual laboratories.

b. "#Laboratories in item" is a count of how many laboratories will be assessed the fee for that item for a fiscal year, based on accreditations currently held.

c. The RV units for each fee item, "RV of item," are listed in Table 3. The total number of RV units is the sum of all base fee RV, matrix fee RV, and technology or class fee RV.

Note: For example, given the following,

Base RV: #Labs Registered (5 RV) = 225; #Labs Certified (10 RV) = 141;

Matrix RV = 341 Aqueous, 71 Solid and 43 Drinking water.
 Technology/Class RV = 2612 Aqueous, 943 Solid and 575 Drinking water.
 Base Fee RV Units = (225 x 5) + (141 x 10) = 2535
 Matrix RV Units = (341 + 71 + 43) = 455 x 5 RV/Matrix = 2275
 Technology/Class RV Units = (2612 + 943 + 575) = 4130 RV
 Total # RV Units = 2535 + 2275 + 4130 = 8940 RV Units.

3. Cost per RV = Fee Revenue Required / Total # RV Units. The Cost per RV is the dollar value assigned to one RV unit and is used to establish all fees for items in Table 3 of this section. The cost per RV is rounded to the nearest \$0.50 to simplify fee statements.

Note: For example, given the following,
 Fee Revenue Required = \$563,361
 Total # RV Units = 8940 RVU
 Cost per RV (\$/RV) = \$563,361 / 8940 RVU = \$63.01/RV Units; rounded to the nearest \$0.50 = \$63.00/RV Units.

4. Laboratory fees = (# RV units for a given laboratory) x (Cost per RV). The sum of base, matrix, technology, and class fees for a given laboratory is multiplied by the cost per RV to determine the fee for each laboratory. Any outstanding administrative fees may also be added.

Note: For example, given the following for Pinestump Wastewater Treatment Plant Laboratory,
 Base RV: 5 (registered)
 Matrix RV = 5 (aqueous matrix only)
 Technology/Class RV = 4 (Oxygen Demand Assays = 3 RV; Gravimetric Assays - Residue = 1 RVU).
 Total # RV Units = 5 + 5 + 4 = 14 RV Units
 Laboratory fee = 14 RV Units x \$63.00/RV Units = \$882.00.

(2) ADMINISTRATIVE FEES. The department shall assess fees to recover the cost of specified administrative functions specified in this section, Table 1. Any outstanding administrative fees may be included as part of the annual fee.

Table 1 - Administrative Fees

Item	RV Units
Discretionary Acceptance (s. NR 149.11)	Actual Cost
Evaluation Cancellation ¹	Incurred Costs
Evaluation for Enforcement Follow-Up	Actual Cost
Evaluation of Out-of-State Laboratories	Travel Cost
Late Renewal Fee ²	2

¹ Out-of-state laboratories may be required to reimburse the laboratory accreditation program for travel costs incurred by the cancellation or postponement of an evaluation, including airfare, hotel, and rental car expenses.

² Assessed 30 days after payment due date.

(3) APPLICATION FEES. (a) The department shall assess fees for all applications specified in this section, Table 2.

(b) The fee for an application also includes matrix and technology or class fees when a laboratory applies for a new matrix, technology, or class. The matrix fee is not required if a laboratory is applying for additional technologies or analytes within a matrix for which the lab already holds accreditation. Technology fees are not required if a laboratory already holds accreditation for that matrix and technology or matrix and class, for drinking water combination.

Note: Example – The application fee for a laboratory applying to add ammonia by colorimetry under the aqueous matrix is based on only the number of RV units for a revised application since the lab has already paid for the aqueous matrix and colorimetry technology as part of its renewal fees.

(c) Application fees are not refundable in either whole or part.

(d) If an application is not completed within a single fiscal year, the department may adjust the fees on the application to recover the difference in fees between the year the application was submitted and the year the application was completed. The laboratory shall pay this difference prior to receiving accreditations.

Table 2 - Application Fees

Item	RV Units
Initial Application	6
Revised Application	3
Reciprocity Application	4
Transfer of Ownership Application	4

(4) ANNUAL FEES. The department shall assess an annual fee to each laboratory holding accreditations under this chapter either directly or through agreements. A laboratory's annual fee shall be the sum of all the following:

(a) The base fees for accreditation. The department shall assess a base fee to all laboratories holding accreditations under this chapter. The number of RV units assigned to each type of base fee is specified in Table 3 of this subchapter.

(b) The matrix fees. The department shall assess a fee per matrix type to all accredited laboratories. The number of RV units assigned to each type of matrix fee is specified in this section, Table 3.

(c) The analytical fees.

1. Analytical technology fees. The department shall assess a fee for each analytical technology, per matrix, to all accredited laboratories, in fields involving the aqueous and non-aqueous matrices. The assessed fee shall be based on the RV units specified in this section, Table 3.

2. Analytical class fees. The department shall assess a fee, per analytical class, to all certified laboratories in fields involving the drinking water matrix. The assessed fee shall be based on the RV units specified in this section, Table 3.

(d) Any outstanding administrative fees.

Table 3 - Annual Fees for Accreditation

	Item	RV Units
A.	Administrative Fees	
	Outstanding administrative fees	per Table 1 of this subchapter
B.	Base Fees	
	Base Fee, Certification	10
	Base Fee, Registration	5
C.	Matrix Fees	
	Matrix Fee, Aqueous	5
	Matrix Fee, Drinking Water	5
	Matrix Fee, Non-Aqueous	5
D.	Analytical Technology Fees for Aqueous and Non-Aqueous Matrices	

	General Chemistry	
	Oxygen Demand Assays (BOD or cBOD)	3
	Colorimetric or Turbidimetric	2
	Electrometric Assays (i.e. ion-selective electrodes)	1
	Gravimetric Assays - Residues (solids)	1
	Extraction/Gravimetric Assays - Oil & Grease as Hexane Extractable Materials (HEM)	2
	Titrimetric or Potentiometric Titration Assays	1
	Flow Injection-Gas Diffusion-Amperometry	4
	Nondispersive Infrared (NDIR) or Microcoulometry	2
	Ion Chromatography (IC)	4
	Metals	
	Flame Atomic Absorption Spectrophotometry (FLAA)	2
	Flame Photometry Spectrophotometry (FP)	2
	Gaseous Hydride Atomic Absorption Spectrophotometry (GHAA)	3
	Graphite Furnace Atomic Absorption Spectrophotometry (GFAA)	3
	Cold Vapor Atomic Absorption Spectrophotometry (CVAA)	3
	Cold Vapor Atomic Fluorescence Spectrophotometry (CVAFS)	3
	Thermal Decomposition Atomic Absorption Spectrophotometry (TDAA)	3
	Inductively Coupled Plasma Emission Spectrophotometry (ICP)	4
	Inductively Coupled Plasma-Mass Spectrometry (ICP/MS)	5
	Organics	
	Gas Chromatography (GC)	4
	Gas Chromatography-Mass Spectrometry (GC/MS)	5
	Liquid Chromatography (LC)	4
	Liquid Chromatography-Mass Spectrometry (LC/MS)	5
	High Resolution Gas Chromatography-Mass Spectrometry (HRGC/MS)	10
	Other	
	Hazardous Waste Characteristics	2
	Solid Waste Leaching Procedures	2
	Whole Effluent Toxicity Assays	5
	Other	Not to exceed 10 ¹
E.	Analytical Class Fees for Drinking Water Matrix	
	Disinfection By-products	5
	Primary Inorganic Contaminants (Non-Metals)	3
	Primary Inorganic Contaminants (Metals)	6
	Secondary Contaminants (Non-Metals)	2
	Secondary Contaminants (Metals)	3
	Synthetic Organic Contaminants (SOC) – Dioxin	8
	Synthetic Organic Contaminants (SOC) – Organochlorine Pesticides	3
	Synthetic Organic Contaminants (SOC) – Nitrogen-Phosphorus Pesticides	3
	Synthetic Organic Contaminants (SOC) – Herbicides	3
	Synthetic Organic Contaminants (SOC) – Miscellaneous	4
	Volatile Organic Compounds (VOC)	4
	Other	Not to exceed 10 ¹

¹ Actual cost will be determined by the department considering the complexity of the technology.

SUBCHAPTER V - PROFICIENCY TESTING

NR 149.22 Required proficiency testing samples and frequency of analysis. (1) REQUIREMENTS. (a) A laboratory shall participate in at least one PT sample study per accreditation period as specified in sub.(2), subject to all the following:

1. For aqueous and non-aqueous matrices, a laboratory shall analyze aqueous matrix PT samples for each combination of technology and analyte or analyte group in its fields of accreditation.

2. For the drinking water matrix, a laboratory shall analyze PT samples for each combination of method and analyte or analyte group in its fields of certification. Acceptance criteria for these samples are set in s. NR 149.27.

(b) PT samples may be those offered by approved PT sample providers at regular intervals, as “rapid response” PT samples, or as custom formulations approved by the department.

(c) A laboratory shall report a proper method code, which matches the technology and analyte or analyte group for which accreditation is held, with results for PT samples.

Note: A link to the universal list of method codes for methods and technologies is available from the NELAC Institute (TNI) which can be found on the Wisconsin department of natural resources laboratory accreditation program website.

(2) LISTS OF REQUIRED PT SAMPLES AND APPROVED PT SAMPLE PROVIDERS. (a) The department shall seek the advice of the council prior to identifying required PT samples and approved PT sample providers.

(b) The list shall identify matrix-specific PT samples required for submittal for renewal of accreditation or with initial or revised applications and the specific PT sample providers approved for supplying each required PT sample.

Note: Lists of required PT samples and approved PT sample providers can be found on the Wisconsin department of natural resources laboratory accreditation program website.

NR 149.23 Approval of proficiency testing sample providers. (1) CRITERIA FOR APPROVAL. (a) When evaluating a PT sample provider for approval, the department shall consider criteria including all the following:

1. The PT sample provider accreditation status by national accreditation programs.

2. The PT sample provider use of techniques for calculating acceptance limits as specified in s. NR 149.27.

(2) PROFICIENCY TESTING SAMPLE PROVIDER REQUIREMENTS. Approved PT sample providers shall submit all PT sample results to the department electronically in a format specified by the department.

NR 149.24 Proficiency testing samples for applications and accreditation renewal. (1) PT SAMPLE ACCEPTANCE CRITERIA. The department may not grant or renew an accreditation unless the associated PT sample results meet the criteria specified in s. NR 149.27.

(2) PT SAMPLE STUDY CLOSE DATE. Acceptable PT samples shall have a PT sample study close date no more than six months prior to the date of application.

(3) PT SAMPLE DUE DATE FOR RENEWAL. For renewal of accreditations, which begin on September 1 of each calendar year, acceptable PT sample results shall have been reported electronically to the department by an

approved PT sample provider no sooner than January 1 or later than August 31 of the same calendar year. Preliminary reports from approved PT sample providers may not be used for renewal of accreditation.

Note: For example, to renew accreditation for any analyte effective for the period from September 1, 2009 to August 31, 2010, a laboratory shall have successfully analyzed a PT sample for that analyte reported between January 1 and August 31, 2009.

(4) PT SAMPLES FOR APPLICATIONS. A laboratory submitting initial or revised applications for accreditation shall analyze PT samples from an approved PT sample provider and submit acceptable results for any of the following:

(a) For aqueous and non-aqueous matrices, acceptable PT sample results are required for each combination of technology and analyte or analyte group for which the department has identified that PT samples are required.

(b) For the drinking water matrix, acceptable PT sample results, from a water supply study, are required for each combination of method and analyte or analyte group.

(5) PT SAMPLES FOR RENEWAL. A laboratory wishing to renew its accreditation shall analyze PT samples from an approved PT sample provider and submit acceptable results for any of the following:

(a) For aqueous and non-aqueous matrices, acceptable PT sample results from a water pollution study are required for each combination of technology and analyte or analyte group for which the department has identified that PT samples are required.

(b) For the drinking water matrix, acceptable PT sample results from a water supply study are required for each combination of method and analyte or analyte group.

Note: The department does not accept PT samples prepared in a non-aqueous matrix to obtain or renew accreditation for analytes or analyte groups under the non-aqueous matrix.

(6) RENEWAL REQUIREMENTS FOR MULTIPLE SUCCESSIVE PT SAMPLE FAILURES. A laboratory that experiences multiple successive PT sample failures shall submit two consecutive acceptable PT samples from an approved PT sample provider to renew its accreditation. Consecutive PT samples shall be two unique studies received by the laboratory at least ten business days apart. The laboratory may not prepare or analyze the two PT samples in the same batch.

(a) For aqueous and non-aqueous matrices, PT sample failure means failing three consecutive PT samples for any combination of technology and analyte or analyte group.

(b) For the drinking water matrix, PT sample failure means failing two consecutive PT samples for any combination of method and analyte or analyte group.

NR 149.25 Treatment of proficiency testing samples. (1) PT samples shall be subjected to any preparatory steps undergone by analytical samples of that matrix, unless the preparation instructions submitted by a PT sample provider specifically instruct omitting a preparatory step.

Note: Preparatory steps include digestions, distillations, extractions, concentrations, and dilutions.

(2) A laboratory may report multiple results for a single PT sample when the laboratory maintains accreditations for multiple technologies for any analyte or analyte group in aqueous and non-aqueous matrices.

(3) A laboratory may report multiple results of a single PT sample when the laboratory maintains certifications for multiple methods for any analyte or analyte group in the drinking water matrix.

(4) Prior to submitting PT sample results to a PT sample provider, all the following apply:

(a) A laboratory may not send a PT sample, or portion of a PT sample, to another laboratory for analysis.

(b) A laboratory may not knowingly analyze a PT sample, or a portion of a PT sample, from another laboratory.

(c) Until a PT sample study has been closed, a laboratory may not share results of a PT sample from that study to any party other than the PT sample provider or regulatory agency.

NR 149.26 Reporting proficiency testing sample results. (1) A laboratory shall submit PT sample results to PT sample providers in accordance with the dates specified by the PT sample providers.

(2) PT sample reports may be submitted to the department directly from the PT sample provider or by the laboratory, but it is the laboratory's responsibility to ensure the department receives the necessary reports for initial and revised applications. The laboratory shall submit PT sample reports in their entirety, without modification, to the department.

(3) Results from all PT sample reports issued to the department by PT sample providers shall be used to determine a laboratory's accreditation status.

(4) The department may only accept amended and reissued PT sample reports if the reissue is due to an error made by the PT sample provider and revised reports are all the following:

(a) Clearly labeled as revised or reissued.

(b) Directly submitted to the department by the PT sample provider.

(c) Accompanied by an explanation of the PT sample provider's error.

Note: Re-issued reports are acceptable in cases when the laboratory neglected to instruct the PT sample provider to report results to the department.

NR 149.27 Proficiency testing sample acceptance limits and grading. (1) ACCEPTANCE LIMITS. A laboratory's result for any analyte or analyte group is considered unacceptable if it meets any of the following conditions:

(a) The result falls outside the acceptance limits.

(b) The laboratory reports a result for an analyte not present in the PT sample.

(c) The laboratory does not report a result for an analyte present in the PT sample.

(d) The laboratory fails to submit its results to the PT sample provider on or before the deadline for the PT sample study.

(e) The laboratory reports a method code for either an unapproved method or the method code reported is not appropriate for the technology-analyte or method-analyte combination.

(f) The laboratory fails to meet department specified grading criteria for multi-analyte PT samples.

Note: Department grading criteria can be found on the Wisconsin department of natural resources laboratory accreditation program website.

(2) GRADING. (a) PT samples for analytes in aqueous and non-aqueous matrices shall be graded in accordance with acceptance limits established by the department considering criteria developed by the EPA.

(b) When the EPA has not developed acceptance limits for required PT sample analytes, the department may develop acceptance limits based on its experience or information supplied by approved PT sample providers.

(c) When an insufficient number of laboratories participate in a study to generate peer-based acceptance limits in a PT sample with analytes for which the EPA has not established acceptance limits, the department may grade results using fixed acceptance limits.

(d) PT sample analytes in drinking water shall be graded in accordance with the acceptance limits established in 40 CFR 141.23 (k) (3) (ii), 40 CFR 141.24 (f) (17) (i) (C) and (D), 40 CFR 141.24 (f) (17) (ii) (B), 40 CFR 141.24 (h) (19) (i) (A) and (B), and 40 CFR 141.89 (a) (1) (ii), and 40 CFR 141.131 (b) (2) (ii) and (iii).

Note: Links to the 40 CFR Part 141 can be found on the Wisconsin department of natural resources laboratory accreditation program website.

(e) When accreditation in an analyte group is based on passing a representative PT sample containing more than one analyte, the laboratory shall report acceptable results on at least 80% of the analytes to achieve acceptable results for that sample. The department may investigate repeated failures for specific analytes and direct enforcement action in the event of two consecutive failures in the drinking water matrix or three consecutive failures in the aqueous matrix.

(f) The department shall establish procedures for evaluating false positives and false negatives reported in analyzed PT samples.

NR 149.28 Procedure for correcting unacceptable proficiency testing sample results. (1) AQUEOUS AND NON-AQUEOUS MATRICES. (a) If a laboratory does not meet the acceptance limits for an analyte or analyte group and the laboratory does not have acceptable results on a previous sample analyzed during the same accreditation period, the laboratory shall analyze a second PT sample for that analyte or analyte group.

(b) If the results of a second PT sample do not meet the acceptance limits, the department may initiate an assessment of the laboratory's quality control records if this action is necessary to validate data generated by the laboratory. If two consecutive PT samples do not meet acceptance limits, the laboratory shall do all the following:

1. Prepare a corrective action report and initiate an action plan to correct the problems within 30 days of the date of notification of the second failure. This action plan shall include a timetable for correcting the problems and obtaining a third PT sample.

2. Analyze a third PT sample within 60 days of the date of notification of the second failure. If the results of the third PT sample do not meet the acceptance limits, the laboratory shall analyze two subsequent and consecutive acceptable PT samples.

(c) The department may not renew accreditation of those analytes or analyte groups for which a laboratory has failed three consecutive PT samples and has not successfully analyzed two subsequent and consecutive PT samples for those analytes or analyte groups prior to September 1.

(d) When applying to have an analyte or analyte group reinstated after non-renewal for failing three consecutive PT samples, the laboratory shall provide acceptable results on two subsequent and consecutive PT sample studies for that analyte or analyte group. The consecutive PT samples shall be two unique studies received by the laboratory at least ten business days apart. The laboratory may not prepare or analyze the two PT samples in the same batch.

(2) DRINKING WATER. If a certified laboratory does not meet the acceptance limits that have been established by the department, the department shall require the laboratory to analyze a second PT sample and may require the laboratory to submit a corrective action report. If the results of the second sample do not meet the acceptance limits, the department may not renew the laboratory's certification and may revoke the laboratory's certification as specified in s. NR 149.10. To re-instate the certification for the affected method-analyte or analyte group, the laboratory shall submit a revised application, pay the revised application fee, and provide acceptable results on two subsequent and consecutive PT sample studies for that method-analyte or analyte group. The consecutive PT samples shall be two unique studies received by the laboratory at least ten business days apart. The laboratory may not prepare or analyze the two PT samples in the same batch.

SUBCHAPTER VI - ON-SITE LABORATORY EVALUATIONS

NR 149.29 Purpose, type and frequency. (1) The department shall perform on-site evaluations to determine a laboratory's potential, actual, or continued ability to comply with the provisions of this chapter.

(2) The department shall conduct announced on-site evaluations of laboratories once every three years and when any of the following occurs:

(a) A laboratory applies to become certified or registered in any field of accreditation unless the department waives the requirement to perform an on-site evaluation. When the department does not waive an evaluation, the evaluation shall be performed within 90 days after the department determines that a received application is complete.

(b) A laboratory changes its location, ownership or key personnel, unless the department waives the requirement to perform an on-site evaluation. When the department does not waive an evaluation, the evaluation shall be performed within 90 days after the department receives notification of these changes.

(c) The department determines that an on-site evaluation is necessary to verify corrective action implemented by a laboratory to address deficiencies identified in a previous on-site evaluation.

(d) The department has reason to believe that a laboratory is not in compliance with this chapter.

(3) The department may conduct unannounced on-site evaluations of a laboratory to verify compliance with this chapter after a notice of violation has been issued to a laboratory.

NR 149.30 Evaluation procedures and appraisal. (1) The department shall perform on-site evaluations of laboratories to evaluate systems, practices, procedures, and documentation in a laboratory and to identify deficiencies according to documented procedures that promote consistency in determining a laboratory's potential, actual, or continued ability to comply with this chapter.

(2) If, in performing an on-site evaluation, the department finds that the laboratory is implementing a procedure that is neither allowed nor disallowed by method or this chapter, the department will assess the scientific validity of the procedure. The department may seek the advice of the council in making determinations under this paragraph.

(3) The department shall provide laboratories with a survey to allow them to appraise the evaluation process.

NR 149.31 Evaluation reports. (1) The department shall document the deficiencies identified during an on-site evaluation under s. NR 149.30 in reports issued to the evaluated laboratory.

(2) The report of an on-site evaluation shall be issued to a laboratory within 30 days of the conclusion of the on-site visit. When the department finds it necessary to issue an evaluation report at a date later than 30 days after the conclusion of an on-site visit, the department shall notify the laboratory about the delay. The notice shall include an expected delivery date for the report.

NR 149.32 Evaluation corrective action. (1) A laboratory shall take corrective action to address all deficiencies discovered during an on-site evaluation under s. NR 149.30 and contained in an evaluation report under s. NR 149.31.

(2) A laboratory shall submit to the department, within 30 days from the evaluation report's date, a plan of corrective action to address all the deficiencies noted in the report. When a laboratory finds it necessary to submit a corrective action plan at a date later than 30 days after the evaluation report's date, the laboratory shall notify the department about the delay and provide an expected delivery date in consultation with the department.

(3) The department shall review the corrective action plan submitted by a laboratory under sub. (2) and inform the laboratory whether the submitted plan addresses satisfactorily all noted deficiencies, or whether additional action or documentation is necessary to determine the laboratory's ability to comply with this chapter, subject to all the following:

(a) When the department determines that the submitted corrective action plan addresses all noted deficiencies satisfactorily, the department shall inform the laboratory in writing within 30 days that the plan is acceptable.

(b) When the department determines that additional action or documentation is needed to evaluate compliance with this chapter, the department, in consultation with the laboratory, shall set a date for the laboratory to submit a second corrective action plan.

1. If the department determines that the second corrective action plan submitted under sub. (3) (b) addresses all noted deficiencies satisfactorily, the department shall inform the laboratory in writing that the evaluation process has concluded.

2. If the department determines that the second corrective action plan submitted under sub. (3) (b) does not address all the noted deficiencies satisfactorily, the department may schedule another on-site evaluation to determine the laboratory's compliance with this chapter, terminate any outstanding application that led to the original on-site evaluation, or direct enforcement to the laboratory.

3. If a second on-site evaluation is scheduled as a follow-up to a second corrective action plan submitted under sub. (3) (b), the department shall establish deadlines that resolve any remaining unresolved deficiencies expeditiously, but no later than 90 days after the conclusion of the follow-up visit.

NR 149.33 Conflicts of interest. (1) The department shall establish procedures to ensure and document that laboratory evaluators under its employment are free of any conflicts that would render the laboratory evaluator incapable of performing an objective and unbiased evaluation of a laboratory.

(2) A laboratory may request information and documents used by the department to establish that any evaluator assigned to perform the laboratory's evaluation is free of any conflicts of interest.

NR 149.34 Evaluator qualifications. (1) The department shall develop procedures to establish and evaluate the education, experience, and credentials of the laboratory evaluators under its employment.

(2) A laboratory may request information and documents used by the department to establish that any evaluator assigned to perform the laboratory's evaluation has the necessary education, experience, or credentials to perform evaluations competently.

SUBCHAPTER VII - QUALITY SYSTEMS

NR 149.35 General requirements. (1) **SCOPE.** This subchapter establishes personnel, quality assurance, quality control, method selection, sample handling, and documentation requirements for laboratories.

(2) **RESPONSIBILITY FOR QUALITY SYSTEM.** A laboratory shall conduct analytical activities under a quality system that incorporates the provisions of this subchapter. At least one individual within a laboratory's organization or under the laboratory's employment shall be identified to the department as responsible for establishing, implementing, assessing, and revising, as needed, a laboratory's quality system.

NR 149.36 Laboratory personnel. (1) **MANAGEMENT AND ANALYTICAL STAFF.** The laboratory shall have personnel with education, training, or experience that allows the laboratory to comply with the requirements of this chapter. Contractors, external to the laboratory, may serve in key laboratory roles. When external contractors serve in essential laboratory roles, the contracts shall be available to the department to ensure that contractual specifications satisfy the requirements of this chapter.

Note: For requirements regarding changes in key personnel see s. NR 149.155.

(2) **DEMONSTRATION OF CAPABILITY.** (a) When a laboratory references a method that contains procedures for demonstrating initial capability, continuing capability or both, personnel performing analyses using these methods shall perform the procedures, meet any associated evaluation criteria, and document the results. When initial demonstrations of capability include the analysis of samples, the samples shall be prepared from a clean matrix and processed through all method preparation steps.

(b) When a laboratory references a method that does not contain procedures for demonstrating initial capability, the laboratory shall establish initial demonstration of capability criteria for determining that each person who performs testing on compliance samples using the method has demonstrated the necessary skills and expertise required to generate quality analytical results. The laboratory shall retain documentation that each person performing a given test on compliance samples has satisfied the demonstration of capability criteria established by the laboratory.

NR 149.365 Laboratory ethics. All the following practices are prohibited and may result in enforcement action under s. NR 149.10:

(1) Fabrication, falsification, or misrepresentation of data.

(2) Improper instrument clock setting, termed time traveling, or improper recording of date or time.

(3) Unwarranted manipulation of samples, software, peak integration, or analytical conditions.

(4) Concealing or failing to report a known improper or unethical behavior or action associated with sample analysis.

NR 149.37 Quality manual. (1) PURPOSE AND GENERAL PROVISIONS. Each laboratory shall define its quality system in a quality manual. All policies and procedures governing the laboratory's quality system shall be documented or referenced in the quality manual. All laboratory personnel shall follow the policies and procedures established by the quality manual.

(2) FORMAT. The quality manual shall have a format that addresses the content elements specified in this section. Content elements may be presented in narrative, tabular, schematic, or graphical form. The manual shall be a document in hard copy or electronic format traceable to the laboratory.

(3) CONTENT. Unless included in other standard operating procedures maintained under s. NR 149.40, the quality manual shall include, address, or refer to all the following elements:

(a) Procedures for retention, control, and maintenance of documents used in or associated with analysis.

(b) Procedures for achieving traceability of standards, reagents, and reference materials used to derive any results or measurements.

(c) Procedures for handling samples.

(d) Procedures for calibration, verification, and maintenance of support equipment.

(e) Procedures for evaluating quality control samples.

(f) Procedures for initiating, following up on, and documenting corrective action, addressing quality assurance and quality control failures, and any discrepancies or nonconformances.

(4) REVISIONS. The quality manual shall be kept current. All editions or versions of the quality manual shall indicate the dates in which the quality manual was issued or revised.

NR 149.38 Corrective action. (1) The laboratory shall take corrective action in response to any nonconformances including all the following:

(a) Departures from established procedures in the quality system are identified.

(b) Quality control samples fail, unless immediate reanalysis of the affected sample resolves the issue.

(2) The corrective action under sub. (1) shall identify the problem, determine the most probable cause of the problem, implement solutions to correct the problem, and include a mechanism to verify that the action has had the desired effect.

(3) The laboratory shall document corrective action taken to address the nonconformance under sub. (1) and any other changes resulting from corrective action investigations. Changes implemented to address failures of quality control samples shall be those that resolve or address the failure. Changes shall be implemented to minimize the number of affected results reported by a laboratory.

(4) The laboratory shall monitor the effectiveness of implemented corrective action changes and take additional corrective action when initial or subsequent corrective action fails to resolve the nonconformance.

Note: The analyst may not always be able to identify the cause of isolated nonconformance incidents.

(5) Root cause analysis shall be performed when there is recurrence.

NR 149.39 Records and documents. (1) RECORDS AND DOCUMENTS RETENTION AND CONTROL. (a) The laboratory shall establish procedures to control and manage all records and documents that form part of its quality system and that are required to demonstrate compliance with this chapter.

(b) The procedures shall ensure that documents required to perform analyses and to ensure the quality of generated data are available to laboratory personnel, and that records and documents are reviewed periodically for continuing suitability and, when necessary, revised to facilitate compliance with the requirements of this chapter.

(c) The laboratory shall retain all records and documents, which are part of its quality system, and that are required to demonstrate compliance with this chapter, for a minimum of three years after the generation of the last entry in an associated record or document. The laboratory shall retain records and documents for a longer minimum period if the records and documents are necessary to reconstruct analytical results generated during a three-year period.

(d) The department may require, in writing, that records be retained for a longer period than that specified in par. (c) if the department has initiated legal action involving test results or the accreditation status of the laboratory.

(e) The laboratory shall identify to the department a responsible party for retaining documents and records for the required period in the event the laboratory changes ownership or ceases to be accredited.

(f) Records and documents shall be handled and stored in a manner that ensures permanence and security for the required retention period and that facilitates retrieval to demonstrate compliance with this chapter.

(g) All records shall allow for reconstruction of reported results from raw data. Records and documents shall be legible, and entries shall be safeguarded against obliteration, erasures, overwriting, and corruption and are subject to all the following requirements:

Note: The determination of legibility includes concerns regarding the quality and permanence of records and the ability to decipher numbers and letters. For example, thermal paper ages and eventually becomes unreadable, so thermal paper printouts should ultimately be scanned or copied to ensure permanence.

1. Handwritten records shall be recorded in ink.

2. Records and documents that are stored only on electronic media shall be supported by the hardware and software necessary for retrieval and reproduction into hard copy.

3. Corrections or other alterations made to entries in records or documents may not obscure the original entry.

4. The laboratory shall have procedures to prevent unauthorized access or amendments to records and documents.

(2) ADMINISTRATIVE RECORDS. A laboratory shall maintain all the following administrative records:

(a) Certificates of accreditation issued by the department unless the department has requested a laboratory to return the certificates to the department.

(b) Certificates issued to the laboratory by entities with which the department has entered into a reciprocal agreement under s. NR 149.08, if a laboratory is accredited for this chapter under any existing agreement.

(c) Records of personnel qualifications, experience, and training when personnel are required to possess or maintain specific credentials by s. NR 149.36 (2).

(d) Copies of, or access to, other regulations, standards, and documents necessary for the laboratory to operate or to maintain compliance with this chapter.

(3) REAGENT AND STANDARD RECORDS AND REFERENCE MATERIALS. The laboratory shall document the identity, source, and purity of standards and reagents used in the methods performed. The laboratory shall retain records of reference materials and certificates of analysis when the records are provided by the supplier and are necessary to establish the identity, source, or purity of standards and reagents.

(a) Reagent containers shall be labeled with an expiration date, chemical name, and concentration. Except for instrument vials, standard containers shall be labeled with an expiration date, chemical name, and concentration.

(b) The laboratory shall document the lot number, manufacturer, chemical name, concentration, and the date of expiration for standards and reagents purchased from a manufacturer. These records shall be separate from the container labels.

Note: An expiration date is not required when one is not provided by the supplier.

(c) The laboratory shall document the preparation details of all prepared standards and reagents. These records shall link the prepared standards and reagents to the respective originating stocks or neat compounds and shall indicate the date of preparation, date of expiration, and the identity of the preparer.

(d) The laboratory may not use any standards and reagents beyond the expiration dates unless the laboratory is using the standard and reagents for qualitative determinations.

(e) Certificates for all reference materials shall be maintained.

(4) ANALYTICAL AND TECHNICAL RECORDS. The format of the analytical and technical records of a laboratory shall facilitate access to the information in this subsection and may be contained in bench sheets, log books, notebooks, journals, manuals, standard operating procedures under s. NR 149.40, and forms, in hard copy or electronic media.

(5) SAMPLE COLLECTION RECORDS. The laboratory shall retain records supplied by the collector to allow the laboratory to evaluate collection information against the laboratory's sample acceptance policy.

NR 149.40 Standard operating procedures. (1) A laboratory shall maintain written standard operating procedures that document or reference activities needed to maintain its quality systems and that enable performing or reproducing an analysis in its entirety as performed at the laboratory. Each laboratory shall develop, maintain, and keep current its standard operating procedures for both sample preparation and analysis.

Note: Sample preparation includes digestions, distillations, extractions, concentrations, dilutions, and clean-up performed on samples prior to the determinative analytical step.

(2) Standard operating procedures may be documents written by laboratory personnel or may consist entirely of copies of published documents, manuals, or procedures if the laboratory follows the chosen source exactly.

(3) Standard operating procedures may consist, in part, of copies of published documents, manuals, or procedures if all the following conditions are met:

(a) Modifications to the published source are described in writing in additional documents.

(b) Clarifications, changes, or choices are completely described in additional documents, when published sources offer multiple options, ambiguous directives, or insufficient detail to perform or reproduce an analysis.

(4) Standard operating procedures shall indicate the dates of issue or revision.

(5) When the standard operating procedure is written by the laboratory, each standard operating procedure shall include, address, or refer to all the following elements, if applicable:

(a) Identification of the referenced method.

(b) For multi-analyte methods, a list of analytes.

(c) Potential interferences and how the interferences are treated.

(d) Equipment and analytical instruments.

(e) Consumable supplies, reagents, and standards.

(f) Sample preservation, storage, and hold time.

(g) Quality control samples and frequency of the analysis.

(h) Calibration and standardization.

(i) Procedure for analysis.

(j) Data assessment and acceptance criteria for quality control measures.

(k) Corrective actions and contingencies for handling out of control or unacceptable data.

NR 149.41 Method selection. (1) The laboratory shall use methods for environmental testing required by covered programs under this chapter and that are suitable for the matrix, type of analyte, expected level of analyte, regulatory limit, and potential interferences in the samples to be tested.

Note: Sources, including the following as updated, likely contain methods that are acceptable for testing under this chapter: The EPA, the department, Standard Methods for the Examination of Water and Wastewater, Test Methods for Evaluating Solid Waste, Physical/Chemical Methods – SW-846, American Society for Testing and Materials, and the U.S. Geological Survey Agency.

(2) When methods are not specified by covered programs under this chapter or specified in permits issued by the department, the laboratory shall consult with the department to select a method that meets the requirements in sub. (1).

(3) When using methods associated with the methods compendium document, “Test Methods for Evaluating Solid Waste,” the laboratory shall comply with the minimum requirements of the methods as written and state which options are being implemented when options exist.

Note: The document above can be found on the Wisconsin department of natural resources laboratory accreditation program website.

(4) The department will assess the scientific validity of method modifications to determine if the modification is within the scope of a method.

NR 149.42 Alternative methods. (1) The department may allow the use of alternative methods from those required by covered programs, including the safe drinking water program, if a laboratory requests approval and if the EPA has granted approval for the alternative methods.

(2) On a case-by-case basis, the department may allow the use of methods other than those required by covered programs for any of the following situations:

(a) After consultation with the department, the manager of a covered program determines that the allowance does not result in a detrimental effect on the quality and defensibility of the results to be generated.

(b) The request is for approval of a method that employs a new or emerging technology and there is documentation that substantiates the validity of the new or emerging technology for the intended purpose.

(3) Requests to use an alternative method shall include the reason for seeking the approval, a description of the principles of any new or emerging technology involved, and the potential scope of application of the method. The department may establish criteria for validating the method for the specific application and scope requested. If the laboratory's method validation results meet the established validation criteria, the department shall allow the use of the method for the specific application and scope requested.

(4) The department shall approve or deny the request for consideration of approval for use within 90 days from the receipt of the request. The department shall consider in its decision whether the covered programs that would be the recipients of the data generated have a demonstrated need for allowing the alternative method.

(5) The department may charge a fee under s. 299.11 (5) (d), Stats., if it is necessary to verify the results of any validation data submitted by a laboratory requesting use of an alternative method.

NR 149.43 Laboratory facilities. (1) The laboratory shall ensure that the environmental conditions of its facility do not adversely affect the required quality of any measurement.

(a) Laboratory facilities shall ensure effective separation between neighboring areas in which incompatible analytical activities take place. The laboratory shall take measures to prevent cross-contamination.

(b) Access to and use of areas affecting the quality of environmental tests shall be controlled to an extent commensurate with the type of analysis and samples analyzed by a laboratory.

(2) The laboratory shall monitor, control, and record environmental conditions when the environmental conditions are required by the methods or when the environmental conditions influence the quality of test results.

NR 149.44 Laboratory equipment. (1) GENERAL PROVISIONS. (a) The laboratory shall furnish the equipment necessary and required for the correct performance of all the environmental tests and associated preparations and activities it performs.

(b) The laboratory shall use equipment and software for testing and calibration that achieves the accuracy required to comply with the requirements of the methods or specifications relevant to the environmental testing performed by the laboratory.

(2) LABORATORY SUPPORT EQUIPMENT. (a) The laboratory shall use support equipment only for its intended purpose, and it shall keep that equipment in working order by routine and preventive maintenance.

(b) When support equipment leaves the direct control of the laboratory for maintenance or for any other reason, the laboratory shall ensure that the function and calibration status of that equipment is checked or demonstrated to be in working order before the equipment is returned to service.

(3) CALIBRATION AND VERIFICATION OF SUPPORT EQUIPMENT. (a) The laboratory shall calibrate or verify all support equipment within that equipment's range of use using available reference materials traceable to NIST. When reference materials traceable to NIST are not commercially available, the laboratory shall use materials of a quality that will ensure the accuracy of the calibrated or verified support equipment for its intended use.

(b) The acceptability criteria for these calibration or verification checks shall be established by the methods, or in the absence, department guidance.

Note: Department guidance can be found on the Wisconsin department of natural resources laboratory accreditation program website.

(c) The laboratory shall establish a procedure for calibrating or verifying the calibration of support equipment which shall include all the following elements:

1. Procedures used for calibrating or verifying the calibration.
2. Procedures for utilization of correction factors when there is a bias.
3. Evaluation criteria used which defensibly documents the continued accuracy of the equipment.
4. Procedures for addressing equipment which fails to meet calibration or verification requirements.

(d) Minimum verification frequencies include all the following:

1. Annually: devices used to measure atmospheric pressure and temperature.
2. Quarterly: mechanical and automatic volumetric dispensing devices, including pipettes.
3. Monthly: balances, with one weight in the expected range of use. Balance weights shall be all the following:

- a. Handled and stored in a manner that protects the weights' integrity.
- b. Traceable to NIST and of class 2 quality or better. Certified for accuracy every five years by a metrology service outside the laboratory. Alternatively, new weights of class 2 quality or better, traceable to NIST, shall be purchased for use. Weight recertification shall be performed sooner than every five years if balance checks performed using these weights suggest that a change in the certified weights has occurred.

Note: Weights that are currently NIST traceable may be used to verify other weights.

4. Each day of use: when specific temperatures are required by method, regulation, or covered program, the operating temperature of the equipment used to control temperatures shall be checked and documented.

(e) All the following are exempt from accuracy verification under this section:

1. Glass microliter syringes.
2. Disposable pipettes.
3. Automatic titrator systems.

(4) LABORATORY ANALYTICAL INSTRUMENTS. (a) The laboratory shall use personnel properly trained to operate analytical instruments. Instructions on the use and maintenance of equipment shall be available to instrument operators.

(b) The laboratory shall properly maintain, inspect, and clean all instruments. The laboratory shall establish procedures for the maintenance of analytical instruments to prevent contamination or deterioration that may affect reported results.

(c) The laboratory shall remove from service all analytical instruments that give suspect results or that have been shown to be defective or outside of performance specifications.

(d) When analytical instruments leave the direct control of the laboratory for maintenance or for any other reason, the laboratory shall ensure that the instruments are functional and that a new initial calibration has passed to demonstrate that the instruments are in satisfactory working order before returned to service.

NR 149.442 Handling of samples. (1) SAMPLE ACCEPTANCE POLICY AND SAMPLE HANDLING PROCEDURES. (a) The laboratory shall have and follow a written policy that clearly outlines the conditions under which samples will be accepted or rejected for analysis or under which associated reported results will be qualified.

Note: s. NR 149.47 (4) provides rejection criteria.

(b) The laboratory shall receive drinking water samples in a secure manner so that the integrity of the sample is maintained.

(c) When samples received do not conform to the descriptions provided by a collector or do not conform to sample acceptance requirements, the laboratory shall consult with the collector or client to determine the proper processing or disposition of the samples.

(d) The laboratory shall place a unique identification code on a sample container as a durable label. The unique identification code shall be used as a link to associate samples with the complete sample history, including treatment and analysis, while in the laboratory's possession.

(2) SAMPLE PRESERVATION AND HOLDING TIME. (a) A laboratory shall follow the sample preservation procedures and holding times required by state and federal regulations.

Note: Sample preservation procedures and holding times are given in 40 CFR Part 136, 40 CFR Part 141, NR 219, SW -846 "Test Methods for Evaluating Solid Waste" and may be specified in the methods.

Note: Links to the 40 CFR Part 136, 40 CFR Part 141, NR 219, and SW -846 can be found on the Wisconsin department of natural resources laboratory accreditation program website.

(b) The laboratory shall measure and document the sample temperature at the time of receipt when temperature preservation is required.

(c) The laboratory shall consider any sample requiring preservation at ≤ 6 °C to be preserved if the sample is received at a temperature greater than its freezing point to 6 °C. When samples are received on the same day that they were collected, the samples may not yet have reached the appropriate temperature by the time they arrived at the laboratory. These samples may be considered acceptable, without the need to qualify the data, if all the following apply:

1. Samples were placed on ice at the time of sample collection.
2. Samples were received at the laboratory on ice. "Blue ice" packs may not be considered as received on ice.

(d) When sample temperature measurements are required, the laboratory shall record any of the following:

1. The temperature of an actual sample.
2. The temperature of a temperature blank shipped with the samples.
3. The temperature of the melt water in the shipping container.

(e) The laboratory shall verify the pH of each bottle received for samples requiring chemical preservation to a specific pH requirement under this section. Bottles not received at the proper pH may be adjusted at the laboratory provided that the methods allow preservation upon receipt and the lab retains documentation of its actions.

Note: pH verification is only required from the bottle that is analyzed.

(3) SAMPLE RECEIPT DOCUMENTATION. The laboratory shall document the receipt and condition of all samples in chronological hard copy or electronic records. The records may be maintained in any format that retains all the following information:

- (a) The identity of the client or entity submitting samples, or the project associated with the received samples.
- (b) The dates of sample collection.
- (c) The times of sample collection for samples to be analyzed for tests with holding times expressed in hours.
- (d) The unique sample identification code assigned by the laboratory.
- (e) Documentation of sample preservation status and other sample conditions on receipt for all sample containers analyzed for those tests for which it is appropriate.
- (f) An unequivocal link between the sample identification code assigned by the laboratory and the field collection identification code assigned by the collector.

(g) The requested analyses, unless the laboratory collects and analyzes its own samples and analyses are directed by permit.

(h) The reference to requested test methods when the collector or sample originator specifies the methods.

(i) Any comments resulting from the inspection undertaken to determine whether samples meet the policy in sub. (2).

(4) STORAGE OF SAMPLES. (a) The laboratory shall have procedures and appropriate facilities for avoiding deterioration, contamination, loss, or damage of samples during storage.

(b) The laboratory shall store samples requiring thermal preservations at ≤ 6 °C at temperatures from greater than the samples' freezing point to 6 °C.

(c) The laboratory shall store samples separately from all standards, reagents, food, and other potentially contaminating sources. Samples shall be stored in areas that prevent or minimize cross-contamination.

(d) The laboratory shall store sample extracts, digestates, leachates, or concentrates resulting from any initial preparatory step as specified in this subsection.

NR 149.444 Initial instrument calibration requirements. (1) GENERAL PROVISIONS. (a) The laboratory shall calibrate or verify the calibration of all analytical instruments before the instruments are used to provide any quantitative results.

(b) Once a calibration model is selected, a calibration function is established, and an initial calibration is finalized, a laboratory may not change the model or calibration function after samples have been analyzed without performing another initial calibration.

(c) The laboratory shall perform an initial calibration if any of the following apply:

1. After instruments undergo non-routine maintenance.

2. Conditions change the expected behavior of the instrument.

3. When a CCV standard fails and any of the following occur:

a. Corrective action taken does not result in a passing CCV standard.

b. A second consecutive (immediate) CCV standard is performed under the same conditions and it also fails and the corrective action taken does not result in two consecutive passing CCV standards.

(d) The laboratory shall retain all the raw data necessary to reconstruct or reproduce calibration functions associated with initial calibrations.

(e) For colorimetric technologies, the laboratory may not use a method blank to zero the instrument.

Note: For colorimetric technologies, other than those based on inverse chemistries, the instrument is to be zeroed with the matrix of interest which is generally reagent water.

(f) The laboratory may not utilize pre-programmed initial calibrations, provided by the instrument manufacturer, for compliance testing.

(g) The laboratory shall include or reference the details of initial instrument calibration procedures including algorithms, any required equations, and acceptance criteria in the method standard operating procedure.

(h) When required by method, the laboratory shall process each calibration standard in the same manner as samples.

(i) Point-to-point calibrations are not allowed unless otherwise specified in this chapter.

(2) MINIMUM NUMBER OF STANDARDS. To establish calibration, the laboratory shall select the number of non-zero standard concentrations that is appropriate for the calibration model selected and the expected range of concentrations. If a method requires analyzing more than three standards to establish a linear calibration, and the laboratory chooses to narrow the calibration range of the determination to no more than two orders of magnitude, the laboratory may use three standards to generate the initial calibration. The minimum number of non-zero standard concentrations selected to establish calibration shall be three except for all the following:

(a) Dissolved oxygen meters, for which the minimum shall be one. Dissolved oxygen meters shall be calibrated against water-saturated air or air-saturated water at a known temperature and pressure. Alternatively, calibration may be performed using an iodometric method.

(b) Conductivity meters, for which the minimum shall be one. Conductivity meters shall be calibrated by verifying the cell constant or adjusting the meter based on the analysis of a potassium chloride standard solution.

(c) Inductively coupled plasma emission spectrophotometers and inductively coupled plasma mass spectrometers, for which the minimum number shall be one.

(d) pH meters, for which the minimum number shall be two.

(e) Quadratic calibration models, for which the minimum shall be five.

(f) Cubic calibration models, for which the minimum shall be seven.

(3) CONCENTRATION LEVELS OF STANDARDS. The concentration of the standards chosen to establish a calibration function shall be within the same orders of magnitude as the expected concentration of samples.

(4) CALIBRATION MODELS. The laboratory shall select a calibration model that is appropriate for the expected behavior of the analytical instrument to be calibrated. To generate a calibration model, the laboratory shall select a

reduction technique or algorithm that is appropriate for the calibration model and the number of non-zero standards used, subject to all the following:

(a) The selected algorithm or reduction technique shall be describable mathematically and shall provide equations, coefficients, or parameters necessary to characterize the calibration function uniquely, unless an analytical instrument is tuned to conform to a universally accepted scientific law or scale.

Note: The response of dissolved oxygen meters is generally adjusted to conform to the concentration of oxygen allowable in a given liquid at a specified temperature and pressure. The response of an ion selective electrode is generally tuned to conform to the Nernst equation. The response of a pH meter is tuned to conform to the universally accepted pH scale. When these instruments are adjusted or tuned according to these principles, characterizing the calibration reduction algorithm mathematically is not necessary.

(b) Non-linear functions may not be used to compensate for instrument saturation, insensitivity, or malfunction.

(c) The laboratory may use weighted algorithms, unless the weighted algorithms are chosen to compensate for deviations from the expected behavior of a detector of an analytical instrument resulting from saturation, insensitivity, or malfunction.

(d) Except for methods that allow average response factors and average calibration factors, the laboratory may not use reiterative reduction techniques or algorithms that force calibration functions through zero.

Note: Reiterative reduction techniques or algorithms that force the calibration function through zero obtain mathematically, by repeated application, a null response for a zero standard that has a non-zero response or adjust calibration parameters to obtain a theoretical null response without analysis of a calibration blank. This paragraph does not prohibit the use of average calibration or response factors or automatic zeroing as part of an initial calibration, when methods, regulations, or covered programs allow those techniques.

(5) EXCLUDING CALIBRATION POINTS. If one or more calibration standards are excluded from the calibration, all the following criteria shall be met:

(a) The rationale for the exclusion is documented.

(b) Any required regulatory limits can still be met.

(c) Except for ICP, ICP/MS, and HRGC/MS, if the highest calibration standard is removed, the linear range shall be limited to the remaining high standard concentration.

(6) EVALUATING ALGORITHM VALIDITY. The laboratory shall establish acceptability criteria for initial calibrations. The type of criteria chosen, and the acceptance range shall be appropriate for the type of analytes to be quantitated, the calibration model selected, and reduction technique or algorithm chosen. Acceptability criteria shall be established using any of the following:

(a) When the x-intercept is used to evaluate the calibration, then the value of the x-intercept of the calibration function for each analyte may not exceed its LOD.

(b) Unless otherwise specified by the method, when RSE is used to evaluate the calibration, the relative standard deviation may not exceed 15% for inorganic analytes or 20% for organic analytes.

(c) Unless otherwise specified by the method, when residuals of each calibration standard are used to evaluate the calibration, the standard recovery for all but the lowest calibration point shall fall within 90% to 110% for

inorganic analytes or within 70% to 130% for organic analytes. Recovery for the lowest calibration point shall be within 80% to 120% for inorganic analytes or 50% to 150% for organic analytes.

(d) When average response factors are used to reduce calibration data, the relative standard deviation of the response factors may not exceed 20% unless the method allows a larger percentage.

(e) When linear regression or least squares analysis is used to reduce calibration data, the correlation coefficient (r) of the resultant calibration shall be at least 0.995 for inorganic analytes or 0.99 for organic analytes.

(f) When quadratic (2nd order) or cubic (3rd order) analysis is used to reduce calibration data, the coefficient of determination (r^2) of the resultant calibration shall be at least 0.995 for inorganic analytes or 0.99 for organic analytes.

(7) VERIFYING ACCURACY. Except for calibrations generated using dissolved oxygen meters, pH meters, or conductivity meters, the laboratory shall verify all initial instrument calibrations after the calibrations are generated, but before the calibrations are used to quantitate any samples, with a second source standard, referred to as an ICV standard. ICV standards shall be treated in the same manner as the standards analyzed for the initial calibration. Unless otherwise required by method, regulation, or covered program, the acceptance criteria for the ICV standard shall be all the following:

(a) Obtaining concentrations within 10% of the theoretical concentrations of all reportable inorganic analytes.

(b) Obtaining concentrations within 20% of the theoretical concentrations of all reportable organic analytes.

(8) EVALUATING SENSITIVITY. When methods require an ICB be analyzed after the initial calibration, the ICB shall be treated in the same manner as the initial calibration standards. The concentration of an analyte in an ICB may not exceed its LOD.

NR 149.446 Continuing instrument calibration requirements. (1) GENERAL PROVISIONS. When an initial instrument calibration is not performed on the day of analysis, the continuing validity of the initial calibration shall be verified prior to analyzing any batch quality control or environmental samples by the analysis of one or more CCV standards, subject to all the following:

(a) Except for multi-peak analytes, CCV standards shall contain all analytes to be reported and may be prepared from the same standards used to generate the initial calibration. CCV standards are required for multi-peak analytes when the analytes are detected and reported in the samples.

(b) CCV standards shall be treated the same as the standards used in the initial calibration. When the method requires that the standards be treated the same as samples, the CCV standards shall be performed with the associated batch so that the CCV standards and samples are all processed together.

(c) Continuing calibration verification is not required for technologies when there are no initial calibrations established.

(d) If an LCS also serves as a CCV standard, the acceptance criteria of the CCV standard shall be used.

(2) FREQUENCY. (a) Continuing calibration verification shall be performed at least once on each analysis day when an initial calibration is not performed and prior to sample analysis and batch quality control analysis.

(b) Continuing calibration verification shall be performed after the consecutive analysis of each group of 20 environmental samples, if 20 or more samples constitute an analytical batch, unless otherwise required by method, regulation, or covered program.

(3) MINIMUM NUMBER OF STANDARDS AND CONCENTRATION LEVELS. (a) For linear and quadratic model calibration functions, the laboratory shall analyze at least a single CCV standard. The concentration of the standard shall be within the range established during the initial calibration.

Note: Linear calibration models include electrometric technologies (pH and ion selective electrode), average response factor, average calibration factor, linear regression, and least squares analysis.

(b) For cubic model calibration functions or third order polynomials, the laboratory shall analyze at least two CCV standards in each instance when a single CCV standard is required by method, regulation, or covered program.

(4) VERIFYING ACCURACY. (a) Unless otherwise required by method, regulation, or covered program, the acceptance criteria for CCV standards shall be within 10% of the theoretical concentrations of all reportable inorganic analytes from an initial calibration.

(b) Unless otherwise required by method, regulation, or covered program, the acceptance criteria for CCV standards shall be within 20% of the theoretical concentrations of all reportable organic analytes from an initial calibration.

(5) ACCURACY CORRECTIVE ACTION. (a) When a CCV standard fails, the laboratory shall do any of the following:

1. Perform corrective action and reanalyze the CCV standard. If the CCV standard does not pass, an initial calibration shall be performed.

2. Perform a second consecutive (immediate) CCV standard under the same conditions. If the second CCV standard also fails, then corrective action shall be performed and two consecutive CCV standards shall pass or an initial calibration shall be performed.

(6) EVALUATING SENSITIVITY. When the method requires that the standards be treated the same as the samples and when the method requires a CCB, the CCB shall be performed with the associated batch so that the CCB and samples are all processed together. The CCB is processed at the same frequency as the CCV standard. The CCB is subject to the same criteria specified in s. NR 149.48 (5) (d).

NR 149.45 Measurement traceability. The laboratory shall maintain all analytical and technical records containing raw and derived data or original observations necessary to allow historical reconstruction of all laboratory activities that contributed to generating reported results. Observations, data, and calculations shall be recorded at the time they are made. At a minimum, the laboratory shall ensure that results of analyses can be linked

to sample collection data, preparation records, calibration records, analytical records, test reports, corrective action, and any chemicals used.

NR 149.47 Reporting results. (1) GENERAL PROVISIONS. (a) The laboratory shall report results of each test performed by the laboratory in accordance with any requirements or instructions specified in the methods or by the department.

(b) The laboratory shall quantitate sample results only from initial instrument calibrations, unless otherwise allowed by method, regulation, or covered program or unless any of the following applies:

1. Samples analyzed by inductively coupled plasma emission spectrophotometers and inductively coupled plasma mass spectrometers having responses at or greater than 90% of the established upper limit of the linear dynamic range of the instruments shall be diluted and reanalyzed.

2. When an analyte does not perform as well as most of the analytes in a multi-analyte initial calibration, analysis may proceed, and results reported for these analytes, provided that the results are appropriately qualified as required in this section.

(c) When samples cannot be diluted and reanalyzed, the laboratory shall report sample results with appropriate qualifiers.

(d) The laboratory shall establish procedures for reporting results for samples analyzed by dual column and dual detector systems. These procedures shall establish all the following prior to analysis:

1. A primary column or primary detector from which results shall be reported.

2. The conditions under which a presumptive identification is confirmed and reported from the secondary column or detector.

(e) When results are greater than the LOQ on dual column or dual detector systems, and the RPD exceeds 40%, then the higher of the two results shall be reported unless the analyst defensibly documents that the higher result is biased due to interference. In this case the laboratory may report the lower result with a qualifier indicating the value of the higher result or report both results.

(f) Excluding microbiological results, MCL exceedances for any regulated analyte associated with ch. NR 809 compliance monitoring shall be reported by the laboratory to the affected water supply facility within 48 hours of completing sample results.

Note: Laboratories performing bacteriological testing for a covered program are certified or approved under ch. ATCP 77 by the department of agriculture, trade, and consumer protection.

(2) FORMAT AND CONTENT. (a) Laboratory test reports shall have formats that facilitate reviewing the content elements specified in this section, unless otherwise provided by pars. (b), (c) and (d). Content elements may be presented in any form, including electronic media.

(b) When tests are performed for internal clients or when a laboratory has a written agreement with a client, the laboratory may issue reports without all the content elements specified in this section. The laboratory shall retain and make available to the department, upon request, records that include the content elements specified in this section.

(c) A laboratory that is operated by a facility whose function is to provide data to monitor the facility's compliance with covered programs shall retain and make available to the department, upon request, records that include the content elements specified in this section. Laboratory reports with all the content elements specified in this section are not required to be issued if any of the following apply:

1. The laboratory is responsible for preparing regulatory reports in a specified format to the department.
2. The laboratory provides information to another individual within the facility for preparation of regulatory reports in a specified format to the department.

(d) Unless otherwise specified by the department, for covered programs that receive data on behalf of a facility, directly from a laboratory, or when provided by pars. (b) and (c), test reports from the laboratory shall include all the following information.

1. The name of the laboratory where the tests were performed.
2. The laboratory's accreditation identification number.
3. The sample identifying information provided by the client or collector.
4. Identification of the methods used for preparation and analysis.
5. The collection date of the samples.
6. Collection, preparation, and analysis times for tests with holding times expressed in hours.
7. The dates of analysis, extraction, or digestion, when a holding time has been established for the preparation step.
8. When non-aqueous sample results are reported, the laboratory shall indicate whether the non-aqueous sample results were reported on a dry weight or wet weight basis.
9. The LOD and LOQ for tests which the department requires reporting to the LOD.
10. Except for HRGC/MS analysis, for sample results requiring adjustments, an indication of whether the LOD and LOQ have been adjusted accordingly.

Note: Sample adjustments are any sample dilutions or sample amounts that were used differently than those used in the initial demonstration of capability and MDL studies.

11. The units of measurement.
12. The date of the test report.
13. Any qualifiers with reported results.
14. The identity of the subcontract laboratory, for each reported result generated by a subcontract laboratory.

(3) AMENDMENTS TO LABORATORY TEST REPORTS. (a) A laboratory may make amendments to a test report already issued by the laboratory in a manner that clearly identifies the reasons for the amendment and that references the original laboratory test report.

(b) Amended reports shall comply with the requirements of this section.

(4) SAMPLE REJECTION OR QUALIFICATION OF RESULTS. The laboratory shall handle results for samples received with insufficient volume to complete the requested analyses, samples received beyond holding time, samples received improperly preserved, samples received in inappropriate containers, or samples received showing evidence that the samples have not been collected according to approved procedures as follows:

(a) Drinking water samples shall be rejected for analysis unless the laboratory has documented instructions from the client to proceed with analyses and all reported results are accompanied by a disclaimer attesting that the results may not be used to determine or evaluate compliance with the safe drinking water act.

(b) Non-drinking water samples shall be rejected for analysis or appropriately qualified.

(5) SAMPLES REQUIRING REANALYSIS OR QUALIFICATION OF RESULTS. Samples shall be re-analyzed, or the affected sample results qualified when any of the following occur:

(a) The concentration of an analyte in the ICB exceeds its LOD.

(b) A CCV standard exceeds limits.

(c) The concentration of an analyte in the CCB or method blank exceeds the criteria specified in s. NR 149.48 (5) (d).

(d) An LCS exceeds limits.

(e) Surrogates or internal standard recoveries exceed limits, unless the failures result from matrix interference, reanalysis is not required but the laboratory shall qualify the results of the affected samples.

(f) When reporting results to the LOD, the concentration of each non-spiked target element in an interference check standard exceeds 10/3 their corresponding LOD for ICP analysis.

Note: The examples for qualifying data listed in this section are common situations. Other situations may exist that could require qualification of data.

NR 149.48 Quality control requirements for chemical testing. (1) GENERAL REQUIREMENTS. (a) A laboratory shall establish a quality control program that includes the analysis of appropriate quality control samples and quality control procedures that define their practices. These quality control procedures shall be used to assess all the following:

1. The level of background contamination associated with the preparation and analysis of all samples.

2. The sensitivity of all tests performed.
3. The level of control of an entire analytical system.
4. The bias contributed to sample results by all preparation and analysis steps.
5. The reproducibility of test results.
6. The selectivity of test methods.

(b) A laboratory may not adjust or correct the sample results by the recoveries of associated quality control samples or surrogates unless otherwise allowed by method, regulation, or covered program. A laboratory may not subtract analyte concentrations found in method blanks from sample results unless otherwise allowed by method, regulation, or covered program.

(c) A laboratory shall document deviations from the laboratory's quality system or exceedances of quality control samples. To the extent the department's data systems allow, the deviations shall be communicated with the results.

(d) A laboratory shall establish procedures for identifying and documenting preparation batches that facilitate determining compliance with the frequencies of quality control samples required under this chapter.

(2) LOD. (a) A laboratory shall determine the LOD for all tests performed except for any of the following:

1. Biochemical oxygen demand and carbonaceous biochemical oxygen demand.
2. Tests for which analyzing a fortified sample is impossible or impractical.
3. Titrimetric tests.
4. Gravimetric tests, other than oil and grease as HEM.

(b) A laboratory shall determine the LOD of an analyte annually by 40 CFR, Part 136, Appendix B. All sample-processing steps of a method shall be included in the determination of a LOD.

Note: Links to the 40 CFR Part 136, Appendix B can be found on the Wisconsin department of natural resources laboratory accreditation program website.

(c) The LOD shall meet the regulatory limits required by the covered programs.

Note: Exemptions to LOD requirements for specific compounds are provided on the Wisconsin department of natural resources laboratory accreditation program website.

(d) The LOD shall be adjusted when the sample amounts used are different than those used for the LOD determination.

(e) For tests exempted from performing an LOD under (a) above, the laboratory shall establish a reporting limit, or an estimate of a test's sensitivity based on the intended use of the data for a given application.

(f) The LOD shall be determined each time there is a change in a method or instrumentation that affects the sensitivity of an analysis.

(g) For HRGC/MS technology, the estimated detection limit is defined in SW-846 8290A and is equivalent to the LOD.

(3) LOQ. (a) A laboratory shall establish the LOQ for all tests performed except for those exempted from an LOD under s. NR 149.48 (2) (a).

(b) The LOQ shall meet the regulatory limits required by the covered programs.

Note: Exemptions to LOQ requirements for specific compounds are provided on the Wisconsin department of natural resources laboratory accreditation program website.

(c) Except for ICP and ICP/MS single point initial calibrations, the LOQ shall be established as 10/3 the LOD or at the concentration of the lowest standard in the initial calibration. For ICP and ICP/MS, when single point initial calibrations are utilized, the LOQ shall be established as 10/3 the LOD or at the “lower limit of quantitation.”

Note: The “lower limit of quantitation” is referenced in SW-846 6010C, 6010D, 6020A, and 6020B.

(d) The LOQ shall be greater than the LOD.

(4) REPORTING LIMITS. (a) Reporting limits are reserved for those analytes exempted under s. 149.48 (2) (a) and shall be established based on a test’s sensitivity and the intended use of the data.

(b) For biochemical oxygen demand and carbonaceous biochemical oxygen demand, the minimum reporting limit is 2 mg/L which is based on a 300 mL sample volume. When no dilution is equal to 300 mL, the reporting limit shall be adjusted based on the lowest dilution reported.

(c) For total suspended solids, the reporting limit shall be determined using the following formula: Reporting Limit (mg/L) = 1000 / (sample volume filtered in mL).

(5) METHOD BLANK. (a) The laboratory shall process method blanks along with and under the same conditions, including all sample preparation steps, as the associated samples in a preparation batch.

Note: Method blanks are not required for analysis of pH, alkalinity, acidity, conductivity, and solids determinations.

(b) The laboratory shall process method blanks at a frequency of at least one per preparation batch up to 20 environmental samples. When samples are analyzed by methods that do not require a preparation step before analysis, a method blank shall be analyzed at the frequency of one per analytical batch up to 20 environmental samples.

(c) Whenever the concentration of the method blank contains analytes of interest greater than the LOD, the laboratory shall evaluate the nature of the interference and its effect on each sample in a preparation batch.

(d) The acceptance criteria for method blanks are analyte and sample specific and are established based on the highest of any of the following:

1. The LOD.
2. Five percent of the regulatory limit for that analyte.

3. Ten percent of the measured concentration in the sample.

(6) LCS. (a) Unless otherwise exempted by this subsection, the laboratory shall process an LCS at a frequency of at least one sample per preparation batch up to 20 environmental samples, along with and under the same conditions as the associated samples in a preparation batch. These conditions shall include all sample preparation steps, except for leaching procedure extractions.

Note: TCLP leachates for metals analysis are fortified after the leaching step is completed and before acid preservation.

(b) The laboratory shall fortify the LCS for the biochemical oxygen demand and carbonaceous biochemical oxygen demand tests with a mixture of glucose-glutamic acid as specified in approved methods of analysis. The LCS shall be processed at a frequency of at least one sample per analytical batch for a laboratory that analyzes more than 20 samples per week. A laboratory that analyzes fewer than 20 samples per week shall analyze one LCS per week.

(c) The laboratory is not required to process an LCS for tests for which analyzing a fortified sample is impossible or impractical.

Note: An LCS need not be analyzed for the following tests: pH, solids determinations, chlorophyll a, and color.

(d) The LCS shall be fortified with the analytes specified by method, regulation, or covered program or all reported analytes, except as allowed in par (e).

(e) For analyses of polychlorinated biphenyls, the laboratory shall fortify an LCS with at least one aroclor per preparation batch. For other tests that determine analytes with responses that encompass more than one chromatographic peak, as in the case of toxaphene and chlordane, the laboratory may fortify an LCS with a single multi-peak analyte per preparation batch. The laboratory shall ensure that all multi-peak analytes detectable by a method are fortified in an LCS at least once every year that any of those analytes are reported at a detectable concentration.

(f) When the method, regulation, or covered program do not specify control limits, the laboratory shall evaluate LCS recoveries and generate in-house control limits, following exclusion of outliers with a statistical technique and using the mean plus or minus 3 times the sample standard deviation. Annually, the laboratory shall review its generated in-house control limits and update those limits whenever the performance characteristics change.

(g) In lieu of using generated in-house control limits for the LCS, the laboratory may opt to use the CCV standard limits.

(7) SELECTIVITY. The laboratory shall establish procedures to confirm the detections of organic analytes determined by technologies that, unlike mass spectrometry or diode array liquid chromatography, do not provide a positive unique identification when a covered program requires it or when the history of a sample source does not suggest the likely presence of the detected analyte.

(a) The laboratory shall develop and document acceptance criteria, which consider retention time shifts, for chromatographic retention time windows.

(b) The laboratory shall document acceptance criteria for mass spectral tuning.

NR 149.50 Technology requirements. The purpose of this section is to establish minimum requirements that can significantly affect data quality but are not always clearly or consistently addressed in all approved methods.

Note: The Department will take the applicability of these requirements into consideration when there are new approved methods or advancements in technology.

(1) OXYGEN DEMAND ASSAYS (BOD OR CBOD). (a) The environmental conditions for the analysis of biochemical oxygen demand and carbonaceous biochemical oxygen demand shall be 17 to 23 °C.

(b) When dissolved oxygen meters are calibrated using a water-saturated air or air-saturated water standard, the laboratory shall verify concentrations in mg/L of those standards by comparing those concentrations to the dissolved oxygen theoretical saturation point. The measured concentration shall be at or near the theoretical saturation point.

(c) The laboratory shall use the theoretical saturation point, based on temperature and barometric pressure, on each day of analysis to assess supersaturation.

Note: When barometric pressure and temperature measurement features are available on the DO meter, they should be taken from the DO meter.

(d) The laboratory shall properly treat supersaturated samples before an initial dissolved oxygen measurement is performed.

(e) When the laboratory uses pipets to deliver sample volumes, the tips shall be manufactured to be wide-bore.

(f) When the laboratory analyzes multiple method blanks and glucose-glutamic acid standards in an analytical batch, each method blank and glucose-glutamic acid standard analyzed shall be assessed individually and associated to the entire analytical batch unless individual method blanks and individual glucose-glutamic acid standards are clearly documented to be traceable to specific groups of 20 samples.

(g) The laboratory shall seed disinfected samples and nitrogenous demand inhibited samples.

(h) The laboratory may not add nitrogenous demand inhibitor to the glucose-glutamic acid standard, to seed material, or method blanks.

(i) The laboratory shall use sample volumes for dilutions that are sufficient to expect 2 mg/L depletion in at least one dilution.

(j) When equipment with multiple dissolved oxygen probes is employed, the laboratory shall calibrate each probe. Sample records shall be traceable to the probe used.

(k) The laboratory shall calibrate dissolved oxygen probes on each day of use.

(l) The laboratory shall use local barometric pressure which has not been adjusted to sea level.

(m) When determining residual chlorine, a minimum detection capability of 0.1 mg/L shall be met.

(2) COLORIMETRIC OR TURBIDIMETRIC. (a) Except for inverse chemistries, the laboratory shall use calibration blanks in the initial calibration of colorimetric or turbidimetric analyses, and those calibration blanks shall be assigned the measured response.

Note: High range chemical oxygen demand and hexavalent chromium are two tests where inverse chemistries are utilized.

(b) When closed vials are digested using block digesters for total phosphorus, the laboratory shall perform the digestion at 150 ± 2 °C for a minimum of 30 minutes.

(c) When the laboratory uses sulfide strips, the sulfide strips shall have a minimum detection capability of 10 mg/L.

(d) The laboratory may not dilute samples after the color reagent has been added to the samples.

(e) The laboratory shall process hexavalent chromium standards the same as samples.

(3) ELECTROMETRIC ASSAYS (I.E. ION-SELECTIVE ELECTRODE). When the laboratory performs electrometric assays, the laboratory shall perform an initial calibration each day of analysis.

(4) GRAVIMETRIC ASSAYS – RESIDUE (SOLIDS). (a) The laboratory may not use Buchner funnels or Gooch crucibles for determination of total suspended solids or total dissolved solids.

(b) When the laboratory uses pipets to deliver sample volumes for total solids and total suspended solids, the pipet tips shall be manufactured to be wide-bore.

(5) GRAVIMETRIC ASSAYS – OIL & GREASE AS HEXANE EXTRACTABLE MATERIALS (HEM). (a) When using the solid phase extraction technique, the laboratory may not allow polar solvents to contact the sample.

(b) The laboratory shall use activated silica gel for silica gel-treated determinations.

(6) TITRIMETRIC OR POTENTIOMETRIC TITRATION ASSAYS. When standardization is required by method, the laboratory shall standardize all titrants monthly, unless all the following are met:

1. Unused titrant is never poured back into the original container.

2. Titrants shall always be protected from light.

3. LCS recovery control limits shall be set at 90 to 110%, or tighter, and the recovery is achieved.

(7) NONDISPERSIVE IR OR MICROCOULOMETRY. (a) For total organic carbon determinations, the laboratory shall perform an inorganic carbon removal check with each analysis batch.

(b) For aqueous samples with results greater than or equal to the LOQ, the laboratory shall perform duplicate injections until the relative percent difference is 10% or less.

(8) ION CHROMATOGRAPHY (IC). The width of the retention time window that the laboratory uses to make identifications shall be based upon measurements of actual retention time variations of standards over the course of a day unless analyst experience provides for another defensible procedure.

(9) FLAME ATOMIC ABSORPTION SPECTROPHOTOMETRY (FLAA). (a) The laboratory shall perform at least two consecutive readings for all samples, standards, and quality control samples, and the laboratory shall use the average for calculating results.

(b) When sample concentrations are greater than the LOQ, the laboratory shall use a control limit of 10% or less for the relative percent difference between replicate aspirations.

(c) The laboratory shall include the same acid types and concentrations in calibration standards as those used in samples.

(10) GRAPHITE FURNACE ATOMIC ABSORPTION SPECTROPHOTOMETRY (GFAA). (a) The laboratory shall use at least two firings for all samples, standards, and quality control samples.

(b) When sample concentrations are greater than the LOQ, the laboratory shall use a control limit of 10% or less for the relative standard deviation of replicate firings.

(c) When elements are measured at wavelengths lower than 200 nm, the laboratory shall analyze the samples with an instrument equipped with Zeeman background correction or equivalent.

(d) The laboratory shall include the same acid types and concentrations in calibration standards as those used in samples.

(11) COLD VAPOR ATOMIC ABSORPTION SPECTROPHOTOMETRY (CVAA). The laboratory shall ensure that potassium permanganate is present after the two-hour digestion for Hg, or the sample shall be redigested using a smaller sample amount until potassium permanganate remains. Instead, the laboratory could choose to add more potassium permanganate to the affected samples and method blank and digest for an additional two hours.

(12) INDUCTIVELY COUPLED PLASMA EMISSION SPECTROPHOTOMETRY (ICP). (a) The laboratory shall perform a spectral interference identification study before performing any sample analysis using the following single element standards: Ag, Al, As, B, Ba, Be, Ca, Cd, Ce, Co, Cr, Cu, Fe, K, Mg, Mn, Mo, Na, Ni, Pb, Sb, Se, SiO₂, Sn, Sr, Ti, Tl, V, and Zn. When other interferences have been identified, the laboratory shall perform appropriate spectral interference identification studies for those interferences. The laboratory shall analyze the interfering elements to identify potential interelement interferences for each mode and wavelength used. This requirement applies to each instrument used for analysis.

(b) The concentration of single element standards used in the spectral interference identification study shall be at or greater than the maximum concentrations encountered in samples.

(c) At the beginning of each analysis day, the laboratory shall verify that interference corrections and background corrections are working properly through the analysis of interference check standards. The interference check standards shall include all the identified interferences at the maximum concentrations encountered in samples.

(d) Interference correction is only valid to the concentration tested in the spectral interference identification study. Samples with interferences present greater than the concentrations tested shall be reanalyzed at a dilution, or if the instrument is capable, the laboratory may analyze a single element standard at the level in the sample to demonstrate that the apparent concentration is less than the LOQ; if it is not less than the LOQ, the interelement correction factors may be updated, and all of the associated data can be reprocessed.

(e) When reporting results to the LOD, the concentration of each non-spiked target element in the interference check standard shall be less than 10/3 their corresponding LOD.

(f) Adjusting background correction shall require re-evaluation of any interference corrections that are affected by the element to which the background correction was made.

(13) INDUCTIVELY COUPLED PLASMA–MASS SPECTROMETRY (ICP/MS). (a) Only those masses listed in methods may be used for identification and quantitation unless the laboratory has supporting data that identifies the potential interfering species for the masses used, and the correction equations needed to resolve the interferences are employed.

(b) All quality control samples shall be performed on the isotope used for identification and quantitation.

(14) GAS CHROMATOGRAPHY (GC). (a) For non-aqueous volatiles analysis, the laboratory shall ensure that the calibration standards contain the same preservative type as the samples, such as methanol, sodium bisulfate, and reagent water.

(b) When the laboratory analyzes multi-peak compounds, such as aroclors, toxaphene, and technical chlordane, the laboratory shall document in its standard operating procedures all the following:

1. For each compound reported, the process used to determine which peaks are used to identify and quantitate the compound.

2. For each compound reported, the process used to determine how the laboratory quantitates the compounds when the compound exhibits weathering, degradation, or positive interferences.

3. For aroclors, the process used to determine how the laboratory quantitates each aroclor when more than one aroclor is present in the sample.

(15) GAS CHROMATOGRAPHY–MASS SPECTROMETRY (GC/MS). (a) The laboratory shall meet full scan tune requirements before selective ion monitoring analysis begins.

(b) For non-aqueous samples, the laboratory shall ensure that the calibration standards shall contain the same preservative type as the samples, such as methanol, sodium bisulfate, and reagent water.

(16) HAZARDOUS WASTE CHARACTERISTICS. (a) The laboratory shall stir samples during pH measurements for toxicity characteristic leaching procedure fluid type determinations.

(b) The laboratory shall perform a flashpoint standard suitable for ignitability determinations for each batch of samples analyzed for flashpoint analysis.

(17) PREPARATORY METHODS. (a) Unless otherwise required by the method, the laboratory shall fortify any quality control sample prior to the addition of the preparation reagents.

(b) The laboratory shall perform microwave preparations with instruments that utilize temperature feedback control.

APPENDIX I

TABLE 1A: List of analytes in aqueous and non-aqueous matrices by class and technology.

Analytes are available in both the aqueous and non-aqueous matrices unless identified by footnote.

Oxygen Demand Assays (BOD or cBOD) Technology

Class: General Chemistry			
	Biochemical Oxygen Demand (BOD) ¹	Carbonaceous Biochemical Oxygen Demand (cBOD) ¹	

Colorimetric or Turbidimetric Technology

Class: General Chemistry			
	Alkalinity ¹	Fluoride	Phosphorus, Total
	Ammonia as N	Hardness, Total as CaCO ₃ ¹	Silica ¹
	Chemical Oxygen Demand (COD) ¹	Kjeldahl Nitrogen, Total	Sulfate
	Chloride	Nitrate	Sulfide
	Chlorine, Total Residual (TRC) ¹	Nitrate + Nitrite	Surfactants ¹
	Chlorophyll ¹	Nitrite	Turbidity ¹
	Cyanide, Available	Orthophosphate	
	Cyanide, Total	Phenolics, Total	
Class: Metals			
	Chromium, Hexavalent		

Electrometric Assays (i.e. ion-selective electrode) Technology

Class: General Chemistry			
	Ammonia as N	Fluoride	pH
	Chloride	Kjeldahl Nitrogen, Total	Specific Conductance
	Chlorine, Total Residual (TRC) ¹	Nitrate	Sulfide
	Cyanide, Total	Oxygen, Dissolved ¹	

Gravimetric Assays – Residue (solids) Technology

Class: General Chemistry			
	Residue, Filterable (TDS) ¹	Residue, Total	Residue, Volatile, Nonfilterable (TVSS) ¹
	Residue, Nonfilterable (TSS) ¹	Residue, Volatile (TVS)	

Extraction/Gravimetric Assays – Oil & Grease as Hexane Extractable Materials (HEM) Technology

Class: General Chemistry		
	Oil & Grease as Hexane Extractable Material (HEM) ¹	

Titrimetric or Potentiometric Titration Assays Technology

Class: General Chemistry			
	Acidity as CaCO ₃ ¹	Chloride	Kjeldahl Nitrogen, Total
	Alkalinity ¹	Chlorine, Total Residual (TRC) ¹	Sulfide
	Ammonia as N	Cyanide, Available	Sulfides, Acid-soluble and Acid-insoluble
	Bromide	Cyanide, Total	Sulfite ¹
	Chemical Oxygen Demand (COD)	Hardness, Total as CaCO ₃ ¹	Calcium
	Percent Water by Karl Fischer Titration ²		

Flow Injection - Gas Diffusion – Amperometry Technology

Class: General Chemistry			
	Cyanide, Available ¹	Cyanide, Total ¹	

Nondispersive Infrared (NDIR) or Microcoulometry Technology

Class: General Chemistry			
	Organic Halides (TOX and AOX)		
	Organic Carbon, Total (TOC)		

Ion Chromatography (IC) Technology

Class: General Chemistry			
	Ammonia as N	Fluoride	Nitrite
	Bromide	Nitrate	Orthophosphate
	Chloride	Nitrate + Nitrite	Sulfate

Flame Atomic Absorption Spectrophotometry (FLAA) Technology

Class: General Chemistry			
	Hardness, Total as CaCO ₃ ¹		
Class: Metals			
	Aluminum	Iridium	Potassium

	Antimony	Iron	Rhodium
	Barium	Lead	Ruthenium
	Beryllium	Lithium	Silver
	Bismuth	Magnesium	Sodium
	Cadmium	Manganese	Strontium
	Calcium	Molybdenum	Thallium
	Chromium, Total	Nickel	Tin
	Cobalt	Osmium	Titanium
	Copper	Palladium	Vanadium
	Gold	Platinum	Zinc

Flame Photometry Spectrophotometry (FP) Technology

Class: Metals

	Calcium	Potassium	Sodium
	Magnesium		

Gaseous Hydride Atomic Absorption Spectrophotometry Technology

Class: Metals

	Antimony	Arsenic	Selenium

Graphite Furnace Atomic Absorption Spectrophotometry (GFAA) Technology

Class: Metals

Aluminum	Gold	Platinum
Antimony	Iridium	Rhodium
Arsenic	Iron	Ruthenium
Barium	Lead	Selenium
Beryllium	Lithium	Silver
Bismuth	Manganese	Thallium
Cadmium	Molybdenum	Tin
Chromium, Total	Nickel	Titanium
Cobalt	Osmium	Vanadium
Copper	Palladium	Zinc

Cold Vapor Atomic Absorption Spectrophotometry (CVAA) Technology

Class: Metals

Mercury	Mercury, Low Level	

Cold Vapor Atomic Fluorescence Spectrophotometry (CVAFS) Technology

Class: Metals

	Mercury	Mercury, Low Level	

Thermal Decomposition Atomic Absorption Spectrophotometry (TDAA) Technology

Class: Metals

	Mercury	Mercury, Low Level	

Inductively Coupled Plasma Emission Spectrophotometry (ICP) Technology

Class: General Chemistry

	Hardness, Total as CaCO ₃ ¹	Phosphorus, Total ²	Silica ¹

Class: Metals

	Aluminum	Iridium	Ruthenium
	Antimony	Iron	Selenium
	Arsenic	Lead	Silicon
	Barium	Lithium	Silver
	Beryllium	Magnesium	Sodium
	Bismuth	Manganese	Strontium
	Boron	Molybdenum	Thallium
	Cadmium	Nickel	Tin
	Calcium	Osmium	Titanium
	Chromium, Total	Palladium	Tungsten
	Cobalt	Platinum	Vanadium
	Copper	Potassium	Zinc
	Gold	Rhodium	Zirconium

Inductively Coupled Plasma - Mass Spectrometry (ICP/MS) Technology

Class: Metals

	Aluminum	Iron	Selenium
	Antimony	Lead	Silicon
	Arsenic	Lithium	Silver
	Barium	Magnesium	Sodium
	Beryllium	Manganese	Strontium
	Bismuth	Mercury	Thallium
	Boron	Molybdenum	Tin
	Cadmium	Nickel	Titanium

	Calcium	Osmium	Tungsten
	Chromium, Total	Palladium	Vanadium
	Cobalt	Platinum	Zinc
	Copper	Potassium	Zirconium
	Gold	Rhodium	
	Iridium	Ruthenium	

Gas Chromatography (GC) Technology

Class: BNA - Phenols

2,3,4,6-Tetrachlorophenol	3,4,5-Trichlorocatechol	4-Chloroguaiacol
2,3,5,6-Tetrachlorophenol	3,4,5-Trichloroguaiacol	4-Chlorophenol
2,4,5-Trichlorophenol	3,4,6-Trichlorocatechol	4-Methylphenol (p-Cresol)
2,4,6-Trichlorophenol	3,4,6-Trichloroguaiacol	4-Nitrophenol
2,4-Dichlorophenol	3,4-Dichlorocatechol	5,6-Dichlorovanillin
2,4-Dimethylphenol	3,4-Dichloroguaiacol	5-Chlorovanillin
2,4-Dinitrophenol	3,6-Dichlorocatechol	6-Chlorovanillin
2,6-Dichlorophenol	3-Methylphenol (m-Cresol)	Dinoseb (2-sec-butyl-4,6-Dinitrophenol)
2,6-Dichlorosyringaldehyde	4,5,6-Trichloroguaiacol	Pentachlorophenol
2-Chlorophenol	4,5-Dichlorocatechol	Phenol
2-Chlorosyringaldehyde	4,5-Dichloroguaiacol	Tetrachlorocatechol
2-Cyclohexyl-4,6-dinitrophenol	4,6-Dichlorocatechol	Tetrachloroguaiacol
2-Methyl-4,6-dinitrophenol	4,6-Dichloroguaiacol	Trichlorosyringol
2-Methylphenol (o-Cresol)	4-Chloro-3-methylphenol (4-Chloro-m-cresol)	
2-Nitrophenol	4-Chlorocatechol	

Class: BNA - Benzidines

3,3'-Dichlorobenzidine	3,3'-Dimethylbenzidine	
3,3'-Dimethoxybenzidine	Benzidine	

Class: BNA - Chlorinated Hydrocarbons

1,2,4,5-Tetrachlorobenzene	1,4-Dichlorobenzene	Hexachlorocyclopentadiene
1,2,4-Trichlorobenzene	Benzyl chloride	Hexachloroethane
1,2-Dichlorobenzene	Hexachlorobenzene	Pentachlorobenzene
1,3-Dichlorobenzene	Hexachlorobutadiene	

Class: BNA - Explosive Residues

1,3,5-Trinitrobenzene	2,4-Dinitrotoluene	Nitrobenzene
1,3-Dinitrobenzene	2,6-Dinitrotoluene	

Class: BNA - Haloethers

4-Bromophenyl phenyl ether	Bis(2-chloroethoxy)methane	Bis(2-chloroisopropyl)ether
4-Chlorophenyl phenyl ether	Bis(2-chloroethyl)ether	

Class: BNA - Nitroaromatics

1,2-Dinitrobenzene	1,4-Dinitrobenzene	Isophorone
1,3-Dinitrobenzene	1,4-Naphthoquinone	Pentachloronitrobenzene (PCNB)

Class: BNA - Nitrosamines

N-Nitrosodiethylamine	N-Nitrosodi-n-propylamine	N-Nitrosomorpholine
N-Nitrosodimethylamine	N-Nitrosodiphenylamine	N-Nitrosopiperidine
N-Nitrosodi-n-butylamine	N-Nitrosomethylethylamine	N-Nitrosopyrrolidine

Class: BNA - Phthalates

Bis(2-ethylhexyl)phthalate	Diethyl phthalate	Di-n-butyl phthalate
Butyl benzyl phthalate	Dimethyl phthalate	Di-n-octyl phthalate

Class: Pesticides - Acid

2,4,5-T	Chloramben	Dinoseb (2-sec-butyl-4,6-Dinitrophenol)
2,4-D	Chlorthal (Dacthal di-acid, DCPA di-acid)	MCPA
2,4-DB	Clopyralid	MCPB
2,4-DB salts and esters	Dalapon	MCPB (Mecoprop)
3,5-Dichlorobenzoic acid	Dicamba	Pentachlorophenol
4-Nitrophenol	Dichlorprop (2,4-DP)	Picloram
5-Hydroxydicamba	Dichlorprop salts and esters	Silvex (2,4,5-TP)
Acifluorfen	Diclofop	Triclopyr

Class: Pesticides - Organochlorine

## PESTICIDES, ORGANOCHLORINE (group)		
4,4'-DDD	Chloroneb	Heptachlor
4,4'-DDE	delta-BHC	Heptachlor epoxide
4,4'-DDT	Dichlone	Isodrin
Aldrin	Dieldrin	Kepone

alpha-BHC	Endosulfan I	Methoxychlor
beta-BHC (β-BHC)	Endosulfan II	Mirex
Captafol	Endosulfan sulfate	Pentachloronitrobenzene (PCNB)
Captan	Endrin	Perthane
Chlordane (alpha)	Endrin aldehyde	Strobane
Chlordane (gamma)	Endrin ketone	Toxaphene
Chlordane (Technical)	gamma-BHC (Lindane)	

Class: Pesticides - Nitrogen

Acetochlor	Chlorothalonil	Norflurazon
Alachlor	Dimethenamid	Pendimethalin
Aspon	Ethalfuralin	Pronamide
Benfluralin	Fenarimol	Propachlor
Bentazon	Hexazinone	Propanil
Bromacil	Isopropalin	Terbacil
Bromoxynil octanoate	Metolachlor	Triadimefon
Butachlor	Metribuzin	Trifluralin
Butylate	Napropamide	

Class: Pesticides - Organophosphorus

Acephate	Dioxathion	Parathion (Parathion ethyl)
Azinphos ethyl	Disulfoton	Parathion methyl
Azinphos methyl (Guthion)	EPN	Phorate
Bolstar	Ethion	Phosalone
Carbophenothion	Ethoprop	Phosmet (Imidan)
Chlorfenvinphos	Famphur	Phosphamidon
Chlorpyrifos	Fenitrothion	Ronnel
Chlorpyrifos methyl	Fensulfothion	Sulfotepp (Tetraethyl dithiopyrophosphate)
Coumaphos	Fenthion	TEPP (Tetraethyl pyrophosphate)
Crotoxyphos	Fonofos	Terbufos
DEF (Butifos)	Hexamethylphosphoramide	Tetrachlorvinphos (Stirofos)
Demeton-O	Leptophos	Thionazin (O,O-Diethyl O-2-pyrazinyl phosphorothioate)
Demeton-S	Malathion	Tokuthion (Prothiofos)
Diazinon	Merphos	Trichloronate
Dichlofenthion	Methamidophos	Trichlorphon

Dichlorvos (DDVP)	Mevinphos	Tri-o-cresylphosphate (TOCP)
Dicrotophos	Monocrotophos	
Dimethoate	Naled	

Class: Pesticides - Triazine

Ametryn	Deethylatrazine	Propazine
Anilazine	Deisopropylatrazine	Simazine
Atraton	Diaminoatrazine	Terbutryn
Atrazine	Prometon	
Cyanazine	Prometryn	

Class: Pesticides - Other

1,2-Dibromo-3-chloropropane (DBCP)	Permethrin	Vapam

Class: Persistent Organic Pollutants

## PCB as AROCLORS (group)		
## PCB CONGENERS (group)		

Class: Volatile Organics

## VOLATILE ORGANICS [VOC] (group)		
1,1,1,2-Tetrachloroethane	Acetone	Isopropyl alcohol (2-Propanol)
1,1,1-Trichloroethane	Acetonitrile	Isopropylbenzene
1,1,2,2-Tetrachloroethane	Acrolein	Malononitrile
1,1,2-Trichloroethane	Acrylonitrile	Methacrylonitrile
1,1-Dichloroethane	Allyl alcohol	Methanol
1,1-Dichloroethylene	Allyl chloride	Methyl acrylate
1,1-Dichloropropene	Benzene	Methyl ethyl ketone (MEK, 2-Butanone)
1,2,3-Trichlorobenzene	Bromoacetone	Methyl methacrylate
1,2,3-Trichloropropane	Bromobenzene	Methyl tert-butyl ether (MtBE)
1,2,4-Trichlorobenzene	Bromochloromethane	Methylene chloride
1,2,4-Trimethylbenzene	Bromodichloromethane	m-Xylene
1,2-Dibromo-3-chloropropane (DBCP)	Bromoform	Naphthalene
1,2-Dibromoethane (EDB)	Bromomethane (Methyl bromide)	n-Butyl alcohol (1-Butanol)

	1,2-Dichlorobenzene	Carbon disulfide	n-Butylbenzene
	1,2-Dichloroethane	Carbon tetrachloride	n-Propylbenzene
	1,2-Dichloroethene (cis)	Chlorobenzene	o-Xylene
	1,2-Dichloroethene (trans)	Chloroethane	Paraldehyde
	1,2-Dichloropropane	Chloroform	p-Isopropyltoluene
	1,3,5-Trimethylbenzene	Chloromethane (Methyl chloride)	Propargyl alcohol
	1,3-Dichloro-2-propanol	Chloromethyl methyl ether	Propionitrile (Ethyl cyanide)
	1,3-Dichlorobenzene	Chloroprene	Propylene glycol
	1,3-Dichloropropane	Crotonaldehyde	p-Xylene
	1,3-Dichloropropylene (cis)	Dibromochloromethane	sec-Butylbenzene
	1,3-Dichloropropylene (trans)	Dibromomethane (Methylene bromide)	β -Propiolactone
	1,3-Propanediol	Dichlorodifluoromethane	Styrene
	1,4-Dichlorobenzene	Diethyl ether (Ethyl ether)	t-Butyl alcohol
	1,4-Dioxane	Epichlorohydrin	tert-Butylbenzene
	2,2-Dichloropropane	Ethanol	Tetrachloroethene
	2,3-Dichloropropene	Ethyl acetate	Toluene
	2-Chloroethanol	Ethyl methacrylate	Trichloroethene
	2-Chloronaphthalene	Ethylbenzene	Trichlorofluoromethane
	2-Chlorotoluene	Ethylene glycol	Vinyl acetate
	2-Hexanone	Ethylene oxide	Vinyl chloride
	2-Pentanone	Hexachlorobutadiene	Xylenes, Total
	4-Chlorotoluene	Iodomethane (Methyl iodide)	
	4-Methyl-2-pentanone (Methyl isobutyl ketone)	Isobutyl alcohol (2-Methyl-1-propanol)	
Class: Solvent Scans			
	Qualitative FID Fingerprint		

Gas Chromatography - Mass Spectroscopy (GC/MS) Technology

Class: Base, Neutral, and Acid Extractable Semivolatile Compounds

SEMIVOLATILES [BNA] (group)

Class: BNA - Phenols

2,3,4,6-Tetrachlorophenol	3,4,5-Trichlorocatechol	4-Chloroguaiacol
2,3,5,6-Tetrachlorophenol	3,4,5-Trichloroguaiacol	4-Chlorophenol
2,4,5-Trichlorophenol	3,4,6-Trichlorocatechol	4-Methylphenol (p-Cresol)

2,4,6-Trichlorophenol	3,4,6-Trichloroguaiacol	4-Nitrophenol
2,4-Dichlorophenol	3,4-Dichlorocatechol	5,6-Dichlorovanillin
2,4-Dimethylphenol	3,4-Dichloroguaiacol	5-Chlorovanillin
2,4-Dinitrophenol	3,6-Dichlorocatechol	6-Chlorovanillin
2,6-Dichlorophenol	3-Methylphenol (m-Cresol)	Benzoic acid
2,6-Dichlorosyringaldehyde	4,5,6-Trichloroguaiacol	Dinoseb (2-sec-butyl-4,6-Dinitrophenol)
2-Chlorophenol	4,5-Dichlorocatechol	Pentachlorophenol
2-Chlorosyringaldehyde	4,5-Dichloroguaiacol	Phenol
2-Cyclohexyl-4,6-dinitrophenol	4,6-Dichlorocatechol	Tetrachlorocatechol
2-Methyl-4,6-dinitrophenol	4,6-Dichloroguaiacol	Tetrachloroguaiacol
2-Methylphenol (o-Cresol)	4-Chloro-3-methylphenol (4-Chloro-m-cresol)	Trichlorosyringol
2-Nitrophenol	4-Chlorocatechol	

Class: BNA - Benzidines

	3,3'-Dichlorobenzidine	3,3'-Dimethylbenzidine	
	3,3'-Dimethoxybenzidine	Benzidine	

Class: BNA - Non-Halogenated Organics

	1,4-Dioxane	Diethyl sulfate	p-Benzoquinone
	1-Acetyl-2-thiourea	Diethylstilbestrol	p-Cresidine
	2-Acetylaminofluorene	Dihydroaffrole	Phenacetin
	2-Aminoanthraquinone	Diphenylamine	Phenobarbital
	2-Hydroxypropionitrile	Ethyl methanesulfonate	Phthalic anhydride
	4-Chloroaniline	Fluchloralin	Piperonyl sulfoxide
	4-Dimethylaminoazobenzene	Hydroquinone	Propylthiouracil
	4-Nitroquinoline 1-oxide	Isosafrole	Pyridine
	5,5-Diphenylhydantoin	Maleic anhydride	Resorcinol
	Acetophenone	Mestranol	Safrole
	Aminoazobenzene	Methapyrilene	TEPP (Tetraethyl pyrophosphate)
	Aniline	Methyl methanesulfonate	Tetraethyl dithiopyrophosphate
	Aramite	Nicotine	Thionazin (O,O-Diethyl O-2-pyrazinyl phosphorothioate)
	Azobenzene	Nitrofen	Thiophenol (Benzenethiol)
	Benzyl alcohol	O,O,O-Triethyl phosphorothioate	Toluene diisocyanate
	Biphenyl	o-Anisidine	Trimethyl phosphate

	Carbazole	Octamethyl pyrophosphoramidate	Tri-p-tolyl phosphate
	Dibenzofuran	o-Toluidine	Tris(2,3-dibromopropyl) phosphate
Class: BNA - Chlorinated Hydrocarbons			
	1,2,4,5-Tetrachlorobenzene	2-Chloronaphthalene	Hexachlorocyclopentadiene
	1,2,4-Trichlorobenzene	3-(Chloromethyl)pyridine Hydrochloride	Hexachloroethane
	1,2-Dichlorobenzene	Benzyl chloride	Hexachlorophene
	1,3-Dichlorobenzene	Chlorobenzilate	Hexachloropropene
	1,4-Dichlorobenzene	Hexachlorobenzene	Pentachlorobenzene
	1-Chloronaphthalene	Hexachlorobutadiene	Pentachloroethane
Class: BNA - Explosives Residues			
	1,3,5-Trinitrobenzene	2-Methyl-3-nitroaniline ¹	3-Nitrotoluene ¹
	1,3-Dinitrobenzene	2-Methyl-5-nitroaniline ¹	4-Methyl-2-nitroaniline ¹
	2,3-Dinitrotoluene ¹	2-Methyl-6-nitroaniline ¹	4-Methyl-3-nitroaniline ¹
	2,4-Dinitrotoluene	2-Nitrotoluene ¹	4-Nitrotoluene ¹
	2,5-Dinitrotoluene ¹	3,4-Dinitrotoluene ¹	5-Methyl-2-nitroaniline ¹
	2,6-Dinitrotoluene	3,5-Dinitrotoluene ¹	Nitrobenzene
Class: BNA - Haloethers			
	4-Bromophenyl phenyl ether	Bis(2-chloroethoxy)methane	Bis(2-chloroisopropyl)ether
	4-Chlorophenyl phenyl ether	Bis(2-chloroethyl)ether	
Class: BNA - Nitroaromatics			
	1,2-Dinitrobenzene	2-Methyl-5-nitroaniline ¹	4-Chloro-1,3-phenylenediamine
	1,3,5-Trinitrobenzene	2-Naphthylamine	4-Chloroaniline
	1,3-Dinitrobenzene	2-Nitroaniline	4-Nitroaniline
	1,4-Dinitrobenzene	2-Picoline (2-Methylpyridine)	4-Nitrobiphenyl
	1,4-Naphthoquinone	3-Amino-9-ethylcarbazole	5-Chloro-2-methylaniline
	1,4-Phenylenediamine	3-Nitroaniline	5-Nitroacenaphthene
	1-Naphthylamine	4,4'-Methylenebis (2-chloroaniline)	5-Nitro-o-anisidine
	2,4,5-Trimethylaniline	4,4'-Methylenebis (N,N-dimethylaniline)	5-Nitro-o-toluidine ¹
	2,4-Diaminotoluene	4,4'-Oxydianiline	a,a-Dimethylphenethylamine
	2,4-Dinitrotoluene	4-Aminobiphenyl	Isophorone

	2,6-Dinitrotoluene	4-Chloro-1,2-phenylenediamine	Nitrobenzene
Class: BNA - Nitrosamines			
	N-Nitrosodiethylamine	N-Nitrosodi-n-propylamine	N-Nitrosomorpholine
	N-Nitrosodimethylamine	N-Nitrosodiphenylamine	N-Nitrosopiperidine
	N-Nitrosodi-n-butylamine	N-Nitrosomethylethylamine	N-Nitrosopyrrolidine
Class: BNA - Polynuclear Aromatic Hydrocarbons			
	## PAH (group)		
	1-Methylnaphthalene	Benzo[a]pyrene	Fluoranthene
	2-Methylnaphthalene	Benzo[b]fluoranthene	Fluorene
	3-Methylcholanthrene	Benzo[g,h,i]perylene	Indeno(1,2,3-cd)pyrene
	7,12-Dimethylbenz(a)-anthracene	Benzo[k]fluoranthene	Naphthalene
	Acenaphthene	Chrysene	Phenanthrene
	Acenaphthylene	Dibenz(a,j)acridine	Pyrene
	Anthracene	Dibenzo[a,e]pyrene	
	Benzo[a]anthracene	Dibenzo[a,h]anthracene	
Class: BNA - Phthalates			
	Bis(2-ethylhexyl)phthalate	Diethyl phthalate	Di-n-butyl phthalate
	Butyl benzyl phthalate	Dimethyl phthalate	Di-n-octyl phthalate
Class: Pesticides - Acid			
	2,4,5-T	Clopyralid	MCPB
	2,4-D	Dalapon	MCPP (Mecoprop)
	2,4-DB	Dicamba	Pentachlorophenol
	4-Nitrophenol	Dichlorprop (2,4-DP)	Picloram
	Acifluorfen	Diclofop	Silvex (2,4,5-TP)
	Bromoxynil (Brominal)	Dinoseb (2-sec-butyl-4,6-Dinitrophenol)	Triclopyr
	Chlorthal (Dacthal di-acid, DCPA di-acid)	MCPA	
Class: Pesticides - Organochlorine			
	## PESTICIDES, ORGANOCHLORINE (group)		
	4,4'-DDD	Chlordane (Technical)	gamma-BHC (Lindane)

	4,4'-DDE	delta-BHC	Heptachlor
	4,4'-DDT	Dichlone	Heptachlor epoxide
	Aldrin	Dieldrin	Isodrin
	alpha-BHC	Endosulfan I	Kepone
	beta-BHC (β-BHC)	Endosulfan II	Methoxychlor
	Captafol	Endosulfan sulfate	Mirex
	Captan	Endrin	Pentachloronitrobenzene (PCNB)
	Chlordane (alpha)	Endrin aldehyde	Toxaphene
	Chlordane (gamma)	Endrin ketone	

Class: Pesticides - Nitrogen

	Acetochlor	Chlorothalonil	Norflurazon
	Alachlor	Dimethenamid	Pendimethalin
	Aspon	Ethalfuralin	Pronamide
	Benfluralin	Fenarimol	Propachlor
	Bentazon	Hexazinone	Propanil
	Bromacil	Isopropalin	Terbacil
	Bromoxynil octanoate	Metolachlor	Triadimefon
	Butachlor	Metribuzin	Trifluralin
	Butylate	Napropamide	

Class: Pesticides – Organophosphorus

	Acephate	Dioxathion	Parathion (Parathion ethyl)
	Azinphos ethyl	Disulfoton	Parathion methyl
	Azinphos methyl (Guthion)	EPN	Phorate
	Bolstar	Ethion	Phosalone
	Carbophenothion	Ethoprop	Phosmet (Imidan)
	Chlorfenvinphos	Famphur	Phosphamidon
	Chlorpyrifos	Fenitrothion	Ronnel
	Chlorpyrifos methyl	Fensulfthion	Sulfotepp (Tetraethyl dithiopyrophosphate)
	Coumaphos	Fenthion	TEPP (Tetraethyl pyrophosphate)
	Crotoxyphos	Fonofos	Terbufos
	DEF (Butifos)	Hexamethylphosphoramide	Tetrachlorvinphos (Stirofos)
	Demeton-O	Leptophos	Thionazin (O,O-Diethyl O-2-pyrazinyl phosphorothioate)
	Demeton-S	Malathion	Tokuthion (Prothiofos)
	Diazinon	Merphos	Trichloronate

	Dichlofenthion	Methamidophos	Trichlorphon
	Dichlorvos (DDVP)	Mevinphos	Tri-o-cresylphosphate (TOCP)
	Dicrotophos	Monocrotophos	
	Dimethoate	Naled	
Class: Pesticides - Triazine			
	Ametryn	Deethylatrazine	Propazine
	Anilazine	Deisopropylatrazine	Simazine
	Atraton	Diaminoatrazine	Terbutryn
	Atrazine	Prometon	
	Cyanazine	Prometryn	
Class: Pesticides - Carbamate			
	Barban	Dazomet	Nabam
	Busan 40	Diallate (cis or trans)	Nabonate
	Busan 85	EPTC (Eptam)	Sulfallate (Thioallate)
	Carbam-S	Ethyl Carbamate	Tebuthiuron
	Carbaryl	KN Methyl	Triallate
	Carbofuran	Mexacarbate	Ziram
Class: Pesticides - Other			
	Endothall	Strychnine	
Class: Persistent Organic Pollutants			
	## PCB as AROCLORS (group)		
	## PCB CONGENERS (group)		
Class: Volatile Organics			
	## VOLATILE ORGANICS [VOC] (group)		
	1,1,1,2-Tetrachloroethane	4-Chlorotoluene	Iodomethane (Methyl iodide)
	1,1,1-Trichloroethane	4-Methyl-2-pentanone (Methyl isobutyl ketone)	Isobutyl alcohol (2-Methyl-1-propanol)
	1,1,2,2-Tetrachloroethane	Acetone	Isopropyl alcohol (2-Propanol)
	1,1,2-Trichloroethane	Acetonitrile	Isopropylbenzene
	1,1-Dichloroethane	Acrolein	Malononitrile
	1,1-Dichloroethylene	Acrylonitrile	Methacrylonitrile

	1,1-Dichloropropene	Allyl alcohol	Methanol
	1,2,3,4-Diepoxybutane	Allyl chloride	Methyl acrylate
	1,2,3-Trichlorobenzene	Benzene	Methyl ethyl ketone (MEK, 2-Butanone)
	1,2,3-Trichloropropane	Bis(2-chloroethyl)sulfide	Methyl methacrylate
	1,2,4-Trichlorobenzene	Bromoacetone	Methyl tert-butyl ether (MtBE)
	1,2,4-Trimethylbenzene	Bromobenzene	Methylene chloride
	1,2-Dibromo-3-chloropropane (DBCP)	Bromochloromethane	m-Xylene
	1,2-Dibromoethane (EDB)	Bromodichloromethane	Naphthalene
	1,2-Dichlorobenzene	Bromoform	n-Butyl alcohol (1-Butanol)
	1,2-Dichloroethane	Bromomethane (Methyl bromide)	n-Butylbenzene
	1,2-Dichloroethene (cis)	Carbon disulfide	n-Propylamine
	1,2-Dichloroethene (trans)	Carbon tetrachloride	n-Propylbenzene
	1,2-Dichloropropane	Chlorobenzene	o-Toluidine
	1,3,5-Trimethylbenzene	Chloroethane	o-Xylene
	1,3-Dichloro-2-propanol	Chloroform	Paraldehyde
	1,3-Dichlorobenzene	Chloromethane (Methyl chloride)	Pentachloroethane
	1,3-Dichloropropane	Chloromethyl methyl ether	p-Isopropyltoluene
	1,3-Dichloropropylene (cis)	Chloroprene	Propargyl alcohol
	1,3-Dichloropropylene (trans)	Crotonaldehyde	Propionitrile (Ethyl cyanide)
	1,3-Propanediol	Dibromochloromethane	p-Xylene
	1,4-Dichloro-2-butene (trans)	Dibromomethane (Methylene bromide)	Pyridine
	1,4-Dichlorobenzene	Dichlorodifluoromethane	sec-Butylbenzene
	1,4-Dioxane	Dichlorofluoromethane	β-Propiolactone
	1-Chlorohexane	Diethyl ether (Ethyl ether)	Styrene
	1-Propanol	Diisopropyl ether	t-Butyl alcohol
	2,2-Dichloropropane	Epichlorohydrin	tert-Butylbenzene
	2,3-Dichloropropene	Ethanol	Tetrachloroethene
	2-Chloroethanol	Ethyl acetate	Tetrahydrofuran
	2-Chloronaphthalene	Ethyl methacrylate	Toluene
	2-Chlorotoluene	Ethylbenzene	Trichloroethene
	2-Hexanone	Ethylene glycol	Trichlorofluoromethane
	2-Nitropropane	Ethylene oxide	Vinyl acetate
	2-Pentanone	Hexachlorobutadiene	Vinyl chloride

	2-Picoline (2-Methylpyridine)	Hexachloroethane	Xylenes, Total
	3-Chloropropionitrile	Hexane, n-	

Liquid Chromatography (LC) Technology

Class: Aldehydes & Ketones			
	Acetaldehyde	Formaldehyde	Octanal
	Acetone	Heptanal	o-Tolualdehyde
	Butanal	Hexanal	Pentanal (Valeraldehyde)
	Crotonaldehyde	Isovaleraldehyde	Propanal (Propionaldehyde)
	Cyclohexanone	m-Tolualdehyde	p-Tolualdehyde
	Decanal	Nonanal	
Class: Pesticides - Acid			
	2,4,5-T	Acifluorfen	Diclofop
	2,4,5-T, butoxyethanol ester	Bromoxynil (Brominal)	Dinoseb (2-sec-butyl-4,6-Dinitrophenol)
	2,4,5-T, butyl ester	Chloramben	MCPA
	2,4-D	Chlorthal (Dacthal di-acid, DCPA di-acid)	MCPB
	2,4-D, butoxyethanol ester	Clopyralid	MCPP (Mecoprop)
	2,4-D, ethylhexyl ester	Dalapon	Pentachlorophenol
	2,4-DB	Dicamba	Picloram
	2,4-DB salts and esters	Dichlorprop (2,4-DP)	Silvex (2,4,5-TP)
	3,5-Dichlorobenzoic acid	Dichlorprop salts and esters	Triclopyr
	4-Nitrophenol		
Class: Pesticides - BNA-Benzidines			
	3,3'-Dichlorobenzidine	Benzidine	
Class: BNA - Non-Halogenated Organics			
	Acrolein	Acrylamide	Acrylonitrile
Class: Pesticides - Carbamate			
	3-Hydroxycarbofuran	Diuron	Monuron
	Aldicarb	Fenuron	Oxamyl (Vydate)
	Aldicarb sulfone	Fluometuron	Promecarb
	Aldicarb sulfoxide	Linuron	Propanil
	Baygon (Propoxur)	m-Cumenyl methylcarbamate	Propham

	Bendiocarb	Methiocarb	Siduron
	Carbaryl	Methomyl	Tebuthiuron
	Carbofuran	Metolcarb	Thiodicarb
	Dioxacarb	Mexacarbate	Triallate
Class: BNA - Explosive Residues			
	1,3,5-Trinitrobenzene	2-Amino-4,6-dinitrotoluene	Nitroglycerin
	1,3-Dinitrobenzene	2-Nitrotoluene	PETN (Pentaerythritol tetranitrate)
	2,4,6-Trinitrobenzene	3-Nitrotoluene	Picric Acid (Trinitrophenol)
	2,4,6-Trinitrotoluene	4-Amino-2,6-dinitrotoluene	RDX
	2,4-Diamino-6-nitrotoluene	4-Nitrotoluene	Tetryl
	2,4-Dinitrotoluene	HMX	
	2,6-Dinitrotoluene	Nitrobenzene	
Class: Metals			
	Mercury	Organomercury	
Class: Pesticides - Nitrogen			
	Bentazon	Bromoxynil (Brominal)	Secbumeton
	Bromacil	Butylate	TCMTB
Class: Pesticides - Organophosphorus			
	Dichlorvos (DDVP)	Fensulfothion	Parathion methyl
	Dimethoate	Merphos	Phorate
	Disulfoton	Monocrotophos	Trichlorphon
	Famphur	Naled	
Class: Polynuclear Aromatic Hydrocarbons			
	## PAH (group)		
	1-Methylnaphthalene	Benzo[a]pyrene	Fluoranthene
	2-Methylnaphthalene	Benzo[b]fluoranthene	Fluorene
	Acenaphthene	Benzo[g,h,i]perylene	Indeno(1,2,3-cd)pyrene
	Acenaphthylene	Benzo[k]fluoranthene	Naphthalene
	Anthracene	Chrysene	Phenanthrene
	Benzo[a]anthracene	Dibenzo[a,h]anthracene	
Class: Pesticides - Other			
	Pyrene	Glyphosate	Pyrethrin II

	Diquat	Paraquat
	Fenvalerate	Pyrethrin I

Class: BNA - Phenols

	Dinoseb (2-sec-butyl-4,6-Dinitrophenol)
--	--

Liquid Chromatography - Mass Spectroscopy (LC/MS) Technology

Class: Pesticides - Acid

	2,4,5-T	2,4-DB salts and esters	Dichlorprop salts and esters
	2,4,5-T, butoxyethanol ester	3,5-Dichlorobenzoic acid	Dinoseb (2-sec-butyl-4,6-Dinitrophenol)
	2,4,5-T, butyl ester	Acifluorfen	MCPA
	2,4-D	Chloramben	MCPP (Mecoprop)
	2,4-D, butoxyethanol ester	Dalapon	Picloram
	2,4-D, ethylhexyl ester	Dicamba	Silvex (2,4,5-TP)
	2,4-DB	Dichlorprop (2,4-DP)	

Class: BNA - Benzidines

	3,3'-Dichlorobenzidine	3,3'-Dimethylbenzidine	Benzidine
	3,3'-Dimethoxybenzidine		

Class: Pesticides - Carbamate

	3-Hydroxycarbofuran	Chloroxuron	Neburon
	Aldicarb	Diuron	o-Chlorophenyl thiourea
	Aldicarb sulfone	EPTC (Eptam)	Oxamyl (Vydate)
	Aldicarb sulfoxide	Fenuron	Pebulate
	Aminocarb	Fenuron-TCA	Propham
	Asulam	Fluometuron	Prosulfocarb
	Barban	Linuron	Siduron
	Baygon (Propoxur)	m-Cumenyl methylcarbamate	Tebuthiuron
	Bendiocarb	Methiocarb	Thiodicarb
	Benomyl	Methomyl	Thiofanox
	Carbaryl	Metolcarb	Thiophanate-methyl
	Carbendazim	Mexacarbate	Triallate
	Carbofuran	Molinate	Vernolate
	Carbosulfan	Monuron	
	Chloroprotham	Monuron-TCA	

Class: Pesticides - Nitrogen			
	Alachlor-ESA (Alachlor ethane sulfonic acid)	Bromacil	Propachlor
	Benzoylprop ethyl	Butylate	
Class: Pesticides - Organophosphorus			
	Dichlorvos (DDVP)	Fensulfothion	Parathion methyl
	Dimethoate	Merphos	Phorate
	Disulfoton	Monocrotophos	Trichlorphon
	Famphur	Naled	Rotenone

High Resolution Gas Chromatography - Mass Spectrometry (HRGC/MS) Technology

Class: Persistent Organic Pollutants			
	## DIOXINS & FURANS (group)		
	## PCB AROCLORS (group)		
	## PCB CONGENERS (group)		

Hazardous Waste Characteristics Technology


Class: Hazardous Waste Characteristics			
	Corrosivity, Toward Steel ²	Ignitability, Setaflash Closed Cup ²	Ignitability, Small Scale Closed Cup ²
	Corrosivity, Liquids ²	Ignitability, Pensky-Martens Closed Cup ²	Toxicity Characteristic Leaching Procedure (TCLP) Extraction ^{2, 3}

Solid Waste Leaching Procedures Technology

Class: Leaching Procedures			
	SPLP Extraction ^{2,3}	Reagent Water Shake Extraction (ASTM Leach) ^{2,3}	EPTOX Extraction ^{2,3}

Whole Effluent Toxicity Assays

Class: Toxicity, Acute			
	Ceriodaphnia dubia ¹	Pimephales promelas ¹	
Class: Toxicity, Chronic			
	Ceriodaphnia dubia ¹	Pimephales promelas ¹	Selenastrum capricornutum ¹



1 = accreditation available in the aqueous matrix only

2 = accreditation available in the non-aqueous only

3 = Leaching extractions require that laboratories also maintain accreditation for any analyte to be determined in the resulting leachate.

TABLE 1B: List of analytes in the drinking water matrix by class and method.

Analyte (group) - Method

Class: Disinfection By-products	
## HALOACETIC ACIDS (5) - EPA 552.1	
## HALOACETIC ACIDS (5) - EPA 552.2	
## HALOACETIC ACIDS (5) - EPA 552.3	
## HALOACETIC ACIDS (5) - EPA 557	
## HALOACETIC ACIDS (5) - SM 6251B	
## HALOACETIC ACIDS (5) - SM 6610B	
## THM (group) - EPA 502.2 ## THM (group) - EPA 524.2 ## THM (group) - EPA 524.3 ## THM (group) - EPA 551.1	
Bromate - ASTM D 6581 Bromate - EPA 300.1 Bromate - EPA 302.0 Bromate - EPA 317.0 Rev. 2.0 Bromate - EPA 321.8 Bromate - EPA 326.0 Bromate - EPA 557	
Bromide - ASTM D 6581 Bromide - EPA 300.0 Bromide - EPA 300.1 Bromide - EPA 326.0 Bromide - EPA 327.0 Rev. 1.1	
Bromodichloromethane - EPA 502.2 Bromodichloromethane - EPA 524.2 Bromodichloromethane - EPA 524.3 Bromodichloromethane - EPA 551.1	
Bromoform - EPA 502.2 Bromoform - EPA 524.2 Bromoform - EPA 524.3 Bromoform - EPA 551.1	
Chlorate - EPA 300.1	
Chlorine Dioxide - EPA 327.0, Rev.1 Chlorine Dioxide - SM 4500-CIO2 C Chlorine Dioxide - SM 4500-CIO2 D Chlorine Dioxide - SM 4500-CIO2 E	

	Chlorite - ASTM D 6581
	Chlorite - EPA 300.0 Chlorite - EPA 300.1 Chlorite - EPA 317.0 Rev. 2.0 Chlorite - EPA 326.0
	Chlorite - EPA 327.0 Rev. 1.1
	Chlorite - SM 4500-CIO2 E
	Chloroform - EPA 502.2 Chloroform - EPA 524.2 Chloroform - EPA 524.3 Chloroform - EPA 551.1
	Dibromochloromethane - EPA 502.2 Dibromochloromethane - EPA 524.2 Dibromochloromethane - EPA 524.3 Dibromochloromethane - EPA 551.1
	Ozone - SM 4500-O3 B
	Class: Primary Inorganics Contaminants; Non-metals
	Cyanide - ALPKEM OIA-77 Cyanide - ASTM D2036 (A) Cyanide - ASTM D2036 (B) Cyanide - ASTM D6888 Cyanide - EPA 335.4 Cyanide - Kelada 01 Cyanide - ME355.01
	Cyanide - QuikChem 10-204-00-1-X
	Cyanide - SM 4500-CN- C,E
	Cyanide - SM 4500-CN- C,F
	Cyanide - USGS I-3300-85
	Cyanide, Amenable - SM4500-CN- C,G
	Fluoride - ASTMD1179 (B) Fluoride - ASTMD4327 Fluoride - ASTMD6508, Rev 2 Fluoride - EPA 300.0 Fluoride - EPA 300.1 Fluoride - HACH Method 10225 Fluoride - SM 4110B Fluoride - SM 4500F- B, D Fluoride - SM 4500F- C Fluoride - SM 4500F- E

Fluoride - Technicon 129-71W Fluoride - Technicon 380-75WE
Nitrate - ASTM D3867 (A) Nitrate - ASTM D3867 (B) Nitrate - ASTM D4327 Nitrate - ASTM D6508, Rev 2 Nitrate - EPA 300.0 Nitrate - EPA 300.1 Nitrate - EPA 353.2 Nitrate - Hach Method 10206 Nitrate - Orion 601 Nitrate - SM 4110B Nitrate - SM 4500-NO3- D Nitrate - SM 4500-NO3- E Nitrate - SM 4500-NO3- F
Nitrate - Systea Easy
Nitrate - Waters B-1011
Nitrate + Nitrite - ASTM D3867 (A) Nitrate + Nitrite - ASTM D3867 (B) Nitrate + Nitrite - ASTM D4327 Nitrate + Nitrite - ASTM D6508, Rev 2 Nitrate + Nitrite - EPA 300.0 Nitrate + Nitrite - EPA 300.1 Nitrate + Nitrite - EPA 353.2
Nitrate + Nitrite - SM 4110B
Nitrate + Nitrite - SM 4500-NO3- E
Nitrate + Nitrite - SM 4500-NO3- F
Nitrate + Nitrite - Waters B-1011
Nitrite - ASTM D3867 (A) Nitrite - ASTM D3867 (B) Nitrite - ASTM D4327 Nitrite - ASTM D6508, Rev 2
Nitrite - EPA 300.0
Nitrite - EPA 300.1
Nitrite - EPA 353.2
Nitrite - SM 4110B
Nitrite - SM 4500-NO2- B
Nitrite - SM 4500-NO3- E
Nitrite - SM 4500-NO3- F

	Nitrite - Systema Easy
	Nitrite - Waters B-1011
	Class: Primary Inorganics Contaminants; Metals
	Antimony - ASTM D3697
	Antimony - EPA 200.5 Axial ICP
	Antimony - EPA 200.8
	Antimony - EPA 200.9
	Antimony - SM 3113B
	Arsenic - ASTM D2972 (B)
	Arsenic - ASTM D2972 (C)
	Arsenic - EPA 200.5 Axial ICP
	Arsenic - EPA 200.8
	Arsenic - EPA 200.9
	Arsenic - SM 3113B
	Arsenic - SM 3114B
	Barium - EPA 200.5 Axial ICP
	Barium - EPA 200.7
	Barium - EPA 200.8
	Barium - SM 3111D
	Barium - SM 3113B
	Barium - SM 3120B
	Beryllium - ASTM D3645 (B)
	Beryllium - EPA 200.5 Axial ICP
	Beryllium - EPA 200.7
	Beryllium - EPA 200.8
	Beryllium - EPA 200.9
	Beryllium - SM 3113B
	Beryllium - SM 3120B
	Cadmium - EPA 200.5 Axial ICP
	Cadmium - EPA 200.7
	Cadmium - EPA 200.8
	Cadmium - EPA 200.9
	Cadmium - SM 3113B
	Chromium - EPA 200.5 Axial ICP
	Chromium - EPA 200.7
	Chromium - EPA 200.8
	Chromium - EPA 200.9
	Chromium - SM 3113B
	Chromium - SM 3120B

	Copper - ASTM D1688 (A)
	Copper - ASTM D1688 (C)
	Copper - EPA 200.5 Axial ICP
	Copper - EPA 200.7
	Copper - EPA 200.8
	Copper - EPA 200.9
	Copper - SM 3111B
	Copper - SM 3113B
	Copper - SM 3120B
	Lead - ASTM D3559 (D)
	Lead - EPA 200.5 Axial ICP
	Lead - EPA 200.8
	Lead - EPA 200.9
	Lead - Palintest 1001
	Lead - SM 3113B
	Mercury - ASTM D3223
	Mercury - EPA 200.8
	Mercury - EPA 245.1
	Mercury - EPA 245.2
	Mercury - SM 3112B
	Nickel - EPA 200.5 Axial ICP
	Nickel - EPA 200.7
	Nickel - EPA 200.8
	Nickel - EPA 200.9
	Nickel - SM 3111B
	Nickel - SM 3113B
	Nickel - SM 3120B
	Selenium - ASTM D3859 (A)
	Selenium - ASTM D3859 (B)
	Selenium - EPA 200.5 Axial ICP
	Selenium - EPA 200.8
	Selenium - EPA 200.9
	Selenium - SM 3113B
	Selenium - SM 3114B
	Thallium - EPA 200.8
	Thallium - EPA 200.9
	Class: Secondary Inorganics Contaminants; Non-metals
	Alkalinity - ASTM D1067 (B)
	Alkalinity - SM 2320B

Alkalinity - USGS I-1030-85
Chloride - ASTM D4327
Chloride - ASTM D512 (B)
Chloride - ASTM D6508, Rev. 2
Chloride - EPA 300.0
Chloride - EPA 300.1
Chloride - SM 4110B
Chloride - SM 4500-Cl- B
Chloride - SM 4500-Cl- D
Chlorine, Combined - ASTMD1253
Chlorine, Combined - SM4500-Cl D
Chlorine, Combined - SM4500-Cl F
Chlorine, Combined - SM4500-Cl G
Chlorine, Free - ASTMD1253
Chlorine, Free - Chlorosense
Chlorine, Free - EPA 334.0
Chlorine, Free - SM4500-Cl D
Chlorine, Free - SM4500-Cl F
Chlorine, Free - SM4500-Cl G
Chlorine, Free - SM4500-Cl H
Chlorine, Total - ASTM D1253
Chlorine, Total - Chlorosense
Chlorine, Total - EPA 334.0
Chlorine, Total - SM 4500-Cl D
Chlorine, Total - SM 4500-Cl E
Chlorine, Total - SM 4500-Cl F
Chlorine, Total - SM 4500-Cl G
Chlorine, Total - SM 4500-Cl I
Conductivity - ASTMD1125 (A)
Conductivity - SM 2510 B
Diss. Org. Carbon (DOC) - EPA 415.3
Diss. Org. Carbon (DOC) - SM5310 B
Diss. Org. Carbon (DOC) - SM5310 C
Diss. Org. Carbon (DOC) - SM5310 D
Foaming agents (MBAS) - SM 5540C
Orthophosphate - ASTMD4327
Orthophosphate - ASTMD515 (A)
Orthophosphate - ASTMD6508, Rev. 2

Orthophosphate - EPA 300.0
Orthophosphate - EPA 300.1
Orthophosphate - EPA 365.1
Orthophosphate - SM 4110B
Orthophosphate - SM 4500-P E
Orthophosphate - SM 4500-P F
Orthophosphate - USGS I-1601-85
Orthophosphate - USGS I-2598-85
Orthophosphate - USGS I-2601-90
pH - ASTM D1293
pH - EPA 150.1
pH - EPA 150.2
pH - SM 4500-H+ B
Sulfate - ASTM D4327
Sulfate - ASTM D516
Sulfate - ASTM D6508, Rev. 2
Sulfate - EPA 300.0
Sulfate - EPA 300.1
Sulfate - EPA 375.2
Sulfate - SM 4110B
Sulfate - SM 4500-SO42- C, D
Sulfate - SM 4500-SO42- E
Sulfate - SM 4500-SO42- F
SUVA (calc.) - EPA 415.3
TDS (Total Dissolved Solids) - SM 2540C
Total Organic Carbon (TOC) - EPA 415.3
Total Organic Carbon (TOC) - SM5310 B
Total Organic Carbon (TOC) - SM5310 C
Total Organic Carbon (TOC) - SM5310 D
Turbidity - AMI Turbiwell
Turbidity - EPA 180.1
Turbidity - GLI Method 2
Turbidity - HACH FilterTrak 10133
Turbidity - Mitchell M5271
Turbidity - Mitchell M5331
Turbidity - Orion AQ4500
Turbidity - SM 2130B
UV254 - EPA 415.3

UV254 - SM5910 B
Class: Secondary Inorganics Contaminants; Metals
Aluminum - EPA 200.5 Axial ICP
Aluminum - EPA 200.7
Aluminum - EPA 200.8
Aluminum - EPA 200.9
Aluminum - SM 3111D
Aluminum - SM 3113B
Aluminum - SM 3120B
Calcium - ASTM D511 (A)
Calcium - ASTM D511 (B)
Calcium - ASTM D6919
Calcium - EPA 200.5 Axial ICP
Calcium - EPA 200.7
Calcium - SM 3111B
Calcium - SM 3120B
Calcium - SM 3500-Ca B
Calcium - SM 3500-Ca D
Iron - EPA 200.5 Axial ICP
Iron - EPA 200.7
Iron - EPA 200.9
Iron - SM 3111B
Iron - SM 3113B
Iron - SM 3120B
Magnesium - ASTM D511 (A)
Magnesium - ASTM D511 (B)
Magnesium - ASTM D6919
Magnesium - EPA 200.5 Axial ICP
Magnesium - EPA 200.7
Magnesium - SM 3111 B
Magnesium - SM 3120 B
Magnesium - SM 3500-Mg B
Manganese - EPA 200.5 Axial ICP
Manganese - EPA 200.7
Manganese - EPA 200.8
Manganese - EPA 200.9
Manganese - SM 3111B
Manganese - SM 3113B

Manganese - SM 3120B
Silica - ASTM D859
Silica - EPA 200.5 Axial ICP
Silica - EPA 200.7
Silica - SM 3120B
Silica - SM 4500-Si D
Silica - SM 4500-Si E
Silica - SM 4500-Si F
Silica - SM 4500-SiO2 C
Silica - SM 4500-SiO2 D
Silica - SM 4500-SiO2 E
Silica - USGS I-1700-85
Silica - USGS I-2700-85
Silver - EPA 200.5 Axial ICP
Silver - EPA 200.7
Silver - EPA 200.8
Silver - EPA 200.9
Silver - SM 3111B
Silver - SM 3113B
Silver - SM 3120B
Silver - USGS I-3720-85
Sodium - ASTM D6919
Sodium - EPA 200.5 Axial ICP
Sodium - EPA 200.7
Sodium - SM 3111B
Zinc - EPA 200.5 Axial ICP
Zinc - EPA 200.7
Zinc - EPA 200.8
Zinc - SM 3111B
Zinc - SM 3120B
Class: Synthetic Organic Contaminants (SOC) – Dioxin 2,3,7,8-TCDD (Dioxin) - EPA 1613
Class: Synthetic Organic Contaminants (SOC) – Organochlorine Pesticides Aldrin - EPA 505
Aldrin - EPA 508
Aldrin - EPA 508.1
Aldrin - EPA 525.2
Chlordane - EPA 505

	Chlordane - EPA 508
	Chlordane - EPA 508.1
	Chlordane - EPA 525.2
	Chlordane - EPA 525.3
	Dieldrin - EPA 505
	Dieldrin - EPA 508
	Dieldrin - EPA 508.1
	Dieldrin - EPA 525.2
	Endrin - EPA 505
	Endrin - EPA 508
	Endrin - EPA 508.1
	Endrin - EPA 525.2
	Endrin - EPA 525.3
	Endrin - EPA 551.1
	Heptachlor - EPA 505
	Heptachlor - EPA 508
	Heptachlor - EPA 508.1
	Heptachlor - EPA 525.2
	Heptachlor - EPA 525.3
	Heptachlor - EPA 551.1
	Heptachlor epoxide - EPA 505
	Heptachlor epoxide - EPA 508
	Heptachlor epoxide - EPA 508.1
	Heptachlor epoxide - EPA 525.2
	Heptachlor epoxide - EPA 525.3
	Heptachlor epoxide - EPA 551.1
	Lindane (gamma-BHC) - EPA 505
	Lindane (gamma-BHC) - EPA 508
	Lindane (gamma-BHC) - EPA 508.1
	Lindane (gamma-BHC) - EPA 525.2
	Lindane (gamma-BHC) - EPA 525.3
	Lindane (gamma-BHC) - EPA 551.1
	Methoxychlor - EPA 505
	Methoxychlor - EPA 508
	Methoxychlor - EPA 508.1
	Methoxychlor - EPA 525.2
	Methoxychlor - EPA 525.3
	Methoxychlor - EPA 551.1

Toxaphene - EPA 505
Toxaphene - EPA 508
Toxaphene - EPA 508.1
Toxaphene - EPA 525.2
Toxaphene - EPA 525.3
Class: Synthetic Organic Contaminants (SOC) – Nitrogen-phosphorus Pesticides
Alachlor - EPA 505
Alachlor - EPA 507
Alachlor - EPA 508.1
Alachlor - EPA 525.2
Alachlor - EPA 525.3
Alachlor - EPA 551.1
Atrazine - EPA 505
Atrazine - EPA 507
Atrazine - EPA 508.1
Atrazine - EPA 523
Atrazine - EPA 525.2
Atrazine - EPA 525.3
Atrazine - EPA 536
Atrazine - EPA 551.1
Atrazine - Syngenta AG-625
Butachlor - EPA 507
Butachlor - EPA 508.1
Butachlor - EPA 525.2
Metolachlor - EPA 507
Metolachlor - EPA 508.1
Metolachlor - EPA 525.2
Metolachlor - EPA 551.1
Metribuzin - EPA 507
Metribuzin - EPA 508.1
Metribuzin - EPA 525.2
Metribuzin - EPA 551.1
Propachlor - EPA 507
Propachlor - EPA 508.1
Propachlor - EPA 525.2
Simazine - EPA 505
Simazine - EPA 507
Simazine - EPA 508.1

Simazine - EPA 523	
Simazine - EPA 525.2	
Simazine - EPA 525.3	
Simazine - EPA 536	
Simazine - EPA 551.1	
Class: Synthetic Organic Contaminants (SOC) – Herbicides	
2,4-D - ASTM D5317	
2,4-D - EPA 515.1	
2,4-D - EPA 515.2	
2,4-D - EPA 515.3	
2,4-D - EPA 515.4	
2,4-D - EPA 555	
2,4-D - SM 6640 B	
Dalapon - EPA 515.1	
Dalapon - EPA 515.3	
Dalapon - EPA 515.4	
Dalapon - EPA 552.1	
Dalapon - EPA 552.2	
Dalapon - EPA 552.3	
Dalapon - EPA 557	
Dalapon - SM 6640 B	
Dicamba - EPA 515.1	
Dicamba - EPA 515.2	
Dicamba - EPA 515.3	
Dicamba - EPA 515.4	
Dicamba - EPA 555	
Dinoseb - EPA 515.1	
Dinoseb - EPA 515.2	
Dinoseb - EPA 515.3	
Dinoseb - EPA 515.4	
Dinoseb - EPA 555	
Dinoseb - SM 6640 B	
Pentachlorophenol - ASTMD5317	
Pentachlorophenol - EPA 515.1	
Pentachlorophenol - EPA 515.2	
Pentachlorophenol - EPA 515.3	
Pentachlorophenol - EPA 515.4	
Pentachlorophenol - EPA 525.2	

Pentachlorophenol - EPA 525.3
Pentachlorophenol - EPA 555
Pentachlorophenol - SM 6640B
Picloram - ASTM D5317
Picloram - EPA 515.1
Picloram - EPA 515.2
Picloram - EPA 515.3
Picloram - EPA 515.4
Picloram - EPA 555
Picloram - SM 6640 B
Silvex (2.4.5-TP) - ASTMD5317
Silvex (2.4.5-TP) - EPA 515.1
Silvex (2.4.5-TP) - EPA 515.2
Silvex (2.4.5-TP) - EPA 515.3
Silvex (2.4.5-TP) - EPA 515.4
Silvex (2.4.5-TP) - EPA 555
Silvex (2.4.5-TP) - SM 6640 B
Class: Synthetic Organic Contaminants (SOC) – Miscellaneous
3-Hydroxycarbofuran - EPA 531.1
3-Hydroxycarbofuran - EPA 531.2
3-Hydroxycarbofuran - SM6610
Aldicarb - EPA 531.1
Aldicarb - EPA 531.2
Aldicarb - SM6610
Aldicarb sulfone - EPA 531.1
Aldicarb sulfone - EPA 531.2
Aldicarb sulfone - SM6610
Aldicarb sulfoxide - EPA 531.1
Aldicarb sulfoxide - EPA 531.2
Aldicarb sulfoxide - SM6610
Benzo[a]pyrene - EPA 525.2
Benzo[a]pyrene - EPA 525.3
Benzo[a]pyrene - EPA 550
Benzo[a]pyrene - EPA 550.1
Carbaryl - EPA 531.1
Carbaryl - EPA 531.2
Carbaryl - SM6610
Carbofuran - EPA 531.1

	Carbofuran - EPA 531.2
	Carbofuran - SM 6610B
	Carbofuran - SM 6651B
	Di(2-ethylhexyl)adipate - EPA 506
	Di(2-ethylhexyl)adipate - EPA 525.2
	Di(2-ethylhexyl)adipate - EPA 525.3
	Di(2-ethylhexyl)phthalate - EPA 506
	Di(2-ethylhexyl)phthalate - EPA 525.2
	Di(2-ethylhexyl)phthalate - EPA 525.3
	Dibromochloropropane (DBCP) - EPA 504.1
	Dibromochloropropane (DBCP) - EPA 524.3
	Dibromochloropropane (DBCP) - EPA 551.1
	Diquat - EPA 549.2
	Endothall - EPA 548.1
	Ethylene dibromide (EDB) - EPA 504.1
	Ethylene dibromide (EDB) - EPA 524.3
	Ethylene dibromide (EDB) - EPA 551.1
	Glyphosate - EPA 547
	Glyphosate - SM 6651B
	Hexachlorobenzene - EPA 505
	Hexachlorobenzene - EPA 508
	Hexachlorobenzene - EPA 508.1
	Hexachlorobenzene - EPA 525.2
	Hexachlorobenzene - EPA 525.3
	Hexachlorobenzene - EPA 551.1
	Hexachlorocyclopentadiene - EPA 505
	Hexachlorocyclopentadiene - EPA 508
	Hexachlorocyclopentadiene - EPA 508.1
	Hexachlorocyclopentadiene - EPA 525.2
	Hexachlorocyclopentadiene - EPA 525.3
	Hexachlorocyclopentadiene - EPA 551.1
	Methomyl - EPA 531.1
	Methomyl - EPA 531.2
	Methomyl - SM6610
	Oxamyl (Vydate) - EPA 531.1
	Oxamyl (Vydate) - EPA 531.2
	Oxamyl (Vydate) - SM 6610B
	Oxamyl (Vydate) - SM 6651B

PCBs (as Aroclors) Screening - EPA 505
PCBs (as Aroclors) Screening - EPA 508
PCBs (as Aroclors) Screening - EPA 508.1
PCBs (as Aroclors) Screening - EPA 525.2
PCBs (as Aroclors) Screening - EPA 525.3
PCBs (as Decachlorobiphenyl) - EPA 508A
Class: Volatile Organic Compounds (VOCs)
VOCS, REGULATED (group) - EPA 502.2
VOCS, REGULATED (group) - EPA 524.2
VOCS, REGULATED (group) - EPA 524.3
VOCS, UNREGULATED (group) - EPA 502.2
VOCS, UNREGULATED (group) - EPA 524.2
VOCS, UNREGULATED (group) - EPA 524.3
<u>Regulated VOCs</u>
® 1,1,1-Trichloroethane - EPA 502.2
® 1,1,1-Trichloroethane - EPA 524.2
® 1,1,1-Trichloroethane - EPA 524.3
® 1,1,1-Trichloroethane - EPA 551.1
® 1,1,2-Trichloroethane - EPA 502.2
® 1,1,2-Trichloroethane - EPA 524.2
® 1,1,2-Trichloroethane - EPA 524.3
® 1,1,2-Trichloroethane - EPA 551.1
® 1,1-Dichloroethylene - EPA 502.2
® 1,1-Dichloroethylene - EPA 524.2
® 1,1-Dichloroethylene - EPA 524.3
® 1,2,4-Trichlorobenzene - EPA 502.2
® 1,2,4-Trichlorobenzene - EPA 524.2
® 1,2,4-Trichlorobenzene - EPA 524.3
® 1,2-Dichlorobenzene - EPA 502.2
® 1,2-Dichlorobenzene - EPA 524.2
® 1,2-Dichlorobenzene - EPA 524.3
® 1,2-Dichloroethane - EPA 502.2
® 1,2-Dichloroethane - EPA 524.2
® 1,2-Dichloroethane - EPA 524.3
® 1,2-Dichloroethylene (cis-) - EPA 502.2
® 1,2-Dichloroethylene (cis-) - EPA 524.2
® 1,2-Dichloroethylene (cis-) - EPA 524.3
® 1,2-Dichloroethylene (trans-) - EPA 502.2

® 1,2-Dichloroethylene (trans-) - EPA 524.2
® 1,2-Dichloroethylene (trans-) - EPA 524.3
® 1,2-Dichloropropane - EPA 502.2
® 1,2-Dichloropropane - EPA 524.2
® 1,2-Dichloropropane - EPA 524.3
® 1,4-Dichlorobenzene - EPA 502.2
® 1,4-Dichlorobenzene - EPA 524.2
® 1,4-Dichlorobenzene - EPA 524.3
® Benzene - EPA 502.2
® Benzene - EPA 524.2
® Benzene - EPA 524.3
® Carbon tetrachloride - EPA 502.2
® Carbon tetrachloride - EPA 524.2
® Carbon tetrachloride - EPA 524.3
® Carbon tetrachloride - EPA 551.1
® Chlorobenzene - EPA 502.2
® Chlorobenzene - EPA 524.2
® Chlorobenzene - EPA 524.3
® Dichloromethane - EPA 502.2
® Dichloromethane - EPA 524.2
® Dichloromethane - EPA 524.3
® Ethylbenzene - EPA 502.2
® Ethylbenzene - EPA 524.2
® Ethylbenzene - EPA 524.3
® Styrene - EPA 502.2
® Styrene - EPA 524.2
® Styrene - EPA 524.3
® Tetrachloroethylene - EPA 502.2
® Tetrachloroethylene - EPA 524.2
® Tetrachloroethylene - EPA 524.3
® Tetrachloroethylene - EPA 551.1
® Toluene - EPA 502.2
® Toluene - EPA 524.2
® Toluene - EPA 524.3
® Trichloroethylene - EPA 502.2
® Trichloroethylene - EPA 524.2
® Trichloroethylene - EPA 524.3
® Trichloroethylene - EPA 551.1

® Vinyl chloride - EPA 502.2
® Vinyl chloride - EPA 524.2
® Vinyl chloride - EPA 524.3
® Xylenes (Total) - EPA 502.2
® Xylenes (Total) - EPA 524.2
® Xylenes (Total) - EPA 524.3
<u>Un-regulated VOCs</u>
1,1,1,2-Tetrachloroethane - EPA 502.2
1,1,1,2-Tetrachloroethane - EPA 524.2
1,1,1,2-Tetrachloroethane - EPA 524.3
1,1,2,2-Tetrachloroethane - EPA 502.2
1,1,2,2-Tetrachloroethane - EPA 524.2
1,1,2,2-Tetrachloroethane - EPA 524.3
1,1-Dichloroethane - EPA 502.2
1,1-Dichloroethane - EPA 524.2
1,1-Dichloroethane - EPA 524.3
1,1-Dichloropropene - EPA 502.2
1,1-Dichloropropene - EPA 524.2
1,1-Dichloropropene - EPA 524.3
1,2,3-Trichlorobenzene - EPA 502.2
1,2,3-Trichlorobenzene - EPA 524.2
1,2,3-Trichlorobenzene - EPA 524.3
1,2,3-Trichloropropane - EPA 502.2
1,2,3-Trichloropropane - EPA 524.2
1,2,3-Trichloropropane - EPA 524.3
1,2,4-Trimethylbenzene - EPA 502.2
1,2,4-Trimethylbenzene - EPA 524.2
1,2,4-Trimethylbenzene - EPA 524.3
1,3,5-Trimethylbenzene - EPA 502.2
1,3,5-Trimethylbenzene - EPA 524.2
1,3,5-Trimethylbenzene - EPA 524.3
1,3-Dichlorobenzene - EPA 502.2
1,3-Dichlorobenzene - EPA 524.2
1,3-Dichlorobenzene - EPA 524.3
1,3-Dichloropropane - EPA 502.2
1,3-Dichloropropane - EPA 524.2
1,3-Dichloropropane - EPA 524.3
1,3-Dichloropropylene (cis) - EPA 502.2

1,3-Dichloropropylene (cis) - EPA 524.2
1,3-Dichloropropylene (cis) - EPA 524.3
1,3-Dichloropropylene (trans) - EPA 502.2
1,3-Dichloropropylene (trans) - EPA 524.2
1,3-Dichloropropylene (trans) - EPA 524.3
2,2-Dichloropropane - EPA 502.2
2,2-Dichloropropane - EPA 524.2
2,2-Dichloropropane - EPA 524.3
2-Chlorotoluene - EPA 502.2
2-Chlorotoluene - EPA 524.2
2-Chlorotoluene - EPA 524.3
4-Chlorotoluene - EPA 502.2
4-Chlorotoluene - EPA 524.2
4-Chlorotoluene - EPA 524.3
4-Isopropyltoluene - EPA 502.2
4-Isopropyltoluene - EPA 524.2
4-Isopropyltoluene - EPA 524.3
Bromobenzene - EPA 502.2
Bromobenzene - EPA 524.2
Bromobenzene - EPA 524.3
Bromochloromethane - EPA 502.2
Bromochloromethane - EPA 524.2
Bromochloromethane - EPA 524.3
Bromomethane - EPA 502.2
Bromomethane - EPA 524.2
Bromomethane - EPA 524.3
Chloroethane - EPA 502.2
Chloroethane - EPA 524.2
Chloroethane - EPA 524.3
Chloromethane - EPA 502.2
Chloromethane - EPA 524.2
Chloromethane - EPA 524.3
Dibromomethane - EPA 502.2
Dibromomethane - EPA 524.2
Dibromomethane - EPA 524.3
Dichlorodifluoromethane - EPA 502.2
Dichlorodifluoromethane - EPA 524.2
Dichlorodifluoromethane - EPA 524.3

Fluorotrichloromethane - EPA 502.2
Fluorotrichloromethane - EPA 524.2
Fluorotrichloromethane - EPA 524.3
Hexachlorobutadiene - EPA 502.2
Hexachlorobutadiene - EPA 524.2
Hexachlorobutadiene - EPA 524.3
Isopropylbenzene - EPA 502.2
Isopropylbenzene - EPA 524.2
Isopropylbenzene - EPA 524.3
Methyl tert-butyl ether - EPA 502.2
Methyl tert-butyl ether - EPA 524.2
Methyl tert-butyl ether - EPA 524.3
Naphthalene - EPA 502.2
Naphthalene - EPA 524.2
Naphthalene - EPA 524.3
n-Butylbenzene - EPA 502.2
n-Butylbenzene - EPA 524.2
n-Butylbenzene - EPA 524.3
n-Propylbenzene - EPA 502.2
n-Propylbenzene - EPA 524.2
n-Propylbenzene - EPA 524.3
sec-Butylbenzene - EPA 502.2
sec-Butylbenzene - EPA 524.2
sec-Butylbenzene - EPA 524.3
tert-Butylbenzene - EPA 502.2
tert-Butylbenzene - EPA 524.2
tert-Butylbenzene - EPA 524.3

NR 149 Appendix I Table 2: Analytes and analyte groups available for accreditation

Analyte Groups

Analyte	Class code	Technologies		Class
		Aqueous matrix	Non-aqueous matrix	Drinking Water matrix
## DIOXINS & FURANS (group)	GRP	HRGC/MS	HRGC/MS	--
## HALOACETIC ACIDS (5)	GRP	--	--	EPA 552.1 EPA 552.2 EPA 552.3 EPA 557

				SM 6251B SM 6251 B-94 SM 6610B
## PAH (group)	GRP	GC GC/MS LC	GC GC/MS LC	--
## PCB as AROCLORS (group)	GRP	GC GC/MS	GC GC/MS	--
## PCB CONGENERS (group)	GRP	GC GC/MS HRGC/MS	GC GC/MS HRGC/MS	--
## PESTICIDES, ORGANOCHLORINE (group)	GRP	GC GC/MS	GC GC/MS	--
## SEMIVOLATILES [BNA] (group)	GRP	GC GC/MS	GC GC/MS	--
## THM (group) - EPA 502.2	GRP	--	--	EPA 502.2 EPA 524.2 EPA 524.3 EPA 551.1
## VOLATILE ORGANICS [VOC] (group)	GRP	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
Analytes				
1,1,1,2-Tetrachloroethane	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
1,1,1-Trichloroethane	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3 EPA 551.1
1,1,2,2-Tetrachloroethane	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
1,1,2-Trichloroethane	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3 EPA 551.1
1,1-Dichloroethane	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
1,1-Dichloroethylene	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
1,1-Dichloropropene	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
1,2,3,4-Diepoxybutane	VOC	GC/MS	GC/MS	--
1,2,3-Trichlorobenzene	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3

1,2,3-Trichloropropane	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
1,2,4,5-Tetrachlorobenzene	CHLH	GC GC/MS	GC GC/MS	--
1,2,4-Trichlorobenzene	CHLH VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
1,2,4-Trimethylbenzene	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
1,2-Dibromo-3-chloropropane (DBCP), (Dibromochloropropane)	PEST SOCM VOC	GC	GC	EPA 504.1 EPA 524.3 EPA 551.1
1,2-Dibromoethane (EDB), Ethylene dibromide	VOC	GC GC/MS	GC GC/MS	EPA 504.1 EPA 524.3 EPA 551.1
1,2-Dichlorobenzene	CHLH VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
1,2-Dichloroethane	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
1,2-Dichloroethene (cis)	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
1,2-Dichloroethene (trans)	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
1,2-Dichloropropane	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
1,2-Dinitrobenzene	NAROM	GC GC/MS	GC GC/MS	--
1,3,5-Trimethylbenzene	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
1,3,5-Trinitrobenzene	EXPLO NAROM	GC GC/MS LC	GC GC/MS LC	--
1,3-Dichloro-2-propanol	VOC	GC GC/MS	GC GC/MS	--
1,3-Dichlorobenzene	CHLH VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
1,3-Dichloropropane	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
1,3-Dichloropropylene (cis)	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
1,3-Dichloropropylene (trans)	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
1,3-Dinitrobenzene	EXPLO NAROM	GC GC/MS LC	GC GC/MS LC	--

1,3-Propanediol	VOC	GC GC/MS	GC GC/MS	--
1,4-Dichloro-2-butene (trans)	VOC	GC/MS	GC/MS	--
1,4-Dichlorobenzene	CHLC VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
1,4-Dinitrobenzene	NAROM	GC GC/MS	GC GC/MS	--
1,4-Dioxane	BNANH VOC	GC GC/MS	GC GC/MS	--
1,4-Naphthoquinone	NAROM	GC GC/MS	GC GC/MS	--
1,4-Phenylenediamine	NAROM	GC GC/MS	GC GC/MS	--
1-Acetyl-2-thiourea	BNANH	GC/MS	GC/MS	--
1-Chlorohexane	VOC	GC/MS	GC/MS	--
1-Chloronaphthalene	CHLH	GC/MS	GC/MS	--
1-Methylnaphthalene	PAH	GC GC/MS LC	GC GC/MS LC	--
1-Naphthylamine	NAROM	GC GC/MS	GC GC/MS	--
1-Propanol	VOC	GC/MS	GC/MS	--
2,2-Dichloropropane	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
2,3,4,6-Tetrachlorophenol	PHEN	GC GC/MS	GC GC/MS	--
2,3,5,6-Tetrachlorophenol	PHEN	GC GC/MS	GC GC/MS	--
2,3,7,8-TCDD (Dioxin)	SOCD	--	--	EPA 1613
2,3-Dichloropropene	VOC	GC GC/MS	GC GC/MS	--
2,3-Dinitrotoluene	EXPLO	GC/MS	--	--
2,4,5-T	APEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
2,4,5-T, butoxyethanol ester	APEST	LC LC/MS	LC LC/MS	--
2,4,5-T, butyl ester	APEST	LC LC/MS	LC LC/MS	--
2,4,5-Trichlorophenol	PHEN	GC GC/MS	GC GC/MS	--
2,4,5-Trimethylaniline	NAROM	GC GC/MS	GC GC/MS	--
2,4,6-Trichlorophenol	PHEN	GC GC/MS	GC GC/MS	--
2,4,6-Trinitrobenzene	EXPLO	LC	LC	--
2,4,6-Trinitrotoluene	EXPLO	LC	LC	--
2,4-D	APEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	ASTM D5317 EPA 515.1 EPA 515.2 EPA 515.3

				EPA 515.4 EPA 555 SM 6640 B
2,4-D, butoxyethanol ester	APEST	LC LC/MS	LC LC/MS	--
2,4-D, ethylhexyl ester	APEST	LC LC/MS	LC LC/MS	--
2,4-DB	APEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
2,4-DB salts and esters	APEST	GC LC LC/MS	GC LC LC/MS	--
2,4-Diamino-6-nitrotoluene	EXPLO	LC	LC	--
2,4-Diaminotoluene	NAROM	GC GC/MS	GC GC/MS	--
2,4-Dichlorophenol	PHEN	GC GC/MS	GC GC/MS	--
2,4-Dimethylphenol	PHEN	GC GC/MS	GC GC/MS	--
2,4-Dinitrophenol	PHEN	GC GC/MS	GC GC/MS	--
2,4-Dinitrotoluene	EXPLO NAROM	GC GC/MS LC	GC GC/MS LC	--
2,5-Dinitrotoluene	EXPLO	GC/MS	--	--
2,6-Dichlorophenol	PHEN	GC GC/MS	GC GC/MS	--
2,6-Dichlorosyringaldehyde	PHEN	GC GC/MS	GC GC/MS	--
2,6-Dinitrotoluene	EXPLO NAROM	GC GC/MS LC	GC GC/MS LC	--
2-Acetylamino fluorene	BNANH	GC/MS	GC/MS	--
2-Amino-4,6-dinitrotoluene	EXPLO	LC	LC	--
2-Aminoanthraquinone	BNANH	GC/MS	GC/MS	--
2-Chloroethanol	VOC	GC GC/MS	GC GC/MS	--
2-Chloronaphthalene	CHLH VOC	GC GC/MS	GC GC/MS	--
2-Chlorophenol	PHEN	GC GC/MS	GC GC/MS	--
2-Chlorosyringaldehyde	PHEN	GC GC/MS	GC GC/MS	--
2-Chlorotoluene	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
2-Cyclohexyl-4,6-dinitro-phenol	PHEN	GC GC/MS	GC GC/MS	--
2-Hexanone	VOC	GC GC/MS	GC GC/MS	--
2-Hydroxypropionitrile	BNANH	GC/MS	GC/MS	--
2-Methyl-3-nitroaniline	EXPLO	GC/MS	--	--

2-Methyl-4,6-dinitrophenol	PHEN	GC GC/MS	GC GC/MS	--
2-Methyl-5-nitroaniline	NAROM EXPLO	GC/MS	--	--
2-Methyl-6-nitroaniline	EXPLO	GC/MS	--	--
2-Methylnaphthalene	PAH	GC GC/MS LC	GC GC/MS LC	--
2-Methylphenol (o-Cresol)	PHEN	GC GC/MS	GC GC/MS	--
2-Naphthylamine	NAROM	GC/MS	GC/MS	--
2-Nitroaniline	NAROM	GC/MS	GC/MS	--
2-Nitrophenol	PHEN	GC GC/MS	GC GC/MS	--
2-Nitropropane	VOC	GC/MS	GC/MS	--
2-Nitrotoluene	EXPLO	GC/MS LC	LC	--
2-Pentanone	VOC	GC GC/MS	GC GC/MS	--
2-Picoline (2-Methylpyridine)	NAROM VOC	GC/MS	GC/MS	--
3-(Chloromethyl)pyridine hydrochloride	CHLH	GC/MS	GC/MS	--
3,3'-Dichlorobenzidine	BENZ	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
3,3'-Dimethoxybenzidine	BENZ	GC GC/MS LC/MS	GC GC/MS LC/MS	--
3,3'-Dimethylbenzidine	BENZ	GC GC/MS LC/MS	GC GC/MS LC/MS	--
3,4,5-Trichlorocatechol	PHEN	GC GC/MS	GC GC/MS	--
3,4,5-Trichloroguaiacol	PHEN	GC GC/MS	GC GC/MS	--
3,4,6-Trichlorocatechol	PHEN	GC GC/MS	GC GC/MS	--
3,4,6-Trichloroguaiacol	PHEN	GC GC/MS	GC GC/MS	--
3,4-Dichlorocatechol	PHEN	GC GC/MS	GC GC/MS	--
3,4-Dichloroguaiacol	PHEN	GC GC/MS	GC GC/MS	--
3,4-Dinitrotoluene	EXPLO	GC/MS	--	--
3,5-Dichlorobenzoic acid	APEST	GC LC LC/MS	GC LC LC/MS	--
3,5-Dinitrotoluene	EXPLO	GC/MS	--	--
3,6-Dichlorocatechol	PHEN	GC GC/MS	GC GC/MS	--
3-Amino-9-ethylcarbazole	NAROM	GC/MS	GC/MS	--
3-Chloropropionitrile	VOC	GC/MS	GC/MS	--

3-Hydroxycarbofuran	CARB	LC LC/MS	LC LC/MS	EPA 531.1 EPA 531.2 SM 6610
3-Methylcholanthrene	PAH	GC/MS	GC/MS	--
3-Methylphenol (m-Cresol)	PHEN	GC GC/MS	GC GC/MS	--
3-Nitroaniline	NAROM	GC/MS	GC/MS	--
3-Nitrotoluene	EXPLO	GC/MS LC	GC/MS LC	--
4,4'-DDD	CPEST	GC GC/MS	GC GC/MS	--
4,4'-DDE	CPEST	GC GC/MS	GC GC/MS	--
4,4'-DDT	CPEST	GC GC/MS	GC GC/MS	--
4,4'-Methylenebis (2-chloroaniline)	NAROM	GC/MS	GC/MS	--
4,4'-Methylenebis(N,N-dimethylaniline)	NAROM	GC/MS	GC/MS	--
4,4'-Oxydianiline	NAROM	GC/MS	GC/MS	--
4,5,6-Trichloroguaiacol	PHEN	GC GC/MS	GC GC/MS	--
4,5-Dichlorocatechol	PHEN	GC GC/MS	GC GC/MS	--
4,5-Dichloroguaiacol	PHEN	GC GC/MS	GC GC/MS	--
4,6-Dichlorocatechol	PHEN	GC GC/MS	GC GC/MS	--
4,6-Dichloroguaiacol	PHEN	GC GC/MS	GC GC/MS	--
4-Amino-2,6-dinitrotoluene	EXPLO	LC	LC	--
4-Aminobiphenyl	NAROM	GC/MS	GC/MS	--
4-Bromophenyl phenyl ether	HALO	GC GC/MS	GC GC/MS	--
4-Chloro-1,2-phenylenediamine	NAROM	GC/MS	GC/MS	--
4-Chloro-1,3-phenylenediamine	NAROM	GC/MS	GC/MS	--
4-Chloro-3-methylphenol (4-Chloro-m-cresol)	PHEN	GC GC/MS	GC GC/MS	--
4-Chloroaniline	BNANH NAROM	GC/MS	GC/MS	--
4-Chlorocatechol	PHEN	GC GC/MS	GC GC/MS	--
4-Chloroguaiacol	PHEN	GC GC/MS	GC GC/MS	--
4-Chlorophenol	PHEN	GC GC/MS	GC GC/MS	--
4-Chlorophenyl phenyl ether	HALO	GC GC/MS	GC GC/MS	--
4-Chlorotoluene	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
4-Dimethylaminoazobenzene	BNANH	GC/MS	GC/MS	--
4-Methyl-2-nitroaniline	EXPLO	GC/MS	--	--
4-Methyl-2-pentanone (Methyl isobutyl ketone)	VOC	GC GC/MS	GC GC/MS	--

4-Methyl-3-nitroaniline	EXPLO	GC/MS	--	--
4-Methylphenol (p-Cresol)	PHEN	GC GC/MS	GC GC/MS	--
4-Nitroaniline	NAROM	GC/MS	GC/MS	--
4-Nitrobiphenyl	NAROM	GC/MS	GC/MS	--
4-Nitrophenol	APEST PHEN	GC GC/MS LC	GC GC/MS LC	--
4-Nitroquinoline 1-oxide	BNANH	GC/MS	GC/MS	--
4-Nitrotoluene	EXPLO	GC/MS LC	LC	--
5,5-Diphenylhydantoin	BNANH	GC/MS	GC/MS	--
5,6-Dichlorovanillin	PHEN	GC GC/MS	GC GC/MS	--
5-Chloro-2-methylaniline	NAROM	GC/MS	GC/MS	--
5-Chlorovanillin	PHEN	GC GC/MS	GC GC/MS	--
5-Hydroxydicamba	APEST	GC	GC	--
5-Methyl-2-nitroaniline	EXPLO	GC/MS	--	--
5-Nitroacenaphthene	NAROM	GC/MS	GC/MS	--
5-Nitro-o-anisidine	NAROM	GC/MS	GC/MS	--
5-Nitro-o-toluidine	NAROM	GC/MS	--	--
6-Chlorovanillin	PHEN	GC GC/MS	GC GC/MS	--
7,12-Dimethylbenz(a)-anthracene	PAH	GC/MS	GC/MS	--
a,a-Dimethylphenethylamine	NAROM	GC/MS	GC/MS	--
Acenaphthene	PAH	GC GC/MS LC	GC GC/MS LC	--
Acenaphthylene	PAH	GC GC/MS LC	GC GC/MS LC	--
Acephate	OPEST	GC GC/MS	GC GC/MS	--
Acetaldehyde	ALDKE	LC	LC	--
Acetochlor	NPEST	GC GC/MS	GC GC/MS	--
Acetone	ALDKE VOC	GC GC/MS LC	GC GC/MS LC	--
Acetonitrile	VOC	GC GC/MS	GC GC/MS	--
Acetophenone	BNANH	GC/MS	GC/MS	--
Acidity as CaCO ₃	GC	Titration	--	--
Acifluorfen	APEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
Acrolein	BNANH VOC	GC GC/MS	GC GC/MS	--
Acrylamide	BNANH	GC/MS	GC/MS	--

Acrylonitrile	BNANH VOC	GC GC/MS	GC GC/MS	--
Alachlor	NPEST SOCN	GC GC/MS	GC GC/MS	EPA 505 EPA 507 EPA 508.1 EPA 525.2 EPA 525.3 EPA 551.1
Alachlor-ESA (Alachlor ethane sulfonic acid)	NPEST	LC/MS	LC/MS	--
Aldicarb	CARB	LC LC/MS	LC LC/MS	EPA 531.1 EPA 531.2 SM 6610
Aldicarb sulfone	CARB	LC LC/MS	LC LC/MS	EPA 531.1 EPA 531.2 SM 6610
Aldicarb sulfoxide	CARB	LC LC/MS	LC LC/MS	EPA 531.1 EPA 531.2 SM 6610
Aldrin	CPEST	GC GC/MS	GC GC/MS	EPA 505 EPA 508 EPA 508.1 EPA 525.2
Alkalinity	GC SCNM	Colorimetry Titration	--	ASTM D1067 (B) SM 2320B SM online 2320B-97 USGS I-1030-85
Allyl alcohol	VOC	GC GC/MS	GC GC/MS	--
Allyl chloride	VOC	GC GC/MS	GC GC/MS	--
alpha-BHC	CPEST	GC GC/MS	GC GC/MS	--
Aluminum	M	Colorimetry FLAA GFAA ICP ICP/MS	Colorimetry FLAA GFAA ICP ICP/MS	EPA 200.5 Axial EPA 200.7 EPA 200.8 EPA 200.9 SM 3111D SM 3111D-99 SM 3113B SM 3113B-99 SM 3120B SM 3120B-99
Ametryn	TPEST	GC GC/MS	GC GC/MS	--
Aminoazobenzene	BNANH	GC/MS	GC/MS	--
Aminocarb	CARB	LC/MS	LC/MS	--
Ammonia as N	GC	Colorimetry ISE Titration	Colorimetry ISE Titration	--
Anilazine	TPEST	GC GC/MS	GC GC/MS	--
Aniline	BNANH	GC/MS	GC/MS	--
Anthracene	PAH	GC GC/MS LC	GC GC/MS LC	--

Antimony	M	FLAA GFAA HydrideAA ICP ICP-MS	FLAA GFAA HydrideAA ICP ICP-MS	ASTM D3697 EPA 200.5 Axial EPA 200.8 EPA 200.9 SM 3113B SM 3113B-99
Aramite	BNANH	GC/MS	GC/MS	--
Arsenic	M	Colorimetry FLAA GFAA HydrideAA ICP ICP/MS	FLAA GFAA ICP ICP/MS	ASTM D2972 (B) ASTM D2972 (C) EPA 200.5 Axial EPA 200.8 EPA 200.9 SM 3113B SM 3113B-99 SM 3114B SM 3114B-97
Aspon	NPEST	GC GC/MS	LC/MS GC/MS	--
Asulam	CARB	LC/MS	LC/MS	--
Atraton	TPEST	GC GC/MS	GC GC/MS	--
Atrazine	TPEST	GC GC/MS	GC GC/MS	EPA 505 EPA 507 EPA 508.1 EPA 523 EPA 525.2 EPA 525.3 EPA 536 EPA 551.1 Syngenta AG-625
Azinphos ethyl	OPEST	GC GC/MS	GC GC/MS	--
Azinphos methyl (Guthion)	OPEST	GC GC/MS	GC GC/MS	--
Azobenzene	BNANH	GC GC/MS	GC GC/MS	--
Barban	CARB	GC GC/MS LC/MS	GC GC/MS LC/MS	--
Barium	M	FLAA GFAA ICP ICP/MS	FLAA GFAA ICP ICP/MS	EPA 200.5 Axial EPA 200.7 EPA 200.8 SM 3111D SM 3111D-99 SM 3113B SM 3113B-99 SM 3120B SM 3120B-99
Bay gon (Propoxur)	CARB	LC LC/MS	LC LC/MS	--
Bendiocarb	CARB	LC LC/MS	LC LC/MS	--
Benfluralin	NPEST	GC GC/MS	GC GC/MS	--
Benomyl	CARB	LC LC/MS	LC LC/MS	--

Bentazon	NPEST APEST	GC GC/MS LC	GC GC/MS LC	--
Benzene	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
Benzidine	BENZ	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
Benzo[a]anthracene	PAH	GC GC/MS LC	GC GC/MS LC	--
Benzo[a]pyrene	PAH SOCM	GC GC/MS LC	GC GC/MS LC	EPA 525.2 EPA 525.3 EPA 550 EPA 550.1
Benzo[b]fluoranthene	PAH	GC GC/MS LC	GC GC/MS LC	--
Benzo[g,h,i]perylene	PAH	GC GC/MS LC	GC GC/MS LC	--
Benzo[k]fluoranthene	PAH	GC GC/MS LC	GC GC/MS LC	--
Benzoic acid	PHEN	GC/MS	GC/MS	--
Benzoylprop ethyl	NPEST	LC/MS	LC/MS	--
Benzyl alcohol	BNANH	GC/MS	GC/MS	--
Benzyl chloride	CHLH	GC GC/MS	GC GC/MS	--
Beryllium	M	Colorimetry FLAA GFAA ICP ICP/MS	Colorimetry FLAA GFAA ICP ICP/MS	ASTM D3645 (B) EPA 200.5 Axial EPA 200.7 EPA 200.8 EPA 200.9 SM 3113B SM 3113B-99 SM 3120B SM 3120B-99
beta-BHC (β-BHC)	CPEST	GC GC/MS	GC GC/MS	--
Biochemical Oxygen Demand (BOD)	GC	5-day Assay	--	--
Biphenyl	BNANH	GC/MS	GC/MS	--
Bis(2-chloroethoxy)methane	HALO	GC GC/MS	GC GC/MS	--
Bis(2-chloroethyl)ether	HALO	GC GC/MS	GC GC/MS	--
Bis(2-chloroethyl)sulfide	VOC	GC/MS	GC/MS	--
Bis(2-chloroisopropyl)ether	HALO	GC GC/MS	GC GC/MS	--
Bis(2-ethylhexyl)phthalate, Di(2-ethylhexyl)phthalate	PHTHL SOCM	GC GC/MS	GC GC/MS	EPA 506 EPA 525.2 EPA 525.3

Bismuth	M	FLAA GFAA ICP ICP/MS	FLAA GFAA ICP ICP/MS	--
Bolstar	OPEST	GC GC/MS	GC GC/MS	--
Boron	M	Colorimetry ICP ICP/MS	Colorimetry ICP ICP/MS	--
Bromacil	NPEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
Bromate	DBP	--	--	ASTM D 6581 EPA 300.1 EPA 302.0 EPA 317.0 Rev. 2.0 EPA 321.8 EPA 326.0 EPA 557
Bromide	GC DBP	IC Titration	IC Titration	ASTM D 6581 EPA 300.0 EPA 300.1 EPA 326.0 EPA 327.0 Rev. 1.1
Bromoacetone	VOC	GC GC/MS	GC GC/MS	--
Bromobenzene	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
Bromochloromethane	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
Bromodichloromethane	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3 EPA 551.1
Bromoform	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3 EPA 551.1
Bromomethane (Methyl bromide)	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
Bromoxynil (Brominal)	APEST NPEST	GC/MS LC	GC/MS LC	--
Bromoxynil octanoate	NPEST	GC GC/MS	GC GC/MS	--
Busan 40	CARB	GC GC/MS	GC GC/MS	--
Busan 85	CARB	GC GC/MS	GC GC/MS	--
Butachlor	NPEST SOCN	GC GC/MS	GC GC/MS	EPA 507 EPA 508.1 EPA 525.2
Butanal	ALDKE	LC	LC	--

Butylbenzylphthalate	PHTHL	GC GC/MS	GC GC/MS	--
Butylate	NPEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
Cadmium	M	Colorimetry FLAA GFAA ICP ICP/MS	Colorimetry FLAA GFAA ICP ICP/MS	EPA 200.5 Axial EPA 200.7 EPA 200.8 EPA 200.9 SM 3113B SM 3113B-99
Calcium	M	Colorimetry FLAA FPAA IC ICP ICP/MS	Colorimetry FLAA FPAA ICP ICP/MS	ASTM D511 (A) ASTM D511 (B) ASTM D6919 EPA 200.5 Axial EPA 200.7 SM 3111B SM 3111B-99 SM 3120B SM 3120B-99 SM 3500-Ca B SM 3500-Ca D SM 3500-Ca B-97 SM 3500-Ca D-97
Captafol	CPEST	GC GC/MS	GC GC/MS	--
Captan	CPEST	GC GC/MS	GC GC/MS	--
Carbam-S	CARB	GC GC/MS	GC GC/MS	--
Carbaryl	CARB	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	EPA 531.1 EPA 531.2 SM 6610
Carbazole	BNANH	GC/MS	GC/MS	--
Carbendazim	CARB	LC/MS	LC/MS	--
Carbofuran	CARB	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	EPA 531.1 EPA 531.2 SM 6610 SM 6651B
Carbon disulfide	VOC	GC GC/MS	GC GC/MS	--
Carbon tetrachloride	VOC	GC GC/MS	GC GC/MS	EPA 502.2® EPA 524.2® EPA 524.3® EPA 551.1®
Carbonaceous Biological Oxygen Demand (cBOD)	GC	5-day Assay	--	--
Carbophenothion	OPEST	GC GC/MS	GC GC/MS	--
Carbosulfan	CARB	LC/MS	LC/MS	--
Ceriodaphnia dubia	AT CT	Acute Toxicity Assay Chronic Toxicity Assay	--	--

Chemical Oxygen Demand (COD)	GC	Colorimetry Titration	Titration	--
Chloramben	APEST	GC LC LC/MS	GC LC LC/MS	--
Chlorate	DBP	--	--	EPA 300.1
Chlordane (alpha)	CPEST	GC GC/MS	GC GC/MS	--
Chlordane (gamma)	CPEST	GC GC/MS	GC GC/MS	--
Chlordane (Technical)	CPEST	GC GC/MS	GC GC/MS	EPA 505 EPA 508 EPA 508.1 EPA 525.2 EPA 525.3
Chlorfenvinphos	OPEST	GC GC/MS	GC GC/MS	--
Chloride	GC SCNM	Colorimetry IC ISE Titration	Colorimetry IC ISE Titration	ASTM D4327 ASTM D512 (B) ASTM D6508, Rev. 2 EPA 300.0 EPA 300.1 SM 4110B SM 4500-Cl- B SM 4500-Cl- B-97 SM 4500-Cl- D SM 4500-Cl- D-97 SM 4110B-00
Chlorine dioxide	DBP	--	--	EPA 327.0, Rev.1 SM 4500-ClO2 C SM 4500-ClO2 C-00 SM 4500-ClO2 D SM 4500-ClO2 E SM 4500-ClO2 E-00
Chlorine, Free Residual	SCNM	--	--	SM 4500-Cl D-00 SM 4500-Cl F-00 SM 4500-Cl G-00 SM 4500-Cl H-00
Chlorine, Total Residual (TRC)	SCNM	--	--	SM 4500-Cl D-00 SM 4500-Cl E-00 SM 4500-Cl F-00 SM 4500-Cl G-00 SM 4500-Cl I-00
Chlorine, Combined	SCNM	--	--	ASTM D1253 SM 4500-Cl D SM 4500-Cl F SM 4500-Cl G
Chlorine, Free	SCNM	--	--	ASTM D1253 Chlorosense EPA 334.0 SM 4500-Cl D SM 4500-Cl F SM 4500-Cl G SM 4500-Cl H
Chlorine, Total Residual (TRC) Chlorine, Total	SCNM	Colorimetry ISE Titration	--	ASTM D1253 Chlorosense EPA 334.0

					SM 4500-CI D SM 4500-CI E SM 4500-CI F SM 4500-CI G SM 4500-CI I
Chlorite	SCNM	--	--	ASTM D 6581 EPA 300.0 EPA 300.1 EPA 317.0 Rev. 2.0 EPA 326.0 EPA 327.0 Rev. 1.1 SM 4500-CIO2 E SM 4500-CIO2E-00	
Chlorobenzene	VOC	GC GC/MS	GC GC/MS	EPA 502.2® EPA 524.2® EPA 524.3®	
Chlorobenzilate	CHLH	GC/MS	GC/MS	--	
Chloroethane	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3	
Chloroform	VOC	GC GC/MS	GC GC/MS	EPA 502.2® EPA 524.2® EPA 524.3® EPA 551.1®	
Chloromethane (Methyl chloride)	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3	
Chloromethyl methyl ether	VOC	GC GC/MS	GC GC/MS	--	
Chloroneb	CPEST	GC	GC	--	
Chlorophyll	GC	Colorimetry	--	--	
Chloroprene	VOC	GC GC/MS	GC GC/MS	--	
Chloropropham	CARB	LC/MS	LC/MS	--	
Chlorothalonil	NPEST	GC GC/MS	GC GC/MS	--	
Chloroxuron	CARB	LC/MS	LC/MS	--	
Chlorpyrifos	OPEST	GC GC/MS	GC GC/MS	--	
Chlorpyrifos methyl	OPEST	GC GC/MS	GC GC/MS	--	
Chlorthal (Dacthal di-acid, DCPA di-acid)	APEST	GC GC/MS LC	GC GC/MS LC	--	
Chromium, Hexavalent	M	Colorimetry FLAA IC	Colorimetry FLAA IC	--	
Chromium, Total	M	Colorimetry FLAA GFAA ICP ICP/MS	Colorimetry FLAA GFAA ICP ICP/MS	EPA 200.5 Axial EPA 200.7 EPA 200.8 EPA 200.9 SM 3113B SM 3113B-99 SM 3120B SM 3120B-99	

Chrysene	PAH	GC GC/MS LC	GC GC/MS LC	--
Clopyralid	APEST	GC GC/MS LC	GC GC/MS LC	--
Cobalt	M	FLAA GFAA ICP ICP/MS	FLAA GFAA ICP ICP/MS	--
Copper	M	Colorimetry FLAA GFAA ICP ICP/MS	Colorimetry FLAA GFAA ICP ICP/MS	ASTM D1688 (A) ASTM D1688 (C) EPA 200.5 Axial EPA 200.7 EPA 200.8 EPA 200.9 SM 3111B SM 3111B-99 SM 3113B SM 3113B-99 SM 3120B SM 3120B-99
Corrosivity	WC	--	pH Steel abrasion	--
Coumaphos	OPEST	GC GC/MS	GC GC/MS	--
Crotonaldehyde	ALDKE VOC	GC GC/MS LC	GC GC/MS LC	--
Crotoxyphos	OPEST	GC GC/MS	GC GC/MS	--
Cyanazine	TPEST	GC GC/MS	GC GC/MS	--
Cyanide (as free Cyanide)	PICNM	--	--	ALPKEM OIA-77 ASTM D2036 (A) ASTM D2036 (B) ASTM D6888 EPA 335.4 Kelada Kelada 01 ME355.01 QuikChem10-204-00-1-X SM 4500-CN- C,E SM 4500-CN- C,E-99 SM 4500-CN- C,F SM 4500-CN- C,F-99 USGS I-3300-85
Cyanide, Amenable	GC	--	--	SM 4500-CN- C,G SM 4500-CN- C,G-99
Cyanide, Available	GC	Colorimetry FIA-Diff.-Amp. Titration	Colorimetry Titration	--
Cyanide, Total	GC	Colorimetry FIA-Diff.-Amp. ISE Titration	Colorimetry ISE Titration	--
Cyclohexanone	ALDKE	LC	LC	--

Dalapon	APEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	EPA 515.1 EPA 515.3 EPA 515.4 EPA 552.1 EPA 552.2 EPA 552.3 EPA 557 SM 6640 B
Dazomet	CARB	GC GC/MS	GC GC/MS	--
Decanal	ALDKE	LC	LC	--
Deethylatrazine	TPEST	GC GC/MS	GC GC/MS	--
DEF (Butifos)	OPEST	GC GC/MS	GC GC/MS	--
Deisopropylatrazine	TPEST	GC GC/MS	GC GC/MS	--
delta-BHC	CPEST	GC GC/MS	GC GC/MS	--
Demeton-O	OPEST	GC GC/MS	GC GC/MS	--
Demeton-S	OPEST	GC GC/MS	GC GC/MS	--
Di(2-ethylhexyl)adipate	SOCM	--	--	EPA 506 EPA 525.2 EPA 525.3
Diallate (cis or trans)	CARB	GC GC/MS	GC GC/MS	--
Diaminoatrazine	TPEST	GC GC/MS	GC GC/MS	--
Diazinon	OPEST	GC GC/MS	GC GC/MS	--
Dibenz(a,j)acridine	PAH	GC/MS	GC/MS	--
Dibenzo[a,e]pyrene	PAH	GC/MS	GC/MS	--
Dibenzo[a,h]anthracene	PAH	GC GC/MS LC	GC GC/MS LC	--
Dibenzofuran	BNANH	GC/MS	GC/MS	--
Dibromochloromethane	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3 EPA 551.1
Dibromomethane (Methylene bromide)	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
Dicamba	APEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	EPA 515.1 EPA 515.2 EPA 515.3 EPA 515.4 EPA 555
Dichlofenthion	OPEST	GC GC/MS	GC GC/MS	--
Dichlone	CPEST	GC GC/MS	GC GC/MS	--

Dichlorodifluoromethane	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
Dichlorprop (2,4-DP)	APEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
Dichlorprop salts and esters	APEST	GC LC LC/MS	GC LC LC/MS	--
Dichlorvos (DDVP)	OPEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
Diclofop	APEST	GC GC/MS LC	GC GC/MS LC	--
Dicrotophos	OPEST	GC GC/MS	GC GC/MS	--
Dieldrin	CPEST	GC GC/MS	GC GC/MS	EPA 505 EPA 508 EPA 508.1 EPA 525.2
Diethyl ether (Ethyl ether)	VOC	GC GC/MS	GC GC/MS	--
Diethyl phthalate	PHTHL	GC GC/MS	GC GC/MS	--
Diethyl sulfate	BNANH	GC/MS	GC/MS	--
Diethylstilbestrol	BNANH	GC/MS	GC/MS	--
Dihydrosaffrole	BNANH	GC/MS	GC/MS	--
Diisopropyl ether	VOC	GC/MS	GC/MS	--
Dimethenamid	NPEST	GC GC/MS	GC GC/MS	--
Dimethoate	OPEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
Dimethyl phthalate	PHTHL	GC GC/MS	GC GC/MS	--
Di-n-butyl phthalate	PHTHL	GC GC/MS	GC GC/MS	--
Di-n-octyl phthalate	PHTHL	GC GC/MS	GC GC/MS	--
Dinoseb (2-sec-butyl-4,6-Dinitrophenol)	APEST PHEN	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	EPA 515.1 EPA 515.2 EPA 515.3 EPA 515.4 EPA 555 SM 6640 B
Dioxacarb	CARB	LC	LC	--
Dioxathion	OPEST	GC GC/MS	GC GC/MS	--
Diphenylamine	BNANH	GC/MS	GC/MS	--
Diquat	PEST SOCM	LC	LC	EPA 549.2

Disulfoton	OPEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
Diuron	CARB	LC LC/MS	LC LC/MS	--
Endosulfan I	CPEST	GC GC/MS	GC GC/MS	--
Endosulfan II	CPEST	GC GC/MS	GC GC/MS	--
Endosulfan sulfate	CPEST	GC GC/MS	GC GC/MS	--
Endothall	PEST SOCM	LC	LC	EPA 548.1
Endrin	CPEST	GC GC/MS	GC GC/MS	EPA 505 EPA 508 EPA 508.1 EPA 525.2 EPA 525.3 EPA 551.1
Endrin aldehyde	CPEST	GC GC/MS	GC GC/MS	--
Endrin ketone	CPEST	GC GC/MS	GC GC/MS	--
Epichlorohydrin	VOC	GC GC/MS	GC GC/MS	--
EPN	OPEST	GC GC/MS	GC GC/MS	--
EPTC (Eptam)	CARB	GC GC/MS LC/MS	GC GC/MS LC/MS	--
EPTOX Extraction	WE	--	Leach Test	--
Ethalfuralin	NPEST	GC GC/MS	GC GC/MS	--
Ethanol	VOC	GC GC/MS	GC GC/MS	--
Ethion	OPEST	GC GC/MS	GC GC/MS	--
Ethoprop	OPEST	GC GC/MS	GC GC/MS	--
Ethyl acetate	VOC	GC GC/MS	GC GC/MS	--
Ethyl carbamate	CARB	GC GC/MS	GC GC/MS	--
Ethyl methacrylate	VOC	GC GC/MS	GC GC/MS	--
Ethyl methanesulfonate	BNANH	GC/MS	GC/MS	--
Ethylbenzene	VOC	GC GC/MS	GC GC/MS	EPA 502.2® EPA 524.2® EPA 524.3®
Ethylene dibromide (EDB)	PEST SOCM	--	--	EPA 504.1 EPA 524.3 EPA 551.1
Ethylene glycol	VOC	GC GC/MS	GC GC/MS	--

Ethylene oxide	VOC	GC GC/MS	GC GC/MS	--
Famphur	OPEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
Fenarimol	NPEST	GC GC/MS	GC GC/MS	--
Fenitrothion	OPEST	GC GC/MS	GC GC/MS	--
Fensulfothion	OPEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
Fenthion	OPEST	GC GC/MS	GC GC/MS	--
Fenuron	CARB	LC LC/MS	LC LC/MS	--
Fenuron-TCA	CARB	LC/MS	LC/MS	--
Fenvalerate	PEST	LC	LC	--
Fluchloralin	BNANH	GC/MS	GC/MS	--
Fluometuron	CARB	LC LC/MS	LC LC/MS	--
Fluoranthene	PAH	GC GC/MS LC	GC GC/MS LC	--
Fluorene	PAH	GC GC/MS LC	GC GC/MS LC	--
Fluoride	GC	Colorimetry IC ISE	Colorimetry IC ISE	ASTM D1179 (B) ASTM D4327 ASTM D6508, Rev 2 EPA 300.0 EPA 300.1 HACH Method 10225 SM 4110B SM 4500F- B, D SM 4500F- B, D-97 SM 4500F- C SM 4500F- C-97 SM 4500F- E SM 4500F- E-97 SM 4110B-00 Technicon 129-71W Technicon 380-75WE
Fonofos	OPEST	GC GC/MS	GC GC/MS	--
Formaldehyde	ALDKE	LC	LC	--
Glyphosate	PEST SOCM	LC	LC	EPA 547 SM 6651B
Gold	M	FLAA GFAA ICP ICP/MS	FLAA GFAA ICP ICP/MS	--
Hardness, Total as CaCO ₃	GC	Colorimetry Titration	--	--

			FLAA ICP		
Heptachlor	CPEST		GC GC/MS	GC GC/MS	EPA 505 EPA 508 EPA 508.1 EPA 525.2 EPA 525.3 EPA 551.1
Heptachlor epoxide	CPEST		GC GC/MS	GC GC/MS	EPA 505 EPA 508 EPA 508.1 EPA 525.2 EPA 525.3 EPA 551.1
Heptanal	ALDKE		LC	LC	--
Hexachlorobenzene	CHLH		GC GC/MS	GC GC/MS	EPA 505 EPA 508 EPA 508.1 EPA 525.2 EPA 525.3 EPA 551.1
Hexachlorobutadiene	CHLH VOC		GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
Hexachlorocyclopentadiene	CHLH		GC GC/MS	GC GC/MS	EPA 505 EPA 508 EPA 508.1 EPA 525.2 EPA 525.3 EPA 551.1
Hexachloroethane	CHLH VOC		GC GC/MS	GC GC/MS	--
Hexachlorophene	CHLH		GC/MS	GC/MS	--
Hexachloropropene	CHLH		GC/MS	GC/MS	--
Hexamethylphosphoramide	OPEST		GC GC/MS	GC GC/MS	--
Hexanal	ALDKE		LC	LC	--
Hexane, n-	VOC		GC/MS	GC/MS	--
Hexazinone	NPEST		GC GC/MS	GC GC/MS	--
HMX	EXPLO		LC	LC	--
Hydroquinone	BNANH		GC/MS	GC/MS	--
Ignitability	WC		--	Pensky-Martens Closed Cup Setaflash Closed Cup Small Scale Closed Cup	--
Indeno(1,2,3-cd)pyrene	PAH		GC GC/MS LC	GC GC/MS LC	--
Iodomethane (Methyl iodide)	VOC		GC GC/MS	GC GC/MS	--
Iridium	M		FLAA GFAA	FLAA GFAA	--

		ICP ICP/MS	ICP ICP/MS	
Iron	M	Colorimetry FLAA GFAA ICP ICP/MS	Colorimetry FLAA GFAA ICP ICP/MS	EPA 200.5 Axial EPA 200.7 EPA 200.9 SM 3111B SM 3111B-99 SM 3113B SM 3113B-99 SM 3120B SM 3120B-99
Isobutyl alcohol (2-Methyl-1-propanol)	VOC	GC GC/MS	GC GC/MS	--
Isodrin	CPEST	GC GC/MS	GC GC/MS	--
Isophorone	NAROM	GC GC/MS	GC GC/MS	--
Isopropalin	NPEST	GC GC/MS	GC GC/MS	--
Isopropyl alcohol (2-Propanol)	VOC	GC GC/MS	GC GC/MS	--
Isopropylbenzene	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
Isosafrole	BNANH	GC/MS	GC/MS	--
Isovaleraldehyde	ALDKE	LC	LC	--
Kepone	CPEST	GC GC/MS	GC GC/MS	--
Kjeldahl Nitrogen, Total (TKN)	GC	Colorimetry ISE Titration	Colorimetry ISE Titration	--
KN Methyl	CARB	GC GC/MS	GC GC/MS	--
Lead	M	Colorimetry FLAA GFAA ICP ICP/MS	Colorimetry FLAA GFAA ICP ICP/MS	ASTM D3559 (D) EPA 200.5 Axial EPA 200.8 EPA 200.9 Palintest 1011 SM 3113B SM 3113B-99
Leptophos	OPEST	GC GC/MS	GC GC/MS	--
Lindane (gamma-BHC)	CPEST	GC GC/MS	GC GC/MS	EPA 505 EPA 508 EPA 508.1 EPA 525.2 EPA 525.3 EPA 551.1
Linuron	CARB	LC LC/MS	LC LC/MS	--
Lithium	M	FLAA GFAA ICP ICP/MS	FLAA GFAA ICP ICP/MS	--
Magnesium	M	FLAA	FLAA	ASTM D511 (A) ASTM D511 (B)

		FPA ICP ICP/MS	FPA ICP ICP/MS	ASTM D6919 EPA 200.5 Axial EPA 200.7 SM 3111B SM 3111B-99 SM 3120B SM 3120B-99 SM 3500-Mg B
Malathion	OPEST	GC GC/MS	GC GC/MS	--
Maleic anhydride	BNANH	GC/MS	GC/MS	--
Malononitrile	VOC	GC GC/MS	GC GC/MS	--
Manganese	M	Colorimetry FLAA GFAA ICP ICP/MS	Colorimetry FLAA GFAA ICP ICP/MS	EPA 200.5 Axial EPA 200.7 EPA 200.8 EPA 200.9 SM 3111B SM 3111B-99 SM 3113B SM 3113B-99 SM 3120B SM 3120B-99
MCPA	APEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
MCPB	APEST	GC GC/MS LC	GC GC/MS LC	--
MCP (Mecoprop)	APEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
m-Cumenyl methylcarbamate	CARB	LC LC/MS	LC LC/MS	--
Mercury	M	CVAA CVAFS LC ICP/MS ThermDecAA	CVAA CVAFS LC ICP/MS ThermDecAA	ASTM D3223 EPA 200.8 EPA 245.1 EPA 245.2 SM 3112B SM 3112B-99
Mercury, Organo-	M	LC	LC	--
Mercury, Trace Level	M	CVAFS LC ICP/MS ThermDecAA	CVAFS LC ICP/MS ThermDecAA	--
Merphos	OPEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
Mestranol	BNANH	GC/MS	GC/MS	--
Methacrylonitrile	VOC	GC GC/MS	GC GC/MS	--
Methamidophos	OPEST	GC GC/MS	GC GC/MS	--

Methanol	VOC	GC GC/MS	GC GC/MS	--
Methapyrilene	BNANH	GC/MS	GC/MS	--
Methiocarb	CARB	LC LC/MS	LC LC/MS	--
Methomyl	CARB	LC LC/MS	LC LC/MS	EPA 531.1 EPA 531.2 SM 6610
Methoxychlor	CPEST	GC GC/MS	GC GC/MS	EPA 505 EPA 508 EPA 508.1 EPA 525.2 EPA 525.3 EPA 551.1
Methylacrylate	VOC	GC GC/MS	GC GC/MS	--
Methyl ethyl ketone (MEK, 2-Butanone)	VOC	GC GC/MS	GC GC/MS	--
Methyl methacrylate	VOC	GC GC/MS	GC GC/MS	--
Methyl methanesulfonate	BNANH	GC/MS	GC/MS	--
Methyl tert-butyl ether (MtBE)	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
Methylene chloride	VOC	GC GC/MS	GC GC/MS	EPA 502.2® EPA 524.2® EPA 524.3®
Metolachlor	NPEST SOCN	GC GC/MS	GC GC/MS	EPA 507 EPA 508.1 EPA 525.2 EPA 551.1
Metolcarb	CARB	LC LC/MS	LC LC/MS	--
Metribuzin	NPEST SOCN	GC GC/MS	GC GC/MS	EPA 507 EPA 508.1 EPA 525.2 EPA 551.1
Mevinphos	OPEST	GC GC/MS	GC GC/MS	--
Mexacarbate	CARB	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
Mirex	CPEST	GC GC/MS	GC GC/MS	--
Molinate	CARB	LC LC/MS	LC LC/MS	--
Molybdenum	M	FLAA GFAA ICP ICP/MS	FLAA GFAA ICP ICP/MS	--
Monocrotophos	OPEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--

Monuron	CARB	LC LC/MS	LC LC/MS	--
Monuron-TCA	CARB	LC/MS	LC/MS	--
m-Tolualdehyde	ALDKE	LC	LC	--
m-Xylene	VOC	GC GC/MS	GC GC/MS	--
Nabam	CARB	GC GC/MS	GC GC/MS	--
Nabonate	CARB	GC GC/MS	GC GC/MS	--
Naled	OPEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
Naphthalene	PAH VOC	GC GC/MS LC	GC GC/MS LC	EPA 502.2 EPA 524.2 EPA 524.3
Napropamide	NPEST	GC GC/MS	GC GC/MS	--
n-Butylalcohol (1-Butanol)	VOC	GC GC/MS	GC GC/MS	--
n-Butylbenzene	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
Neburon	CARB	LC/MS	LC/MS	--
Nickel	M	Colorimetry FLAA GFAA ICP ICP/MS	Colorimetry FLAA GFAA ICP ICP/MS	EPA 200.5 Axial EPA 200.7 EPA 200.8 EPA 200.9 SM 3111B SM 3111B-99 SM 3113B SM 3113B-99 SM 3120B SM 3120B-99
Nicotine	BNANH	GC/MS	GC/MS	--
Nitrate	GC PICNM	Colorimetry IC ISE	Colorimetry IC ISE	ASTM D3867 (A) ASTM D3867 (B) ASTM D4327 ASTM D6508, Rev 2 EPA 300.0 EPA 300.1 EPA 353.2 Hach Method 10206 Orion 601 SM 4110B SM 4110B-00 SM 4500-NO3- D SM 4500-NO3- D-00 SM 4500-NO3- E SM 4500-NO3- E-00 SM 4500-NO3- F SM 4500-NO3-F-00 Systea Easy Waters B-1011

Nitrate + Nitrite				ASTM D3867 (A) ASTM D3867 (B) ASTM D4327 ASTM D6508, Rev 2 EPA 300.0 EPA 300.1 EPA 353.2 SM 4110B SM 4110B-00 SM 4500-NO3- D SM 4500-NO3- D-00 SM 4500-NO3- E SM 4500-NO3- E-00 SM 4500-NO3- F SM 4500-NO3-F-00 Waters B-1011
Nitrite				ASTM D3867 (A) ASTM D3867 (B) ASTM D4327 ASTM D6508, Rev 2 EPA 300.0 EPA 300.1 EPA 353.2 SM 4110B SM 4110B-00 SM 4500-NO2- B SM 4500-NO2- B-00 SM 4500-NO3- E SM 4500-NO3- E-00 SM 4500-NO3- F SM 4500-NO3-F-00 Systea Easy Waters B-1011
Nitrobenzene	EXPLO NAROM	GC GC/MS LC	GC GC/MS LC	--
Nitrofen	BNANH	GC/MS	GC/MS	--
Nitroglycerin	EXPLO	LC	LC	--
N-Nitrosodiethylamine	NSAMI	GC GC/MS	GC GC/MS	--
N-Nitrosodimethylamine	NSAMI	GC GC/MS	GC GC/MS	--
N-Nitrosodi-n-butylamine	NSAMI	GC GC/MS	GC GC/MS	--
N-Nitrosodi-n-propylamine	NSAMI	GC GC/MS	GC GC/MS	--
N-Nitrosodiphenylamine	NSAMI	GC GC/MS	GC GC/MS	--
N-Nitrosomethylethylamine	NSAMI	GC GC/MS	GC GC/MS	--
N-Nitrosomorpholine	NSAMI	GC GC/MS	GC GC/MS	--
N-Nitrosopiperidine	NSAMI	GC GC/MS	GC GC/MS	--
N-Nitrosopyrrolidine	NSAMI	GC GC/MS	GC GC/MS	--

Nonanal	ALDKE	LC	LC	--
Norflurazon	NPEST	GC GC/MS	GC GC/MS	--
n-Propylamine	VOC	GC/MS	GC/MS	--
n-Propylbenzene	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
O,O,O-Triethyl phosphorothioate	BNANH	GC/MS	GC/MS	--
o-Anisidine	BNANH	GC/MS	GC/MS	--
o-Chlorophenyl thiourea	CARB	LC/MS	LC/MS	--
Octamethyl pyrophosphoramidate	BNANH	GC/MS	GC/MS	--
Octanal	ALDKE	LC	LC	--
Oil & Grease, as Hexane Extractable Material (HEM)	GC	Extraction/ Gravimetry	--	--
Organic Carbon, Dissolved (DOC)	SCNM	--	--	EPA 415.3 SM 5310B SM 5310C SM 5310D
Organic Carbon, Total (TOC)	GC SCNM	NonDispersive IR Microcoulometry	NonDispersive IR Microcoulometry	EPA 415.3 SM 5310B SM 5310C SM 5310D
Organic Halides, (Total-TOX and Adsorbable-AOX)	GC	NonDispersive IR Microcoulometry	NonDispersive IR Microcoulometry	--
Orthophosphate	GC SCNM	Colorimetry IC	Colorimetry IC	ASTM D4327 ASTM D515 (A) ASTM D6508, Rev. 2 EPA 300.0 EPA 300.1 EPA 365.1 SM 4110B SM 4110B-00 SM 4500-P E SM 4500-P F USGS I-1601-85 USGS I-2598-85 USGS I-2601-90
Osmium	M	FLAA GFAA ICP ICP/MS	FLAA GFAA ICP ICP/MS	--
o-Tolualdehyde	ALDKE	LC	LC	--
o-Toluidine	BNANH VOC	GC/MS	GC/MS	--
Oxamyl (Vydate)	CARB	LC LC/MS	LC LC/MS	EPA 531.1 EPA 531.2 SM 6610 SM 6651B
Oxygen, Dissolved	GC	ISE	--	--
o-Xylene	VOC	GC GC/MS	GC GC/MS	--
Ozone	DBP	--	--	SM 4500-O3 B-97 SM 4500-O3 B

Palladium	M	FLAA GFAA ICP ICP/MS	FLAA GFAA ICP ICP/MS	--
Paraldehyde	VOC	GC GC/MS	GC GC/MS	--
Paraquat	PEST	LC	LC	--
Parathion (Parathion ethyl)	OPEST	GC GC/MS	GC GC/MS	--
Parathion methyl	OPEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
p-Benzoquinone	BNANH	GC/MS	GC/MS	--
PCBs (as Aroclors) Screening	SOCM	--	--	EPA 505 EPA 508 EPA 508.1 EPA 525.2 EPA 525.3
PCBs (as Decachlorobiphenyl)	SOCM			EPA 508A
p-Cresidine	BNANH	GC/MS	GC/MS	--
Pebulate	CARB	LC/MS	LC/MS	--
Pendimethalin	NPEST	GC GC/MS	GC GC/MS	--
Pentachlorobenzene	CHLH	GC GC/MS	GC GC/MS	--
Pentachloroethane	CHLH VOC	GC/MS	GC/MS	--
Pentachloronitrobenzene (PCNB)	CPEST NAROM	GC GC/MS	GC GC/MS	--
Pentachlorophenol	APEST PHEN	GC GC/MS LC	GC GC/MS LC	ASTM D5317 EPA 515.1 EPA 515.2 EPA 515.3 EPA 515.4 EPA 525.2 EPA 525.3 EPA 555 SM 6640 B
Pentanal (Valeraldehyde)	ALDKE	LC	LC	--
Moisture Content	GC	--	Karl Fischer	--
Percent Solids	GC	--	Gravimetry	--
Permethrin	PEST	GC	GC	--
Perthane	CPEST	GC	GC	--
PETN (Pentaerythritol tetranitrate)	EXPLO	LC	LC	--
pH	GC SCNM	ISE	ISE	ASTM D1293 EPA 150.1 EPA 150.2 SM 4500-H+ B SM 4500-H+ B-00
Phenacetin	BNANH	GC/MS	GC/MS	--

Phenanthrene	PAH	GC GC/MS LC	GC GC/MS LC	--
Phenobarbital	BNANH	GC/MS	GC/MS	--
Phenol	PHEN	GC GC/MS	GC GC/MS	--
Phenolics, Total	GC	Colorimetry	Colorimetry	--
Phorate	OPEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
Phosalone	OPEST	GC GC/MS	GC GC/MS	--
Phosmet (Imidan)	OPEST	GC GC/MS	GC GC/MS	--
Phosphamidon	OPEST	GC GC/MS	GC GC/MS	--
Phosphorus, Total	GC	Colorimetry	Colorimetry ICP	--
Phthalic anhydride	BNANH	GC/MS	GC/MS	--
Picloram	APEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	ASTM D5317 EPA 515.1 EPA 515.2 EPA 515.3 EPA 515.4 EPA 555 SM 6640 B
Picric acid (Trinitrophenol)	EXPLO	LC	LC	--
Pimephales promelas	AT CT	Acute Toxicity Assay Chronic Toxicity Assay	--	--
Piperonyl sulfoxide	BNANH	GC/MS	GC/MS	--
p-Isopropyltoluene (4-Isopropyltoluene)	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
Platinum	M	FLAA GFAA ICP ICP/MS	FLAA GFAA ICP ICP/MS	--
Potassium	M	FPAA FLAA ICP ICP/MS	FPAA FLAA ICP ICP/MS	--
Promecarb	CARB	LC LC/MS	LC LC/MS	--
Prometon	TPEST	GC GC/MS	GC GC/MS	--
Prometryn	TPEST	GC GC/MS	GC GC/MS	--
Pronamide	NPEST	GC GC/MS	GC GC/MS	--
Propachlor	NPEST SOCN	GC GC/MS LC/MS	GC GC/MS LC/MS	EPA 507 EPA 508.1 EPA 525.2

Propanal (Propionaldehyde)	ALDKE	LC	LC	--
Propanil	CARB	LC	LC	--
Propanil	NPEST	GC GC/MS	GC GC/MS	--
Propargyl alcohol	VOC	GC GC/MS	GC GC/MS	--
Propazine	TPEST	GC GC/MS	GC GC/MS	--
Propham	CARB	LC LC/MS	LC LC/MS	--
Propionitrile (Ethyl cyanide)	VOC	GC GC/MS	GC GC/MS	--
Propylene glycol	VOC	GC/MS	GC/MS	--
Propylthiouracil	BNANH	GC/MS	GC/MS	--
Prosulfocarb	CARB	LC/MS	LC/MS	--
p-Tolualdehyde	ALDKE	LC	LC	--
p-Xylene	VOC	GC GC/MS	GC GC/MS	--
Pyrene	PAH	GC GC/MS LC	GC GC/MS LC	--
Pyrethrin I	PEST	LC	LC	--
Pyrethrin II	PEST	LC	LC	--
Pyridine	BNANH VOC	GC/MS	GC/MS	--
Qualitative FID Fingerprint	SSCAN	GC	GC	--
RDX	EXPLO	LC	LC	--
Reagent Water Shake Extraction (ASTM Leach Test)	WE	--	Leach Test	--
Residue, Filterable (TDS)	GC SCNM	Gravimetry	--	SM 2540C SM 2540C-97
Residue, Nonfilterable (TSS)	GC	Gravimetry	--	--
Residue, Settleable	GC	Gravimetry	--	--
Residue, Total	GC	Gravimetry	Gravimetry	--
Residue, Volatile (TVS)	GC	Gravimetry	Gravimetry	--
Residue, Volatile, Nonfilterable (TVSS)	GC	Gravimetry	--	--
Resorcinol	BNANH	GC/MS	GC/MS	--
Rhodium	M	FLAA GFAA ICP ICP/MS	FLAA GFAA ICP ICP/MS	--
Ronnel	OPEST	GC GC/MS	GC GC/MS	--
Rotenone	PEST	LC/MS	LC/MS	--
Ruthenium	M	FLAA GFAA ICP ICP/MS	FLAA GFAA ICP ICP/MS	--
Safrole	BNANH	GC/MS	GC/MS	--
Sebumeton	NPEST	LC	LC	--

sec-Butylbenzene	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
Selenastrum capricornutum	CT	Chronic Toxicity	--	--
Selenium	M	GFAA HydrideAA ICP ICP/MS	GFAA HydrideAA ICP ICP/MS	ASTM D3859 (A) ASTM D3859 (B) EPA 200.5 Axial EPA 200.8 EPA 200.9 SM 3113B SM 3113B-99 SM 3114B SM 3114B-97
Siduron	CARB	LC LC/MS	LC LC/MS	--
Silica	GC	Colorimetry ICP	--	ASTM D859 EPA 200.5 Axial ICP EPA 200.7 SM 3120B SM 3120B-99 SM 4500-Si D SM 4500-Si E SM 4500-Si F SM 4500-SiO2 C SM 4500-SiO2-C-97 SM 4500-SiO2 D SM 4500-SiO2-D-97 SM 4500-SiO2 E SM 4500-SiO2-E-97 USGS I-1700-85 USGS I-2700-85
Silicon	M	Colorimetry ICP ICP/MS	ICP ICP/MS	--
Silver	M	FLAA GFAA ICP ICP/MS	FLAA GFAA ICP ICP/MS	EPA 200.5 Axial EPA 200.7 EPA 200.8 EPA 200.9 SM 3111B SM 3111B-99 SM 3113B SM 3113B-99 SM 3120B SM 3120B-99 USGS I-3720-85
Silvex (2,4,5-TP)	APEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	ASTM D5317 EPA 515.1 EPA 515.2 EPA 515.3 EPA 515.4 EPA 555 SM 6640 B
Simazine	TPEST	GC GC/MS	GC GC/MS	EPA 505 EPA 507 EPA 508.1 EPA 523

				EPA 525.2 EPA 525.3 EPA 536 EPA 551.1
Sodium	M	FPAA FLAA IC ICP ICP/MS	FPAA FLAA IC ICP ICP/MS	ASTM D6919 EPA 200.5 Axial EPA 200.7 EPA 200.8 SM 3111B SM 3111B-99
Specific Conductance (Conductivity)	GC SCNM	ISE	ISE	ASTM D1125 (A) SM 2510 B
SPLP Extraction	WE	--	Leach test	--
β -Propiolactone	VOC	GC GC/MS	GC GC/MS	--
Strobane	CPEST	GC	GC	--
Strontium	M	FLAA ICP ICP/MS	FLAA ICP ICP/MS	--
Strychnine	PEST	GC/MS	GC/MS	--
Styrene	VOC	GC GC/MS	GC GC/MS	EPA 502.2® EPA 524.2® EPA 524.3®
Sulfallate (Thioallate)	CARB	GC GC/MS	GC GC/MS	--
Sulfate	GC SCNM	Colorimetry IC	Colorimetry IC	ASTM D4327 ASTM D516 ASTM D6508, Rev. 2 EPA 300.0 EPA 300.1 EPA 375.2 SM 4110B SM 4110B-00 SM 4500-SO42- C, D SM 4500-SO42- E SM 4500-SO42- F
Sulfide	GC	Colorimetry ISE Titration	Colorimetry ISE Titration	--
Sulfides, Acid-soluble and Acid-insoluble	GC	Titration	Titration	--
Sulfite	GC	Titration	Titration	--
Sulfotepp (Tetraethyl dithiopyrophosphate)	OPEST	GC GC/MS	GC GC/MS	--
Surfactants [Foaming agents (MBAS)]	SCNM	Colorimetry	--	SM 5540C
SUVA (calc.)	SCNM	--	--	EPA 415.3
SUVA (Specific UV Absorbance)	SCNM	--	--	EPA 415.3
t-Butyl alcohol	VOC	GC GC/MS	GC GC/MS	--
TCLP Extraction	WC	--	Leach Test	--
TCMTB	NPEST	LC	LC	--
Tebuthiuron	CARB	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--

TEPP (Tetraethylpyrophosphate)	BNANH OPEST	GC GC/MS	GC GC/MS	--
Terbacil	NPEST	GC GC/MS	GC GC/MS	--
Terbufos	OPEST	GC GC/MS	GC GC/MS	--
Terbutryn	TPEST	GC GC/MS	GC GC/MS	--
tert-Butylbenzene	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
Tetrachlorocatechol	PHEN	GC GC/MS	GC GC/MS	--
Tetrachloroethene	VOC	GC GC/MS	GC GC/MS	EPA 502.2® EPA 524.2® EPA 524.3® EPA 551.1®
Tetrachloroguaiacol	PHEN	GC GC/MS	GC GC/MS	--
Tetrachlorvinphos (Stirofos)	OPEST	GC GC/MS	GC GC/MS	--
Tetraethyl dithiopyrophosphate	BNANH	GC/MS	GC/MS	--
Tetrahydrofuran	VOC	GC/MS	GC/MS	--
Tetryl	EXPLO	LC	LC	--
Thallium	M	FLAA GFAA ICP ICP/MS	FLAA GFAA ICP ICP/MS	EPA 200.8 EPA 200.9
Thiodicarb	CARB	LC LC/MS	LC LC/MS	--
Thiofanox	CARB	LC/MS	LC/MS	--
Thionazin (O,O-Diethyl O-2-pyrazinyl phosphorothioate)	BNANH OPEST	GC GC/MS	GC GC/MS	--
Thiophanate-methyl	CARB	LC/MS	LC/MS	--
Thiophenol (Benzenethiol)	BNANH	GC/MS	GC/MS	--
Tin	M	FLAA GFAA ICP ICP/MS	FLAA GFAA ICP ICP/MS	--
Titanium	M	FLAA GFAA ICP ICP/MS	FLAA GFAA ICP ICP/MS	--
Tokuthion (Prothiofos)	OPEST	GC GC/MS	GC GC/MS	--
Toluene	VOC	GC GC/MS	GC GC/MS	EPA 502.2® EPA 524.2® EPA 524.3®
Toluene diisocyanate	BNANH	GC/MS	GC/MS	--
Toxaphene	CPEST	GC GC/MS	GC GC/MS	EPA 505 EPA 508 EPA 508.1 EPA 525.2 EPA 525.3
Triadimefon	NPEST	GC GC/MS	GC GC/MS	--

Triallate	CARB	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
Trichloroethene	VOC	GC GC/MS	GC GC/MS	EPA 502.2® EPA 524.2® EPA 524.3® EPA 551.1®
Trichlorofluoromethane (Fluorotrichloromethane)	VOC	GC GC/MS	GC GC/MS	EPA 502.2 EPA 524.2 EPA 524.3
Trichloronate	OPEST	GC GC/MS	GC GC/MS	--
Trichlorosyringol	PHEN	GC GC/MS	GC GC/MS	--
Trichlorphon	OPEST	GC GC/MS LC LC/MS	GC GC/MS LC LC/MS	--
Triclopyr	APEST	GC GC/MS LC	GC GC/MS LC	--
Trifluralin	NPEST	GC GC/MS	GC GC/MS	--
Trimethyl phosphate	BNANH	GC/MS	GC/MS	--
Tri-o-cresyl phosphate (TOCP)	OPEST	GC GC/MS	GC GC/MS	--
Tri-p-tolyl phosphate	BNANH	GC/MS	GC/MS	--
Tris(2,3-dibromopropyl) phosphate	BNANH	GC/MS	GC/MS	--
Tungsten	M	ICP ICP/MS	ICP ICP/MS	--
Turbidity	GC SCNM	Colorimetry	--	AMI Turbiwell EPA 180.1 GLI Method 2 HACH FilterTrak 10133 Mitchell M5271 Mitchell M5331 Orion AQ4500 SM 2130B
UV254	SCNM	--	--	EPA 415.3 SM 5910B
Vanadium	M	Colorimetry FLAA GFAA ICP ICP/MS	Colorimetry FLAA GFAA ICP ICP/MS	--
Vapam	PEST	GC	GC	--
Vernolate	CARB	LC/MS	LC/MS	--
Vinyl acetate	VOC	GC GC/MS	GC GC/MS	--
Vinyl chloride	VOC	GC GC/MS	GC GC/MS	EPA 502.2® EPA 524.2® EPA 524.3®

Xylenes, Total	VOC	GC GC/MS	GC GC/MS	EPA 502.2® EPA 524.2® EPA 524.3®
Zinc	M	Colorimetry FLAA GFAA ICP ICP/MS	Colorimetry FLAA GFAA ICP ICP/MS	EPA 200.5 Axial EPA 200.7 EPA 200.8 SM 3111B SM 3111B-99 SM 3120B SM 3120B-99
Ziram	CARB	GC GC/MS	GC GC/MS	--
Zirconium	M	ICP ICP/MS	ICP ICP/MS	--

SECTION 7. NR 157.20 (1) is amended to read

For transformer fluids, waste oils, insulating liquids, and other non-polar liquids containing PCBs the extraction procedures and gas chromatographic analysis methods used shall be as defined in any that is appropriate for the material being analyzed, including “Standard Method for Analysis of Polychlorinated Biphenyls in Insulating Liquids by Gas Chromatography”, ASTM standard D 4059-86, American Society for Testing and Materials, ~~1916 Race Street, Philadelphia, PA 19103~~; or “The Determination of Polychlorinated Biphenyls in Transformer Fluid and Waste Oils”, EPA-600/4-81-045, ~~U.S. EPA Monitoring and Support Laboratory, Cincinnati, OH~~; EPA Method 3580; EPA Method 3580A, EPA Method 8082; or EPA Method 8082A, all found in “Test Methods for Evaluating Solid Waste”, SW-846, U.S. EPA and EPA Method 1668A or EPA Method 1668C.

Note: Links to the standards listed above can be found on the Wisconsin department of natural resources laboratory accreditation program website.

SECTION 8. NR 157.20 (3) is amended to read

For leachate, non-drinking groundwaters, soils, sediments, and sludges containing PCBs not regulated by a Wisconsin pollution discharge elimination system permit, the extraction procedures and gas chromatographic analysis methods used shall be as defined in ~~the method 8080A found in “Test Methods for Evaluating Solid Waste”, SW 846, U.S. EPA, Update I, November 1990, 3rd edition, November 1986~~; any that is appropriate for the material being analyzed as presented in the “Status Tables for SW-846”; EPA Method 1668A; or EPA Methods 1668C.

Note: Available from the Superintendent of Documents, U.S. Government Printing Office, Washington D.C. 20402. Links to the standards listed above can be found on the Wisconsin department of natural resources laboratory accreditation program website.

SECTION 9. NR 200.02 (6) Note is amended to read

Note: The limit of quantitation is ~~10/3 or 3.333~~ times the limit of detection established as defined under s NR 149.48 (3).

SECTION 10. NR 200.22 (1) (e) 6. b. is amended to read

Proper laboratory quality control procedures were used to generate the data. To make this demonstration, the permittee shall supply, for several representative analytical runs, the raw data for samples, instrument

~~initial and continuing calibrations, and all applicable quality control samples, calibration verifications and quality control steps. The raw data for quality control steps shall include results of replicate samples, identity of samples used for replicate samples, matrix spikes, matrix spike concentrations used, reagent blanks, method blanks and quality control limits. Raw data, replicate sample, matrix spike and quality control limit have the meanings specified in s. NR 149.03.~~

~~**Note:** In the revisions to Ch. NR 149, replicate samples are referred to and defined as "replicates".~~

SECTION 11. NR 219.04 (1) is amended to read

ANALYTICAL TEST PROCEDURES. Parameters or pollutants, for which wastewater analytical methods are approved, are listed together with test procedure descriptions and references in tables A to H. Parameters or pollutants, for which sludge analytical methods are approved, are listed together with test procedure descriptions and references in table EM. The discharge values for the listed parameters shall be determined by one of the standard analytical test procedures identified in a table under this subsection or by an alternate test procedure established under ss. NR 219.033 and ~~149.12~~ 149.42.

SECTION 12. NR 247.03 (7) is repealed.

SECTION 13. NR 258.03 (15) is amended to read.

~~"Oil and grease" means those components of process wastewater amenable to measurement by the method described in "Methods for Chemical Analysis of Water and Wastes," 1971, EPA, Analytical quality Control Laboratory, page 217. Copies of this publication are available for inspection at the office of the department of natural resources, the secretary of state's office, and the office of the legislative reference bureau, and may be obtained for personal use from the Superintendent of Documents, U.S. Government Printing Office, Washington, D.C., 20460. the methods approved for oil and grease listed in ch. NR 219.~~

SECTION 14. NR 263.03 (4) is repealed and recreated to read.

(4) OIL AND GREASE ANALYSIS. Follow the approved methods listed in ch. NR 219.

SECTION 15. NR 500.03 (146) is amended to read.

~~"Method blank" has the meaning specified in s. NR 149.03 (15) (b) 149.03 (46).~~

~~**Note:** Section NR 149.03 (15) (b) defines "method blank" to mean "a sample of a matrix devoid of or having a consistent concentration or amount of the analytes of interest processed simultaneously with and under the same conditions, preparatory and analyses steps as the associated samples."~~

SECTION 16. NR 507.17 (4) is amended to read.

ANALYTICAL METHODS. Groundwater, lysimeter and leachate samples shall be handled and analyzed in accordance with the requirements of methods listed in s. NR 149.02 (7) (a) Note "Test Methods for Evaluating Solid Waste, Physical/Chemical Methods," EPA Publication SW-846, third edition, November 1986, as amended by Updates I in July 1992, II in September 1994, IIA in August 1993, IIB in January 1995, III in December 1996 and IIIA in April 1998. The methods used shall be suitable for the matrix,

~~type of analyte, expected level of analyte, regulatory limit, and potential interferences in the samples to be tested.~~ Screening methods may not be used unless approved in writing by the department. Water supply samples shall be handled in accordance with s. NR 507.20. The department may approve alternative analytical methods under s. NR 149.42.

~~**Note:** The test methods are available at no cost at <https://www.epa.gov/hw-sw846/basic-information-about-how-use-sw-846#UseWhich>. Copies of the test methods are available for inspection at the offices of the department of natural resources, the secretary of state and the legislative reference bureau. Copies may be obtained from the superintendent of documents, U.S. government printing office, P.O. Box 371954, Pittsburgh, PA 15250-7954, (866) 512-1800, www.gpo.gov. Copies may also be obtained from the national technical information service, U.S. department of commerce, 5285 Port Royal Road, Springfield, VA 22161, (800) 553-6847, www.ntis.gov.~~

SECTION 17. NR 507.17 (5) is amended to read.

LABORATORY REQUIREMENTS. All chemical analyses shall be conducted by a laboratory certified under s. 299.11, Stats., and ch. NR 149 ~~for that test category~~. The limit of detection and the limit of quantitation shall be determined according to s. NR 149.48 (2) and (3). The analytical laboratory shall meet the requirements of the analytical method and ch. NR 149. Section NR 140.16 (4) applies to analytical results that do not meet the requirements of this subsection.

SECTION 18. NR 507.26 (3) (b) 1. is amended to read.

The limit of detection and the limit of quantitation for each parameter with a public health related groundwater standard. The limit of detection and the limit of quantitation shall be determined in accordance with a method specified by the department as required in s. NR 149.48 (2) and (3).

SECTION 19. NR 700.13 (1m) Note is amended to read.

Note: The “Modified GRO, Method for Determining Gasoline Range Organics: (WI-PUBL-SW-141) and “Modified Diesel Range Organics” (WI-PUBL-SW-14) are available online at <http://dnr.wi.gov/regulations/labcert/documents/methods/drosep95.pdf> and <http://dnr.wi.gov/regulations/labcert/documents/methods/grosep95.pdf>. ~~These methods are referenced in s. NR 149, Appendix III, List of Authoritative Sources.~~ the Wisconsin department of natural resources laboratory accreditation program website.

SECTION 20. EFFECTIVE DATE. This rule takes effect on the first day of the month following publication in the Wisconsin Administrative Register as provided in s. 227.22 (2) (intro.), Stats.

SECTION 21. BOARD ADOPTION. This rule was approved and adopted by the State of Wisconsin Natural Resources Board on December 12, 2018.

Dated at Madison, Wisconsin _____.

STATE OF WISCONSIN DNR
DEPARTMENT OF NATURAL RESOURCES

BY _____

Daniel L. Meyer, Secretary

(SEAL)